



Metropolis Healthcare Ltd., EQAS

REDCLIFFE LIFETECH PRIVATE LIMITED, THANE (997)

Outlier And Analyte Summary Report

Result Details For Cycle No 230412 and Sample No 01

Report Date : 13/02/2024

Instrument : ERBA LAURA Smart urine analyzer

Analyte	Standard Unit	Result Value	Accepted Value
@ pH	-	6	5.5,5.6
@ Specific Gravity	-	1.025	1.020,1.025,1.030
@ Glucose (urine)	-	2+	+2,+3,+4
@ Protein (Urine)	-	2+	+1,+2,+3
@ Ketone (urine)	-	Negative	Neg
@ Blood (urine)	-	1+	+1,+2,+3
@ Bilirubin (urine)	-	Negative	Neg
@ Nitrite (urine)	-	Present	Pre

Legend @ : Acceptable
 @ : Acceptable
 X : Unacceptable
 # : Not Evaluated
 ⊙ : Delayed Result Entry

Total Parameters	8
Not Evaluated Parameters	0
Evaluated Parameters	8
Outlier Parameters (X)	0
EQAS Score Clinical Path - Urine Routine Chemistry	100.00 %

AK Nigam

Dr Puneet Kumar Nigam
 PT coordinator & Technical Manager, MHL EQAS
 Unit No. 409-416.
 Commercial Building - 1A
 Kohinoor Mall, Kirod Road, Kurla (W),
 Mumbai - 400070