



Abbott Healthcare Pvt. Ltd.
Floor 13-19, Godrej BKC, Plot C-68
G Block, Bandra Kurla Complex
Near MCA Club, Bandra East
Mumbai - 400051. India

Registered Office:
4, Corporate Park,
Sion Trombay Road,
Mumbai – 400071. India

T : +91 22 3816 2000 / 1010
F : +91 22 3816 2400 / 2401
Email : webmaster@abbott.co.in
Site : www.abbottdiagnostics.com
CIN : U24200MH1997PTC104834

Calibration Certificate

28th February 2024

To,
Redcliffe Lifetech Pvt Ltd
Thane 400602

Dear Madam,

This is with reference to the Alinity Ci Immunoassay-Chemistry analyzer installed in your laboratory. Please find below the details of the instrument Calibrations and the Verifications performed.

The next calibration will be due on 27th February 2025

Instrument : Abbott Alinity Ci Immunoassay-Chemistry Analyzer.

Serial number : AC05519, Ai24674

Calibration performed on: 28th February 2024

1. System voltage checks performed using Digital Multimeter and confirmed to specifications.
2. Sample & Reagent Pipettor's Calibration Passed.
3. RSM Transport Calibration Passed.
4. Barcode Calibration Passed.
5. Optics Background Passed.

Thanking you,

Yours Truly,

A handwritten signature in black ink, appearing to read 'Parag Mehta'.

Parag Mehta
Technical Service Specialist
Abbott Healthcare Pvt. Ltd.

CALIBRATION DETAILS REPORT

Operator ID: Admin

Site:

System Name: Alinity ci-series

SN: SCM23474

Software Version: 3.5.1

Assay and Reagent Information

Assay Name: Creaz

Assay Number: 1436

Assay Version: 8

Module/SN: 1 / AC05519

Operator ID: Admin

Reagent Lot: 51614UD00

Reagent SN: 16358

Lot Expiration: 18.07.2024

Calibration Information

Cal Status: Active

Cal Type: Full

Cal Method: Linear

Calibrator Lot: 0001275UE

Lot Expiration: 28.12.2024

Calibrator SN:

Cal Date/Time: 26.03.2024 11:27

Archive Status: Not Archived

Archive Date/Time:

Full Calibration Expiration Date/Time: 05.04.2024 11:27
Transmission Status:

Adjust Calibration Expiration Date/Time:
Transmission Date/Time:

Message:

Calibrator Results

CALIBRATOR ID	CONCENTRATION mg/dL	ABSORBANCE	FACTOR	REP 1 ABSORBANCE	REP 2 ABSORBANCE	REP 3 ABSORBANCE
Blank	0.0000	0.0003		0.0002	0.0003	0.0004
Level 1	0.6743	0.0049	137.4422	0.0055	0.0052	0.0052
Level 2	4.1900	0.0233	190.7888	0.0236	0.0236	0.0238

***** End of Report *****

Signature: 

Date:

CALIBRATION DETAILS REPORT

Operator ID: Admin

Site:

System Name: Allinity ci-series

SN: SCM23474

Software Version: 3.5.1

Assay and Reagent Information

Assay Name: ALT2

Assay Number: 1426

Assay Version: 5

Module/SN: 1 / AC05519

Operator ID: Admin

Reagent Lot: 57900UD00

Reagent SN: 14962

Lot Expiration: 21.02.2025

Calibration Information

Cal Status: Active

Cal Type: Full

Cal Method: Linear

Calibrator Lot: 0001275UE

Lot Expiration: 28.12.2024

Calibrator SN:

Cal Date/Time: 26.03.2024 12:45

Archive Status: Not Archived

Archive Date/Time:

Full Calibration
Expiration Date/Time: 25.04.2024 12:45

Adjust Calibration
Expiration Date/Time:

Transmission
Status:

Transmission
Date/Time:

Message:

Calibrator Results

CALIBRATOR ID	CONCENTRATION U/L	ABSORBANCE	FACTOR	REP 1 ABSORBANCE	REP 2 ABSORBANCE	REP 3 ABSORBANCE
Blank	0.0000	-0.0001		-0.0001	-0.0001	-0.0003
Level 1	126.8000	-0.0147	-8,630.6754	-0.0148	-0.0148	-0.0149

***** End of Report *****

Signature:

Date:

CALIBRATION DETAILS REPORT

Operator ID: Admin

Site:

System Name: Allinity ci-series

SN: SCM23474

Software Version: 3.5.1

Assay and Reagent Information

Assay Name: AST2	Module/SN: 1 / AC05519
Assay Number: 1424	Operator ID: Admin
Assay Version: 4	Reagent Lot: 54436UD00
	Reagent SN: 04717
	Lot Expiration: 17.01.2025

Calibration Information

Cal Status: Active	Calibrator Lot: 0001275UE
Cal Type: Full	Lot Expiration: 28.12.2024
Cal Method: Linear	Calibrator SN:
	Full Calibration: 21.04.2024 09:51
	Expiration Date/Time:
	Transmission Status:
	Archive Date/Time:
	Adjust Calibration
	Expiration Date/Time:
	Transmission Date/Time:
	Archive Status: Not Archived
	Cal Date/Time: 22.03.2024 09:51


Message:

Calibrator Results

CALIBRATOR ID	CONCENTRATION U/L	ABSORBANCE	FACTOR	REP 1 ABSORBANCE	REP 2 ABSORBANCE	REP 3 ABSORBANCE
Blank	0.0000	-0.0006		-0.0006	-0.0006	-0.0007
Level 1	159.6000	-0.0166	-9.590.5154	-0.0172	-0.0172	-0.0175

***** End of Report *****

Signature: _____ Date: _____



CALIBRATION DETAILS REPORT

Operator ID: Admin

Site:

System Name: Alinity ci-series

SN: SCM23474

Software Version: 3.5.1

Assay and Reagent Information

Assay Name: Bill Total
 Assay Number: 1094
 Assay Version: 6

Module/SN: 1 / AC05519
 Operator ID: Admin

Reagent Lot: 63906UQ03
 Reagent SN: 04378
 Lot Expiration: 31.08.2024

Calibration Information

Cal Status: Active
 Cal Type: Full
 Cal Method: Linear

Calibrator Lot: 60667FD01
 Lot Expiration: 30.06.2024
 Calibrator SN:

Cal Date/Time: 26.03.2024 11:27
 Archive Status: Not Archived
 Archive Date/Time:

Full Calibration
 Expiration Date/Time: 09.04.2024 11:27
 Transmission Status:

Adjust Calibration
 Expiration Date/Time:
 Transmission Date/Time:


Message:

Calibrator Results

CALIBRATOR ID	CONCENTRATION mg/dL	ABSORBANCE	FACTOR	REP 1 ABSORBANCE	REP 2 ABSORBANCE	REP 3 ABSORBANCE
Blank	0.0000	0.0064		0.0063	0.0064	0.0067
Level 1	1.4900	0.0110	135.7298	0.0174	0.0173	0.0174
Level 2	16.7600	0.1306	127.6545	0.1373	0.1369	0.1370

***** End of Report *****

Signature: _____ Date: _____



CALIBRATION DETAILS REPORT

Operator ID: Admin

SN: SCM23474

Site:

Software Version: 3.5.1

System Name: Alinity ci-series

Assay and Reagent Information

Assay Name: Bill D
 Assay Number: 1065
 Assay Version: 4

Module/SN: 1 / AC05519
 Operator ID: Admin

Reagent Lot: 64359UQ06
 Reagent SN: 07936
 Lot Expiration: 30.11.2024

Calibration Information

Cal Status: Active
 Cal Type: Full
 Cal Method: Linear

Calibrator Lot: 60667FD01
 Lot Expiration: 30.06.2024
 Calibrator SN:

Cal Date/Time: 13.03.2024 09:43
 Archive Status: Not Archived

Full Calibration
 Expiration Date/Time: 27.03.2024 09:43
 Transmission
 Status:

Archive Date/Time:
 Adjust Calibration
 Expiration Date/Time:
 Transmission
 Date/Time:

Message:

Calibrator Results

CALIBRATOR ID	CONCENTRATION mg/dL	ABSORBANCE	FACTOR	REP 1 ABSORBANCE	REP 2 ABSORBANCE	REP 3 ABSORBANCE
Blank	0.0000	0.0000		0.0000	0.0001	0.0000
Level 1	0.8100	0.0174	46.4931	0.0172	0.0175	0.0174
Level 2	8.3800	0.1679	50.2933	0.1679	0.1677	0.1682

***** End of Report *****

Signature _____



Date: _____

CALIBRATION DETAILS REPORT

SN: SCM23474

Software Version: 3.5.1

Operator ID: Admin

Site:

System Name: Alinity ci-series

Assay and Reagent Information

Assay Name: ALKP2

Assay Number: 1448

Assay Version: 4

Module/SN: 1 / AC05519

Operator ID: Admin

Reagent Lot: 56118UD00

Reagent SN: 05595

Lot Expiration: 10.11.2024

Calibration Information

Cal Status: Active

Cal Type: Full

Cal Method: Linear

Calibrator Lot: 0001275UE

Lot Expiration: 28.12.2024

Calibrator SN:

Full Calibration Expiration Date/Time: 27.03.2024 15:34

Transmission Status:

Cal Date/Time: 17.03.2024 15:34

Archive Status: Not Archived

Archive Date/Time: Adjust Calibration Expiration Date/Time:

Transmission Date/Time:

Message:

Calibrator Results

CALIBRATOR ID	CONCENTRATION U/L	ABSORBANCE	FACTOR	REP 1 ABSORBANCE	REP 2 ABSORBANCE	REP 3 ABSORBANCE
Blank	0.0000	0.0022		0.0028	0.0022	0.0021
Level 1	355.8000	0.2004	1,775.1297	0.2026	0.2024	0.2038

***** End of Report *****

Signature: _____ Date: _____



CALIBRATION DETAILS REPORT

Operator ID: Admin

Site:

System Name: Alinity ci-series

SN: SCM23474
Software Version: 3.5.1

Assay and Reagent Information

Assay Name: Trg2
Assay Number: 1417
Assay Version: 5

Module/SN: 1 / AC05519
Operator ID: Admin

Reagent Lot: 56401UD00
Reagent SN: 04237
Lot Expiration: 13.12.2024

Calibration Information

Cal Status: Active
Cal Type: Full
Cal Method: Linear

Calibrator Lot: 0001275UE
Lot Expiration: 28.12.2024
Calibrator SN:

Cal Date/Time: 24.03.2024 09:57
Archive Status: Not Archived
Archive Date/Time:

Full Calibration
Expiration Date/Time: 23.04.2024 09:57
Transmission Status:

Adjust Calibration
Expiration Date/Time: 08.04.2024 09:57
Transmission Date/Time:

Message:

Calibrator Results

CALIBRATOR ID	CONCENTRATION mg/dL	ABSORBANCE	FACTOR	REP 1 ABSORBANCE	REP 2 ABSORBANCE	REP 3 ABSORBANCE
Blank	0.0000	0.0165		0.0167	0.0161	0.0165
Level 1	30.5495	0.0482	634.4653		0.0650	0.0643
Level 2	122.1980	0.1989	607.9503	0.2140	0.2154	0.2161
Level 3	242.0000	0.3929	617.5360	0.4093	0.4100	0.4094

***** End of Report *****

Signature: _____

Date: _____



CALIBRATION DETAILS REPORT

Operator ID: Admin
 Site:
 System Name: Alinity ci-series

SN: SCM23474
 Software Version: 3.5.1

Assay and Reagent Information

Assay Name: CHO12
 Assay Number: 1456
 Assay Version: 3

Module/SN: 1 / AC05519
 Operator ID: Admin

Reagent Lot: 52900UD00
 Reagent SN: 13373
 Lot Expiration: 03.07.2024

Calibration Information

Cal Status: Active
 Cal Type: Full
 Cal Method: Linear

Calibrator Lot: 0001275UE
 Lot Expiration: 28.12.2024
 Calibrator SN:

Cal Date/Time: 14.03.2024 09:31
 Archive Status: Not Archived
 Archive Date/Time:

Full Calibration: 13.04.2024 09:31
 Expiration Date/Time:
 Transmission Status:

Adjust Calibration: 29.03.2024 09:31
 Expiration Date/Time:
 Transmission Date/Time:


Message:

Calibrator Results

CALIBRATOR ID	CONCENTRATION mg/dL	ABSORBANCE	FACTOR	REP 1 ABSORBANCE	REP 2 ABSORBANCE	REP 3 ABSORBANCE
Blank	0.0000	0.0712		0.0714	0.0710	
Level 1	65.8913	0.1854	355.4964	0.2566	0.2565	
Level 2	188.2607	0.5477	337.7573	0.6216	0.6161	
Level 3	281.0000	0.8071	357.4457	0.8779	0.8787	

***** End of Report *****

Signature: _____ Date: _____



CALIBRATION DETAILS REPORT

Operator ID: Admin

Site:

System Name: Alinity ci-series

SN: SCM23474

Software Version: 3.5.1

Assay and Reagent Information

Assay Name: Na-C

Assay Number: 1101

Assay Version: 5

Module/SN: 1 / AC05519

Operator ID: Admin

Reagent Lot: 67538UN22

Reagent SN: 14644

Lot Expiration: 27.12.2024

Calibration Information

Cal Status: Active

Cal Type: Full

Cal Method: ICT

Calibrator Lot: 21596UN23

Lot Expiration: 13.06.2024

Calibrator SN:

Cal Date/Time: 26.03.2024 10:42

Archive Status: Not Archived

Archive Date/Time:

Full Calibration Expiration Date/Time: 27.03.2024 10:42

Transmission Status:

Transmission Date/Time:

Message:

Calibrator Results

CALIBRATOR ID	CONCENTRATION	CAL mV	SLOPE	REP 1 mV	REP 2 mV	REP 3 mV
Low	120.0000	-4.3336	94.5282	-4.2954	-4.4785	-4.3336
High	160.0000	2.9297	94.5282	2.9297	2.9221	3.0365

***** End of Report *****

Signature: _____ Date: _____



CALIBRATION DETAILS REPORT

Operator ID: Admin

Site:

System Name: Alinity ci-series

SN: SCM23474

Software Version: 3.5.1

Assay and Reagent Information

Assay Name: K-C

Assay Number: 1102

Assay Version: 5

Module/SN: 1 / ACO5519

Operator ID: Admin

Reagent Lot: 67538UN22

Reagent SN: 14644

Lot Expiration: 27.12.2024

Calibration Information

Cal Status: Active

Cal Type: Full

Cal Method: ICT

Calibrator Lot: 21596UN23

Lot Expiration: 13.06.2024

Calibrator SN:

Cal Date/Time: 26.03.2024 10:42

Archive Status: Not Archived

Archive Date/Time:

Full Calibration Expiration Date/Time: 27.03.2024 10:42

Transmission Status:

Transmission Date/Time:

Message:

Calibrator Results

CALIBRATOR ID	CONCENTRATION mmol/L	CAL mv	SLOPE	REP 1 mv	REP 2 mv	REP 3 mv
Low	3.4000	-4.3259	96.7456	-4.2115	-4.3717	-4.3259
High	8.0000	17.7844	96.7456	17.7768	17.7844	17.8378

***** End of Report *****

Signature: _____



Date: _____

CALIBRATION DETAILS REPORT

Operator ID: Admin

Site:

System Name: Alinity ci-series

SN: SCM23474

Software Version: 3.5.1

Assay and Reagent Information

Assay Name: Cl-C

Assay Number: 1103

Assay Version: 5

Module/SN: 1 / AC05519

Operator ID: Admin

Reagent Lot: 67538UN22

Reagent SN: 14644

Lot Expiration: 27.12.2024

Calibration Information

Cal Status: Active

Cal Type: Full

Cal Method: ICT

Calibrator Lot: 21596UN23

Lot Expiration: 13.06.2024

Calibrator SN:

Cal Date/Time: 26.03.2024 10:42

Archive Status: Not Archived

Archive Date/Time:

Full Calibration 27.03.2024 10:42
Expiration Date/Time:
Transmission Status:


Transmission Date/Time:

Message:

Calibrator Results

CALIBRATOR ID	CONCENTRATION mmol/L	CAL mV	SLOPE	REP 1 mV	REP 2 mV	REP 3 mV
Low	80.0000	4.4633	78.8341	4.3183	4.4633	4.6845
High	120.0000	-4.0742		-4.0742	-4.1047	-3.9597

***** End of Report *****

Signature: 

Date:

CALIBRATION DETAILS REPORT

Operator ID: Admin

SN: SCM23474

Site:

Software Version: 3.5.1

System Name: Alinity ci-series

Assay and Reagent Information

Assay Name: Ultra HDL

Module/SN: 1 / AC05519

Reagent Lot: 63680UC02

Assay Number: 1293

Operator ID: Admin

Reagent SN: 07973

Assay Version: 7

Lot Expiration: 31.01.2025

Calibration Information

Cal Status: Active

Calibrator Lot: 2976CH

Cal Date/Time: 13.03.2024 11:10

Cal Type: Full

Lot Expiration: 28.03.2025

Archive Status: Not Archived

Cal Method: Linear

Calibrator SN:

Archive Date/Time:

Full Calibration
Expiration Date/Time: 10.04.2024 11:10

Adjust Calibration
Expiration Date/Time:

Transmission
Status:

Transmission
Date/Time:

Message:

Calibrator Results

CALIBRATOR ID	CONCENTRATION mg/dL	ABSORBANCE	FACTOR	REP 1 ABSORBANCE	REP 2 ABSORBANCE	REP 3 ABSORBANCE
Blank	0.0000	0.0033		0.0033	0.0033	0.0033
Level 1	60.0000	0.0695	862.7096	0.0731	0.0728	0.0727

***** End of Report *****

Signature: 

Date:

CALIBRATION DETAILS REPORT

Operator ID: Admin

Site:

System Name: Alinity ci-series

SN: SCM23474

Software Version: 3.5.1

Assay and Reagent Information

Assay Name: Direct LDL

Assay Number: 1840

Assay Version: 5

Module/SN: 1 / AC05519

Operator ID: Admin

Reagent Lot: 63138UQ12

Reagent SN: 10696

Lot Expiration: 04.12.2024

Calibration Information

Cal Status: Active

Cal Type: Full

Cal Method: Linear

Calibrator Lot: 2976CH

Lot Expiration: 28.03.2025

Calibrator SN:

Full Calibration Expiration Date/Time: 16.04.2024 10:00

Transmission Status:

Cal Date/Time: 19.03.2024 10:00

Archive Status: Not Archived

Archive Date/Time:

Adjust Calibration Expiration Date/Time:

Transmission Date/Time:

Message:

Calibrator Results

CALIBRATOR ID	CONCENTRATION mg/dL	ABSORBANCE	FACTOR	REP 1 ABSORBANCE	REP 2 ABSORBANCE	REP 3 ABSORBANCE
Blank	0.0000	0.0006		0.0007	0.0006	0.0006
Level 1	120.0000	0.2338	513.3473	0.2354	0.2336	0.2344

***** End of Report *****

Signature: _____ Date: _____



CALIBRATION DETAILS REPORT

Operator ID: Admin

Site:

System Name: Alinity ci-series

SN: SCM23474
Software Version: 3.5.1

Assay and Reagent Information

Assay Name: Uric2
Assay Number: 1444
Assay Version: 3

Module/SN: 1 / AC05519
Operator ID: Admin

Reagent Lot: 53175UD00
Reagent SN: 04850
Lot Expiration: 28.08.2024

Calibration Information

Cal Status: Active
Cal Type: Full
Cal Method: Linear

Calibrator Lot: 0001275UE
Lot Expiration: 28.12.2024
Calibrator SN:
Full Calibration Expiration Date/Time: 13.04.2024 09:28
Transmission Status:

Cal Date/Time: 14.03.2024 09:28
Archive Status: Not Archived
Archive Date/Time:
Adjust Calibration Expiration Date/Time:
Transmission Date/Time:

Message:

Calibrator Results

CALIBRATOR ID	CONCENTRATION mg/dL	ABSORBANCE	FACTOR	REP 1 ABSORBANCE	REP 2 ABSORBANCE	REP 3 ABSORBANCE
Blank	0.0000	-0.0001		-0.0001	0.0000	
Level 1	3.8818	0.0341	113.7352	0.0342	0.0339	
Level 2	7.4796	0.0661	112.5002	0.0660	0.0661	
Level 3	9.4800	0.0843	110.2076	0.0838	0.0845	

***** End of Report *****

Signature: _____



Date: _____

CALIBRATION DETAILS REPORT

Operator ID: Admin
 Site:
 System Name: Alinity ci-series

SN: SCM23474
 Software Version: 3.5.1

Assay and Reagent Information

Assay Name: GT2
 Assay Number: 1476
 Assay Version: 5

Module/SN: 1 / AC05519
 Operator ID: Admin

Reagent Lot: 53379UD00
 Reagent SN: 20589
 Lot Expiration: 24.08.2024

Calibration Information

Cal Status: Active
 Cal Type: Full
 Cal Method: Linear

Calibrator Lot: 000127SUE
 Lot Expiration: 28.12.2024
 Calibrator SN:

Cal Date/Time: 29.02.2024 09:07
 Archive Status: Not Archived

Full Calibration
 Expiration Date/Time: 30.03.2024 09:07
 Transmission
 Status:

Archive Date/Time:
 Adjust Calibration
 Expiration Date/Time:
 Transmission
 Date/Time:


Message:

Calibrator Results

CALIBRATOR ID	CONCENTRATION U/L	ABSORBANCE	FACTOR	REP 1 ABSORBANCE	REP 2 ABSORBANCE	REP 3 ABSORBANCE
Blank	0.0000	0.0001		0.0001	0.0001	0.0001
Level 1	187.9000	0.0164	11,485.3507	0.0166	0.0164	0.0165

***** End of Report *****

Signature: _____ Date: _____



CALIBRATION DETAILS REPORT

Operator ID: Admin

Site:

System Name: Alinity ci-series

SN: SCM23474

Software Version: 3.5.1

Assay and Reagent Information

Assay Name: Phosphorus

Assay Number: 1011

Assay Version: 5

Module/SN: 1 / AC05519

Operator ID: Admin

Reagent Lot: 60174UN23-13138

Reagent SN: 13138

Lot Expiration: 29.07.2024

Calibration Information

Cal Status: Active

Cal Type: Full

Cal Method: Linear

Calibrator Lot: 36180FD01

Lot Expiration: 07.07.2024

Calibrator SN:

Cal Date/Time: 26.03.2024 13:32

Archive Status: Not Archived

Archive Date/Time:

Adjust Calibration

Expiration Date/Time:

Transmission

Date/Time:

Message:

Calibrator Results

CALIBRATOR ID	CONCENTRATION mg/dL	ABSORBANCE	FACTOR	REP 1 ABSORBANCE	REP 2 ABSORBANCE	REP 3 ABSORBANCE
Blank	0.0000	0.0998		0.0998	0.1000	0.0993
Level 1	3.8900	0.1464	26.5677	0.2462	0.2464	0.2460
Level 2	8.0600	0.3108	25.3730	0.4084	0.4106	0.4109

***** End of Report *****

Signature: _____ Date: _____



Operator ID: Admin
Site:

System Name: Alinity ci-series

CALIBRATION DETAILS REPORT

SN: SCM23474
Software Version: 3.5.1

Assay and Reagent Information

Assay Name: BUN
Assay Number: 1005
Assay Version: 7

Module/SN: 1 / AC05519
Operator ID: Admin

Reagent Lot: 58917UN23-28533
Reagent SN: 28533
Lot Expiration: 13.08.2024

Calibration Information

Cal Status: Active
Cal Type: Full
Cal Method: Linear

Calibrator Lot: 36180F01
Lot Expiration: 07.07.2024
Calibrator SN:
Full Calibration
Expiration Date/Time: 02.04.2024 10:52
Transmission
Status:

Cal Date/Time: 26.03.2024 10:52
Archive Status: Not Archived
Archive Date/Time:
Adjust Calibration
Expiration Date/Time:
Transmission
Date/Time:

Message:

Calibrator Results

CALIBRATOR ID	CONCENTRATION mg/dL	ABSORBANCE	FACTOR	REP 1 ABSORBANCE	REP 2 ABSORBANCE	REP 3 ABSORBANCE
Blank	0.0000	-0.0030		-0.0030	-0.0038	-0.0029
Level 1	18.4000	-0.0420	-437.8219	-0.0450	-0.0452	-0.0447
Level 2	54.4000	-0.1157	-488.5547	-0.1189	-0.1187	-0.1184

***** End of Report *****

Signature: 

Date:

CALIBRATION DETAILS REPORT

Operator ID: Admin

Site:

System Name: Allinity ci-series

SN: SCM23474

Software Version: 3.5.1

Assay and Reagent Information

Assay Name: Calcium

Assay Number: 1066

Assay Version: 5

Module/SN: 1 / AC05519

Operator ID: Admin

Reagent Lot: 26375UN23

Reagent SN: 27111

Lot Expiration: 01.05.2024

Calibration Information

Cal Status: Active

Cal Type: Full

Cal Method: Linear

Calibrator Lot: 99191FD01

Lot Expiration: 25.03.2024

Calibrator SN:

Full Calibration Expiration Date/Time: 06.04.2024 09:03

Transmission Status:

Cal Date/Time: 07.03.2024 09:03

Archive Status: Not Archived

Archive Date/Time:

Adjust Calibration Expiration Date/Time:

Transmission Date/Time:

Message:

Calibrator Results

CALIBRATOR ID	CONCENTRATION mg/dL	ABSORBANCE	FACTOR	REP 1 ABSORBANCE	REP 2 ABSORBANCE	REP 3 ABSORBANCE
Blank	0.0000	0.1626		0.1641	0.1609	0.1626
Level 1	8.8500	0.6743	13.1241	0.8369		
Level 2	11.8300	0.8809	14.4263	1.0435	1.0435	1.0416

***** End of Report *****

Signature:



Date:

CALIBRATION DETAILS REPORT

Operator ID: Admin

SN: SCM23474

Site:

Software Version: 3.5.1

System Name: Alinity ci-series

Assay and Reagent Information

Assay Name: AlbBCG2
 Assay Number: 1400
 Assay Version: 5

Module/SN: 1 / AC05519

Reagent Lot: 51238UD00

Operator ID: Admin

Reagent SN: 00637

Lot Expiration: 05.07.2024

Calibration Information

Cal Status: Active
 Cal Type: Full
 Cal Method: Linear

Calibrator Lot: 0001275UE

Cal Date/Time: 26.03.2024 10:43

Lot Expiration: 28.12.2024

Archive Status: Not Archived

Callibrator SN:

Archive Date/Time:

Full Calibration Expiration Date/Time: 07.05.2024 10:43

Adjust Calibration Expiration Date/Time:

Transmission Status:

Transmission Date/Time:

Message:

Callibrator Results

CALIBRATOR ID	CONCENTRATION g/dL	ABSORBANCE	FACTOR	REP 1 ABSORBANCE	REP 2 ABSORBANCE	REP 3 ABSORBANCE
Blank	0.0000	0.1512		0.1521	0.1502	
Level 1	1.7243	0.6944	2.4831	0.8462	0.8449	
Level 2	2.8900	1.0715	3.0914	1.2233	1.2220	

***** End of Report *****

Signature: _____ Date: _____



CALIBRATION DETAILS REPORT

Operator ID: Admin

Site:

System Name: Alinity ci-series

SN: SCM23474

Software Version: 3.5.1

Assay and Reagent Information

Assay Name: TPRO2

Assay Number: 1406

Assay Version: 5

Module/SN: 1 / AC05519

Operator ID: Admin

Reagent Lot: 50596UD00

Reagent SN: 06248

Lot Expiration: 08.06.2024

Calibration Information

Cal Status: Active

Cal Type: Full

Cal Method: Linear

Calibrator Lot: 0001275UE

Lot Expiration: 28.12.2024

Calibrator SN:

Full Calibration Expiration Date/Time: 07.04.2024 09:40

Transmission Status:

Cal Date/Time: 08.03.2024 09:40

Archive Status: Not Archived

Archive Date/Time:

Adjust Calibration Expiration Date/Time:

Transmission Date/Time:

Message:

Calibrator Results

CALIBRATOR ID	CONCENTRATION g/dL	ABSORBANCE	FACTOR	REP 1 ABSORBANCE	REP 2 ABSORBANCE	REP 3 ABSORBANCE
Blank	0.0000	-0.0839		-0.0839	-0.0839	
Level 1	1.8308	0.1783	10.2708	0.0954	0.0933	
Level 2	2.8294	0.2730	10.5393	0.1894	0.1888	
Level 3	4.7100	0.4500	10.6251	0.3662	0.3660	

***** End of Report *****

Signature: _____ Date: _____



CALIBRATION DETAILS REPORT

Operator ID: Admin

Site:

System Name: Alinity ci-series

SN: SCM23474

Software Version: 3.5.1

Assay and Reagent Information

Assay Name: LDH

Assay Number: 1248

Assay Version: 5

Module/SN: 1 / AC05519

Operator ID: Admin

Reagent Lot: 70279UN22

Reagent SN: 02241

Lot Expiration: 08.11.2023

Calibration Information

Cal Status: Active

Cal Type: Full

Cal Method: Factor

Calibrator Lot:

Lot Expiration:

Calibrator SN:

Full Calibration Expiration Date/Time: 19.04.2024 09:43

Transmission Status:

Cal Date/Time: 20.03.2024 09:43

Archive Status: Not Archived

Archive Date/Time:

Adjust Calibration Expiration Date/Time:

Transmission Date/Time:

Message:

Calibrator Results

CALIBRATOR ID	CONCENTRATION U/L	ABSORBANCE	FACTOR	REP 1 ABSORBANCE	REP 2 ABSORBANCE	REP 3 ABSORBANCE
Blank	0.0000	0.0003	11,180.0000	0.0003	0.0003	0.0005

***** End of Report *****

Signature: _____ Date: _____



CALIBRATION DETAILS REPORT

Operator ID: Admin

Site:

System Name: Alinity ci-series

SN: SCM23474

Software Version: 3.5.1

Assay and Reagent Information

Assay Name: Glucose

Assay Number: 1069

Assay Version: 7

Module/SN: 1 / AC05519

Operator ID: Admin

Reagent Lot: 63822UQ03

Reagent SN: 10382

Lot Expiration: 31.03.2024

Calibration Information

Cal Status: Active

Cal Type: Full

Cal Method: Linear

Calibrator Lot: 99191FD01

Lot Expiration: 25.03.2024

Calibrator SN:

Cal Date/Time: 02.03.2024 10:02

Archive Status: Not Archived

Archive Date/Time:

Full Calibration
Expiration Date/Time: 01.04.2024 10:02

Transmission
Status:

Adjust Calibration
Expiration Date/Time:

Transmission
Date/Time:

Message:

Calibrator Results

CALIBRATOR ID	CONCENTRATION mg/dL	ABSORBANCE	FACTOR	REP 1 ABSORBANCE	REP 2 ABSORBANCE	REP 3 ABSORBANCE
Blank	0.0000	-0.0007		-0.0006	-0.0007	-0.0014
Level 1	93.6000	0.2011	465.4401	0.2004	0.2003	0.2009
Level 2	437.7000	0.8765	509.4759	0.8758	0.8767	0.8758

***** End of Report *****

Signature: _____ Date: _____



QC SUMMARY REPORT

Operator ID: Admin
 Site:
 System Name: Alinity ci-series

SN: SCM23474
 Software Version: 3.5.1

Control Name: BIOCHEM-1
 Control Lot: 89730

Module/SN: 1 / AC05519
 Date Range: 26.03.2024 to 26.03.2024

Assay	Level	N	Flagged	Actual Mean	Actual SD	Actual %CV	Expected Mean	Expected SD	Expected Range
BUN	LEVEL1	1	0	14.00000	0.00000	0.00000	14.00000	1.10000	11.80000 - 16.20000
Calcium	LEVEL1	1	0	9.73000	0.00000	0.00000	9.00000	0.60000	7.80000 - 10.20000
Cl-C	LEVEL1	1	0	96.40000	0.00000	0.00000	90.90000	3.80000	83.30000 - 98.50000
Crea2	LEVEL1	1	0	2.26000	0.00000	0.00000	1.83500	0.22250	1.39000 - 2.28000
Direct LDL	LEVEL1	1	0	105.29000	0.00000	0.00000	103.30000	9.35000	84.60000 - 122.00000
Glucose	LEVEL1	1	0	70.61000	0.00000	0.00000	74.00000	5.80000	62.40000 - 85.60000
K-C	LEVEL1	1	0	3.80000	0.00000	0.00000	3.87500	0.14250	3.59000 - 4.16000
Na-C	LEVEL1	1	0	150.30000	0.00000	0.00000	149.00000	4.50000	140.00000 - 158.00000
Phosphorus	LEVEL1	3	0	3.46000	0.02000	0.57803	3.70500	0.39750	2.91000 - 4.50000
TPRO2	LEVEL1	1	0	6.80000	0.00000	0.00000	6.70000	0.75000	5.20000 - 8.20000
Ultra HDL	LEVEL1	1	0	88.08000	0.00000	0.00000	91.50000	8.25000	75.00000 - 108.00000
Chol2	LEVEL1	1	0	239.94000	0.00000	0.00000	228.00000	15.50000	197.00000 - 259.00000
AlbBCG2	LEVEL1	1	0	4.39000	0.00000	0.00000	4.00000	0.24000	3.52000 - 4.48000
Billi Total	LEVEL1	1	0	1.36000	0.00000	0.00000	1.18150	0.13925	0.90300 - 1.46000
Billi D	LEVEL1	1	0	0.52000	0.00000	0.00000	0.59000	0.06500	0.46000 - 0.72000
Uric2	LEVEL1	1	0	4.39000	0.00000	0.00000	4.68000	0.25500	4.17000 - 5.19000
UIBC	LEVEL1	1	0	89.00000	0.00000	0.00000	76.75000	10.87500	55.00000 - 98.50000
ALT2	LEVEL1	1	0	27.15000	0.00000	0.00000	28.15000	5.12500	17.90000 - 38.40000
AST2	LEVEL1	1	0	50.00000	0.00000	0.00000	45.35000	7.32500	30.70000 - 60.00000
AlKP2	LEVEL1	1	0	120.00000	0.00000	0.00000	119.50000	22.75000	74.00000 - 165.00000
GGT2	LEVEL1	1	0	63.00000	0.00000	0.00000	55.75000	7.12500	41.50000 - 70.00000

Aman Khan

QC SUMMARY REPORT

SN: SCM23474

Operator ID: Admin
Site:

Software Version: 3.5.1

System Name: Alinity ci-series

Assay	Level	N	Flagged	Actual Mean	Actual SD	Actual %CV	Expected Mean	Expected SD	Expected Range
Iron2	LEVEL1	1	0	227.00000	0.00000	0.00000	252.00000	25.00000	202.00000 - 302.00000
Trig2	LEVEL1	1	0	160.00000	0.00000	0.00000	164.00000	20.00000	124.00000 - 204.00000

Amantour

QC SUMMARY REPORT

SN: SCM23474

Software Version: 3.5.1

Operator ID: Admin

Site:

System Name: Alinity ci-series

Control Name: BIOCHEM-1

Control Lot: 89710

Module/SN: 1 / AC05519

Date Range: 26.03.2024 to 26.03.2024

Assay	Level	N	Flagged	Actual Mean	Actual SD	Actual %CV	Expected Mean	Expected SD	Expected Range
BUN	LEVEL2	1	0	43.40000	0.00000	0.00000	43.85000	2.82500	38.20000 - 49.50000
Calcium	LEVEL2	1	0	13.51000	0.00000	0.00000	12.85000	0.77500	11.30000 - 14.40000
Cl-C	LEVEL2	1	0	85.90000	0.00000	0.00000	83.10000	3.90000	75.30000 - 90.90000
Creaz2	LEVEL2	1	0	5.94000	0.00000	0.00000	5.48500	0.50750	4.47000 - 6.50000
Direct LDL	LEVEL2	1	0	60.52000	0.00000	0.00000	61.25000	5.87500	49.50000 - 73.00000
Glucose	LEVEL2	1	0	254.14000	0.00000	0.00000	265.00000	12.00000	241.00000 - 289.00000
K-C	LEVEL2	1	0	6.10000	0.00000	0.00000	5.93000	0.28500	5.36000 - 6.50000
Na-C	LEVEL2	1	0	126.20000	0.00000	0.00000	122.50000	3.75000	115.00000 - 130.00000
Phosphorus	LEVEL2	3	0	7.46333	0.11719	1.57020	7.05000	0.36500	6.32000 - 7.78000
TPRO2	LEVEL2	1	0	4.26000	0.00000	0.00000	3.86500	0.33250	3.20000 - 4.53000
Ultra HDL	LEVEL2	1	0	25.36000	0.00000	0.00000	25.60000	3.05000	19.50000 - 31.70000
Chol2	LEVEL2	1	0	104.73000	0.00000	0.00000	97.90000	6.05000	85.80000 - 110.00000
AlbBCG2	LEVEL2	1	0	2.67000	0.00000	0.00000	2.51500	0.16250	2.19000 - 2.84000
Billi Total	LEVEL2	1	0	4.76000	0.00000	0.00000	4.20000	0.40000	3.40000 - 5.00000
Billi D	LEVEL2	1	0	1.63000	0.00000	0.00000	1.87000	0.24000	1.39000 - 2.35000
Uric2	LEVEL2	1	0	9.43000	0.00000	0.00000	9.45000	0.62500	8.20000 - 10.70000
UIBC	LEVEL2	1	0	122.00000	0.00000	0.00000	101.00000	14.50000	72.00000 - 130.00000
ALT2	LEVEL2	1	0	89.87000	0.00000	0.00000	101.05000	9.47500	82.10000 - 120.00000
AST2	LEVEL2	1	0	245.00000	0.00000	0.00000	231.50000	31.75000	168.00000 - 295.00000
AlKP2	LEVEL2	1	0	424.00000	0.00000	0.00000	425.00000	31.00000	363.00000 - 487.00000
GGT2	LEVEL2	1	0	168.00000	0.00000	0.00000	165.88000	9.04000	147.80000 - 183.96000

Amankar

QC SUMMARY REPORT

SN: SCM23474

Software Version: 3.5.1

Operator ID: Admin

Site:

System Name: Ainity ci-series

Assay	Level	N	Flagged	Actual Mean	Actual SD	Actual %CV	Expected Mean	Expected SD	Expected Range
Iron2	LEVEL2	2	0	68.50000	0.70711	1.03227	69.95000	4.02500	61.90000 - 78.00000
Trig2	LEVEL2	1	0	86.00000	0.00000	0.00000	79.75000	7.27500	65.20000 - 94.30000

***** End of Report *****

Signature: _____ Date: _____

