

Erba Lachema s.r.o.



Brno, 10.11.2016

To,
Redcliffe Lifetech Pvt Ltd.
Thane Mumbai.

Instrument Name: ERBA LAURA
Instrument Serial no: 1137-1-202212

We Erba Lachema SRO with registered office in Czech Republic, Karasek 2219/1D, 621 00 BMO, represented by its Production manager Veronika Pleskova here by confirm as Producer of LAURA instrument that this device is firmly calibrated by the manufacturer and does not require any regular calibration by the user, correct calibration of instrument is controlled by grey strips provided initially together with the instrument.

P. S.
A handwritten signature in blue ink, appearing to be "P. S.", followed by a larger, more complex signature in blue ink.

RNDr. Pavel Suchan Ph.D.

Global Product Manager, Urine Analysis

Erba Lachema s.r.o.

Karásek 1d, Brno 62100, Czech Republic

Erba Lachema s.r.o.
Karásek 1d, 621 00 Brno
DIČ: CZ26918846 -10-



DEKAPHAN LAURA

Seq.No: 0074

URINORM P

26.03.2024

04:59 PM

.....

*BLD	10	Ery/u1	1+
*BIL	3	mg/dl	
*UBG	1	mg/dl	
*KET	52	mg/dl	
*GLU	500	mg/dl	2+
*PRO	100	mg/dl	2+
pH	9		
*NIT	POS		
*LEU	75	Leu/u1	1+
SG	1.000		

QC TEST: OK
<Admin>

DEKAPHAN LAURA

Seq.No: 0073

URINORM N

26.03.2024

04:57 PM

.....

BLD	NEG
BIL	NEG
UBG	NORM
KET	NEG
GLU	NEG
PRO	NEG
pH	6
NIT	NEG
LEU	NEG
SG	1.025

Amankaw

QC TEST: OK
<Admin>