


ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SECTOR-10, NOIDA - 201301 (U.P.)

Mob. : +91-9891121066, 7982584097 • E-mail : advance.cal@gmail.com



Calibration Certificate

CERTIFICATE NO.	: 240302A/N-117	Page	: 1 of 1
CALIBRATED ON.	: 04 / 03 / 2024	Ref. Std.	: IS:12508:1988
CALIBRATION DUE.	: 03 / 03 / 2025	C.P. No.	: ACS/CP/10
SRF No. & Dt. OF RECPT	: 240302A/N Dt. 02.03.2024	Phy. Inspection	: O.K.
ULR No. & ISSUE DATE	: CC342724000000290F Dt. 07.03.2024	Cal. At	: Site / Lab
CALIBRATED FOR	: M/s Redcliffe Lifetech Pvt. Ltd. Excel Laboratory, 3 rd Floor, Arjun Tower, Opposite Rajawat Jeweller, Gokhle Road Naupada, Thane(W) Mumbai - 400602		
ITEM CALIBRATED	: Centrifuge	Make	: Labtop
		Range	: (0 to 4000) RPM
		L. Count	: 1 RPM
		ID No.	: RCL/CP/Cent-01
		Location	: Clinical Pathology
		S. No.	: 16075940
ENVIRONMENT CONDITION:	Temp. { 30°C + 10°C} R.H. {30% to 80%}		
STANDARD INSTRUMENTS USED :			
INSTT. NAME	ID No. / S. No.	TRACEABILITY	CERTIFICATE No.
DUE DATE			
Tachometer	ACS/DTM/01/ 211-0155V	Godrej	G-230908-1-1
			08/09/2024
RESULTS	: Mechanical Discipline		
* U. C. INSTT. READS (RPM)	Standard Measured (RPM)	Error (RPM)	
1000	1001.2	- 1.2	
2000	2002.2	- 2.2	
3000	3003.4	- 3.4	
4000	4004.7	- 4.7	
(average of readings)			

REMARKS : The Expanded uncertainty of measurement at 95% confidence level is ± 3.90 rpm of reading at coverage Factor $k=1.96$.

*U.C. : Under Calibration.

- Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.
- The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
- This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
- The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
- Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use.
- Confirmity statement not provided since not required.

Prepared by



....End of Calibration Certificate....