


# ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SECTOR-10, NOIDA - 201301 (U.P.)

Mob. : +91-9891121066, 7982584097 • E-mail : advance.cal@gmail.com



## Calibration Certificate

<b>CERTIFICATE NO.</b>	: 240302A/N-119	<b>Page</b>	: 1 of 1	
<b>CALIBRATED ON.</b>	: 04 / 03 / 2024	<b>Ref. Std.</b>	: IS:12508:1988	
<b>CALIBRATION DUE.</b>	: 03 / 03 / 2025	<b>C.P. No.</b>	: ACS/CP/10	
<b>SRF No. &amp; Dt. OF RECPT</b>	: 240302A/N Dt. 02.03.2024	<b>Phy. Inspection</b>	: O.K.	
<b>ULR No. &amp; ISSUE DATE</b>	: CC342724000000292F Dt. 07.03.2024	<b>Cal. At</b>	: Site / Lab	
<b>CALIBRATED FOR</b>	: M/s Redcliffe Lifetech Pvt. Ltd. Excel Laboratory, 3 <sup>rd</sup> Floor, Arjun Tower, Opposite Rajawat Jeweller, Gokhle Road Naupada, Thane(W) Mumbai - 400602			
<b>ITEM CALIBRATED</b>	: Centrifuge			
	<b>Make</b>	: Remi	<b>Range</b>	: (0 to 6000) RPM
	<b>L. Count</b>	: 1 RPM	<b>ID No.</b>	: RCL/BIO/Cent-03
	<b>Location</b>	: Clinical Biochemistry	<b>S. No.</b>	: ZHKN-42556
<b>ENVIRONMENT CONDITION:</b>	Temp. { 30°C ± 10°C} R.H. {30% to 80%}			
<b>STANDARD INSTRUMENTS USED :</b>				
<b>INSTT. NAME</b>	<b>ID No. / S. No.</b>	<b>TRACEABILITY</b>	<b>CERTIFICATE No.</b>	<b>DUE DATE</b>
Tachometer	ACS/DTM/01/ 211-0155V	Godrej	G-230908-1-1	08/09/2024
<b>RESULTS</b>	: Mechanical Discipline			
	<b>* U. C. INSTT. READS (RPM)</b>	<b>Standard Measured (RPM)</b>	<b>Error (RPM)</b>	
	1000	1001.1	- 1.1	
	2000	2002.4	- 2.4	
	4000	4005.0	- 5.0	
	6000	6006.3	- 6.3	
	(average of readings)			

### REMARKS

: The Expanded uncertainty of measurement at 95% confidence level is ± 4.68 rpm of reading at coverage Factor k=1.96.

\*U.C. : Under Calibration.

- Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.
- The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
- This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
- The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
- Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use.
- Confirmity statement not provided since not required.



Prepared by

....End of Calibration Certificate....

