


# ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SECTOR-10, NOIDA - 201301 (U.P.)

Mob. : +91-9891121066, 7982584097 • E-mail : advance.cal@gmail.com



## Calibration Certificate

<b>CERTIFICATE NO.</b> : 240302A/N-110	<b>Page</b> : 1 of 1			
<b>CALIBRATED ON.</b> : 02 / 03 / 2024	<b>Ref. Std.</b> : DKD R-5-1			
<b>CALIBRATION DUE.</b> : 01 / 03 / 2025	<b>C.P. No.</b> : ACS/CP/12			
<b>SRF No. &amp; Dt. OF RECPT</b> : 240302A/N Dt. 02.03.2024	<b>Phy. Inspection</b> : O.K.			
<b>ULR No. &amp; ISSUE DATE</b> : CC342724000000283F Dt. 07.03.2024	<b>Cal. At</b> : Site / Lab			
<b>CALIBRATED FOR</b> : M/s Redcliffe Lifetech Pvt. Ltd. Excel Laboratory, 3 <sup>rd</sup> Floor, Arjun Tower, Opposite Rajawat Jeweller, Gokhle Road Naupada, Thane(W) Mumbai - 400602				
<b>ITEM CALIBRATED</b> : Multi Thermometer				
Make : Thermocare	Range : (-50 to 300) °C			
L. Count : 0.1 °C	ID No. : RCL/BIO/D.Temp-01			
Location : Clinical Biochemistry	S.No. : N.A.			
<b>ENVIRONMENT CONDITION:</b> Temp. { 25°C ± 4°C } R.H. { 30% to 75% }				
<b>STANDARD INSTRUMENTS USED :</b>				
<b>INSTT. NAME</b>	<b>ID No. / S. No.</b>	<b>TRACEABILITY</b>	<b>CERTIFICATE No.</b>	<b>DUE DATE</b>
Dig. Thermometer With RTD Sensor	ACS/TIS/01/20C106914	BCL	IALTH/00409/23	09/10/2024
Digital Thermometer with Sensor	ACS/TIS/02/201550503	BCL	30070827	20/07/2024
<b>RESULTS</b> : ET Discipline				
* U. C. INSTT. READS (°C)	STANDARD READING (°C)	ERROR (°C)		
- 39.8	- 39.91	0.11		
0.3	0.01	0.38		
100.6	100.2	0.20		
299.6	299.8	- 0.10		
(average of readings)				

**REMARKS** : The Expanded uncertainty of measurement at 95% confidence level is ± 1.5°C at coverage Factor k=2.

\*U.C. : Under Calibration.

- Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.
- The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
- This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
- The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
- Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use.
- Confirmity statement not provided since not required.

Prepared by



....End of Calibration Certificate....

