


ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SECTOR-10, NOIDA - 201301 (U.P.)

Mob. : +91-9891121066, 7982584097 • E-mail : advance.cal@gmail.com



Calibration Certificate

CERTIFICATE NO. : 240302A/N-106	Page : 1 of 1
CALIBRATED ON. : 02 / 03 / 2024	Ref. Std. : DKD R-5-7
CALIBRATION DUE. : 01 / 03 / 2025	C.P. No. : ACS/CP/11
SRF No. & Dt. OF RECPT : 240302A/N Dt. 02.03.2024	Phy. Inspection : O.K.
ULR No. & ISSUE DATE : CC342724000000279F Dt. 07.03.2024	Cal. At : Site / Lab
CALIBRATED FOR : M/s Redcliffe Lifetech Pvt. Ltd. Excel Laboratory, 3 rd Floor, Arjun Tower, Opposite Rajawat Jeweller, Gokhle Road Naupada, Thane(W) Mumbai - 400602	
ITEM CALIBRATED : Digital Thermo Hygrometer	
Make : Digital Series	Range : (0 to 50)°C, (10-99)%
L. Count :	ID No. : RCL/HEM/HGM-01
Location : Hematology	S. No. : N.A
ENVIRONMENT CONDITION: Temp. { 25°C + 4°C } R.H. {30% to 75%}	

STANDARD INSTRUMENTS USED :				
INSTT. NAME	ID No. / S. No.	TRACEABILITY	CERTIFICATE No.	DUE DATE
RH & Temperature Sensor with Indicator	ACS/RHTIS/01/20365872	DVG	CC352323000001002F	14/07/2024

RESULTS : Thermal Discipline				
* U. C. INSTT. READS	STANDARD READING	ERROR	UNC(+)	
(°C)	15.2	14.89	0.40	2.1
	25.4	24.91	0.30	2.1
	40.5	39.98	0.37	2.1
(average of readings)				
(%)	31	29.92	0.09	2.5
	61	59.95	0.08	2.5
	92	89.88	1.11	2.5
(average of readings)				

REMARKS : The Expanded uncertainty of measurement at 95% confidence level is $\pm 2.1^\circ\text{C}$, 2.5% at coverage Factor $k=1.96$.

*U.C. : Under Calibration.

- Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.
- The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
- This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
- The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
- Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use.
- Confirmity statement not provided since not required.

Prepared by



....End of Calibration Certificate....