


ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SECTOR-10, NOIDA - 201301 (U.P.)

Mob. : +91-9891121066, 7982584097 • E-mail : advance.cal@gmail.com



Calibration Certificate

CERTIFICATE NO.	: 240302A/N-105	Page	: 1 of 1	
CALIBRATED ON.	: 02 / 03 / 2024	Ref. Std.	: ISO:8655:6	
CALIBRATION DUE.	: 01 / 03 / 2025	C.P. No.	: ACS/CP/08	
SRF No.& Dt.OF RECPT	: 240302A/N Dt. 02.03.2024	Phy. Inspection	: O.K.	
ULR No. & ISSUE DATE	: CC342724000000278F Dt. 07.03.2024	Cal. At	: Site / Lab	
CALIBRATED FOR	: M/s Redcliffe Lifetech Pvt. Ltd. Excel Laboratory, 3 rd Floor, Arjun Tower, Opposite Rajawat Jeweller, Gokhle Road Naupada, Thane(W) Mumbai - 400602			
ITEM CALIBRATED	: Micro Pipette			
	Make	: Thermo Scientific	Range	: 500 µl
	L. Count	: fix	ID No.	: RCL/BIO/PIP-05
	Location	: Clinical Biochemistry	S. No.	: NW04748
ENVIRONMENT CONDITION:	Temp. { 25°C ± 4°C } R.H. {30% to 75%}			
STANDARD INSTRUMENTS USED :				
INSTT. NAME	ID No. / S. No.	TRACEABILITY	CERTIFICATE No.	DUE DATE
Weighing Balance	ACS/DB/02/D449927064	BlueBox	BBT/691/JAN/24	04/01/2025
RESULTS : Mechanical Discipline				
Nominal Capacity (µl)	Graduation (µl)	Nominal Value (µl)	Volume at 20 °C (µl)	Error (µl)
500		500	500.37	(-) 0.37

REMARKS : The Expanded uncertainty of measurement at 95% confidence level is ± 2.2 µl at coverage factor k=2.

*U.C. : Under Calibration.

- Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.
- The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
- This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
- The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
- Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use.
- Confirmity statement not provided since not required.

Prepared by



Auth. Signatory
S.K. Sharma
CEO/TM

....End of Calibration Certificate....