

ASIAN TECHNOLOGY





CALIBRATION CERTIFICATE

Customer name	M/S: Rainbow Diagnostic	Service request No. & date	P/06/11/12/23&11/12/2023
And	kshudiram Square(Manjushree)	ULR no.	CC223923000006768 F
Address	Haldia, Dist -purba medinipur, pin-	Cert. No.	AT23000006768
	721602	Date of Receipt of DUC	11/12/2023
		Date of calibration	11/12/2023
		Date of issue	12/12/2023
		Suggested due date	10/12/2024

Instrument Details			
Instrument name	Micro Pipette	Sr. no.	YE186AL0017094
Make / Model No.	Dragon	Location	
Range / Size	100 μΙ	Accuracy	
Least Count	N/A	Visual Inspection	OK
I.D. No.		Calibration Performed at	Lab

Detail of reference standards & Major equipments used			
Equipment Name	A set of weight box	Digital Weighing Balance	
Make	WEIGHTRONICS .	AND	
Model / SR No.		GH-252/ 78002	
Certificate No.	TYCON/WB/12/274	TYCON/WB/02/23/327	1
Calibration Validity	25/12/2023	19/02/2024	
Calibration by	Tycon Engineering	Tycon Engineering	

Environmental	Temperature	(23±3) °C	Calibration Reference	IS:4162 (1985)
Condition	Relative Humidity	(50±20) %	Work Instruction	AT/WI-20M

<u>Cambration Results</u>			
Serial No.	Nominal Capacity in (µl)	Measured Capacity in (μl)	Uncertainty At 95% C.L. (coverage factor K=2)
01.	100.0	99.82	±0.10 μΙ

Remarks:

- (1) Standard equipment use for calibration are traceable to national/international standards.
- (2)The reported expended uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by coverage factork = 2 such that the coverage probability corresponds to approximately 95% (3) The above results are valid at the time of and under the stated conditions measurement.(4)This certificate is refers only to the particular item submitted for calibration. (5) Next calibration due date given as requested by the customer.

Calibrated By (Call bration Engg./TM) (PRASHANT) Form No.- QF-47

pproved By ERAJ TYAGI) Page No. 1 of 1

