

Instrument : Diasys Sys 200

Analyte	Standard Unit	Result Value	Mean	Z-Score	RMZ
X LDH	U/L	44.00	103.60	-7.22	-1.15
# Urea	mg/dl	23.40	21.67		0.04
@ Creatinine	mg/dl	0.80	0.90	0.19	-0.32
@ Bilirubin Total	mg/dl	0.30	0.23	0.90	0.80
@ Total Protein	g/dL	6.00	5.30	0.68	-0.40
@ Albumin	g/dL	3.60	3.10	0.82	0.85
@ Uric Acid	mg/dl	4.40	3.84	0.64	-0.51
Calcium	mg/dl	6.90	6.93	-0.53	-0.24
Phosphorous Inorganic	mg/dl	3.90	3.60	-0.40	-0.09
@ Cholesterol	mg/dl	143.60	135.40	1.23	0.25
@ Triglycerides	mg/dl	125.30	118.40	0.30	-0.25
@ HDL	mg/dl	35.09	33.61	0.87	-0.71
@ ALT	U/L	14.00	13.67	0.13	-0.13
@ AST	U/L	27.00	27.00	0.27	-0.06
@ ALP	U/L	41.00	41.00	-0.52	1.81
@ GGT	U/L	26.00	26.00	0.11	-0.26
@ Amylase	U/L	60.00	58.63	-0.09	-0.27
@ Lipase	U/L	74.50	--	-1.53	2.33

Instrument : ILILYTE

Analyte	Standard Unit	Result Value	Mean	Z-Score	RMZ
Chloride	mmol/L	80.40	79.25	0.08	0.71
Sodium	mmol/L	104.20	110.10	-0.66	-0.39
Potassium	mmol/L	3.55	3.67	0.13	-0.29