



Metropolis Healthcare Ltd., EQAS

KRM HEALTHCARE (1019)

Outlier And Analyte Summary Report

Outlier Details For Cycle No 230412 and Sample No 03

Report Date : 13/02/2024

Analyte	Instrument	Result Value	Standard Unit	Z-Score
* Platelet Count	BC 700	0.11	/ μ l	*
Legend @ : z score \leq 2.0 - Acceptable ! : 2.0 < z score < 3.0 - Warning X : z score \geq 3 - Unacceptable # : Not Evaluated ⌚ : Delayed Result Entry * : Not considered for evaluation.				

Problem Classification: _____

Corrective Action: _____

Reviewed by: _____

Dated: _____

Instrument : BC 700

Analyte	Standard Unit	Result Value	Mean	Z-Score	RMZ
* Platelet Count	1000/ μ L	0.11	--	-9.29 *	-0.59
@ Erythrocyte (RBC) Count	mill/cu.mm	2.05	--	-1.33	-0.77
@ Total Leucocytes (WBC) count	/c.mm	3790.00	--	-0.03	0.17
@ Haemoglobin	g/dL	6.10	--	0.92	1.36
@ MCV (Mean Corpuscular Volume)	fL	147.30	--	0.99	1.30
@ MCH (Mean Corpuscular Hb)	pg	29.90	--	1.07	0.77
@ MCHC (Mean Corpuscular Hb Concentration)	g/dL	20.30	--	-0.16	-0.59
@ Hematocrit	%	30.20	--	0.48	0.64
@ RDW CV%	%	21.00	--	-0.39	-0.85

Legend @ : Acceptable

! : 2.0 < z score < 3.0 - Warning


X : z score \geq 3 - Unacceptable

: Not Evaluated

⌚ : Delayed Result Entry

* : Not considered for evaluation.

Total Parameters	9
Not Evaluated Parameters	0
Evaluated Parameters	9
Outlier Parameters (X)	1
EQAS Score Haematology - CBC	88.89 %



Dr Puneet Kumar Nigam
PT coordinator & Technical Manager, MHL EQAS
Unit No. 409-416.
Commercial Building - 1A
Kohinoor Mall, Kirol Road, Kurla (W),
Mumbai - 400070



Metropolis Healthcare Ltd., EQAS

KRM HEALTHCARE (1019)

Outlier And Analyte Summary Report

Outlier Details For Cycle No 230412 and Sample No 04

Report Date : 13/02/2024

Analyte	Instrument	Result Value	Standard Unit	Z-Score
* Platelet Count	BC 700	0.21	/ μ l	*
Legend @ : z score \leq 2.0 - Acceptable ! : 2.0 < z score < 3.0 - Warning X : z score \geq 3 - Unacceptable # : Not Evaluated ⌚ : Delayed Result Entry * : Not considered for evaluation.				

Problem Classification: _____

Corrective Action: _____

Reviewed by: _____

Dated: _____

Instrument : BC 700

Analyte	Standard Unit	Result Value	Mean	Z-Score	RMZ
* Platelet Count	1000/ μ L	0.21	--	-8.83 *	-0.59
@ Erythrocyte (RBC) Count	mill/cu.mm	2.14	--	-1.07	-0.77
@ Total Leucocytes (WBC) count	/c.mm	8270.00	--	1.03	0.17
@ Haemoglobin	g/dL	7.00	--	1.16	1.36
@ MCV (Mean Corpuscular Volume)	fL	156.80	--	0.71	1.30
@ MCH (Mean Corpuscular Hb)	pg	32.60	--	0.67	0.77
@ MCHC (Mean Corpuscular Hb Concentration)	g/dL	20.80	--	-0.33	-0.59
@ Hematocrit	%	33.50	--	0.33	0.64
@ RDW CV%	%	21.30	--	-1.05	-0.85

Legend @ : Acceptable

! : 2.0 < z score < 3.0 - Warning


X : z score \geq 3 - Unacceptable

: Not Evaluated

⌚ : Delayed Result Entry

* : Not considered for evaluation.

Total Parameters	9
Not Evaluated Parameters	0
Evaluated Parameters	9
Outlier Parameters (X)	1
EQAS Score Haematology - CBC	88.89 %



Dr Puneet Kumar Nigam
PT coordinator & Technical Manager, MHL EQAS
Unit No. 409-416.
Commercial Building - 1A
Kohinoor Mall, Kirol Road, Kurla (W),
Mumbai - 400070



Metropolis Healthcare Ltd., EQAS

KRM HEALTHCARE (1019)

Outlier And Analyte Summary Report

Outlier Details For Cycle No 230412 and Sample No 03

Report Date : 13/02/2024

Analyte	Instrument	Result Value	Standard Unit
No Outlier			
Legend @ : Acceptable X : Unacceptable # : Not Evaluated ⌚ : Delayed Result Entry			

Problem Classification: _____

Corrective Action: _____

Reviewed by: _____


Dated: _____

Instrument : -

Analyte	Standard Unit	Result Value	Accepted Value
@ Diagnosis (Slide)	-	P. falciparum and P. vivax seen	P. falciparum and P. vivax seen

Legend @ : Acceptable
 @ : Acceptable
 X : Unacceptable
 # : Not Evaluated
 ⌚ : Delayed Result Entry

Total Parameters	1
Not Evaluated Parameters	0
Evaluated Parameters	1
Outlier Parameters (X)	0
EQAS Score Haematology- Peripheral smear diagnosis	100.00 %



Dr Puneet Kumar Nigam
 PT coordinator & Technical Manager, MHL EQAS
 Unit No. 409-416.
 Commercial Building - 1A
 Kohinoor Mall, Kiroi Road, Kurla (W),
 Mumbai - 400070



Metropolis Healthcare Ltd., EQAS

KRM HEALTHCARE (1019)

Outlier And Analyte Summary Report

Outlier Details For Cycle No 230412 and Sample No 01

Report Date : 13/02/2024

Analyte	Instrument	Result Value	Standard Unit	Z-Score
No Outlier				
Legend @ : z score \leq 2.0 - Acceptable ! : 2.0 < z score < 3.0 - Warning X : z score \geq 3 - Unacceptable # : Not Evaluated ⌚ : Delayed Result Entry * : Not considered for evaluation.				

Problem Classification: _____

Corrective Action: _____


Reviewed by: _____

Dated: _____

Instrument : S8 Hemoglobin Analyzer

Analyte	Standard Unit	Result Value	Mean	Z-Score	RMZ
@ HbA1c	%	4.70	4.70	-0.52	-0.28

Legend @ : Acceptable	Total Parameters	1
! : 2.0 < z score < 3.0 - Warning	Not Evaluated Parameters	0
X : z score ≥3 - Unacceptable	Evaluated Parameters	1
# : Not Evaluated	Outlier Parameters (X)	0
⌚ : Delayed Result Entry	EQAS Score HbA1c	100.00 %
* : Not considered for evaluation.		



Dr Puneet Kumar Nigam
PT coordinator & Technical Manager, MHL EQAS
Unit No. 409-416.
Commercial Building - 1A
Kohinoor Mall, Kiroil Road, Kurla (W),
Mumbai - 400070



Metropolis Healthcare Ltd., EQAS

KRM HEALTHCARE (1019)

Outlier And Analyte Summary Report

Outlier Details For Cycle No 230412 and Sample No 02

Report Date : 13/02/2024

Analyte	Instrument	Result Value	Standard Unit	Z-Score
No Outlier				
Legend @ : z score \leq 2.0 - Acceptable ! : 2.0 < z score < 3.0 - Warning X : z score \geq 3 - Unacceptable # : Not Evaluated ⌚ : Delayed Result Entry * : Not considered for evaluation.				

Problem Classification: _____

Corrective Action: _____

Reviewed by: _____ **Dated:** _____

Instrument : S8 Hemoglobin Analyzer

Analyte	Standard Unit	Result Value	Mean	Z-Score	RMZ
@ HbA1c	%	5.10	5.05	-0.35	-0.28

Legend @ : Acceptable	Total Parameters	1
! : 2.0 < z score < 3.0 - Warning	Not Evaluated Parameters	0
X : z score ≥3 - Unacceptable	Evaluated Parameters	1
# : Not Evaluated	Outlier Parameters (X)	0
⌚ : Delayed Result Entry	EQAS Score HbA1c	100.00 %
* : Not considered for evaluation.		



Dr Puneet Kumar Nigam
PT coordinator & Technical Manager, MHL EQAS
Unit No. 409-416.
Commercial Building - 1A
Kohinoor Mall, Kiroli Road, Kurla (W),
Mumbai - 400070



Metropolis Healthcare Ltd., EQAS

KRM HEALTHCARE (1019)

Outlier And Analyte Summary Report

Outlier Details For Cycle No 230412 and Sample No 03

Report Date : 13/02/2024

Analyte	Instrument	Result Value	Standard Unit
# Lymphocyte.	-	20-40%	%
Legend @ : Acceptable X : Unacceptable # : Not Evaluated ⌚ Delayed Result Entry			

Problem Classification: _____

Corrective Action: _____

Reviewed by: _____


Dated: _____

Instrument : -

Analyte	Standard Unit	Result Value	Accepted Value
# Lymphocyte.	-	20-40%	
@ Neutrophil.	-	40-70%	40-70%
@ Eosinophil.	-	<6%	<6%

Legend @ : Acceptable
 @ : Acceptable
 X : Unacceptable
 # : Not Evaluated
 Ⓞ : Delayed Result Entry

Total Parameters	3
Not Evaluated Parameters	1
Evaluated Parameters	2
Outlier Parameters (X)	0
EQAS Score Haematology-PS Slide Count Qualitative	100.00 %



Dr Puneet Kumar Nigam
 PT coordinator & Technical Manager, MHL EQAS
 Unit No. 409-416.
 Commercial Building - 1A
 Kohinoor Mall, Kiroi Road, Kurla (W),
 Mumbai - 400070



Metropolis Healthcare Ltd., EQAS

KRM HEALTHCARE (1019)

Outlier And Analyte Summary Report

Outlier Details For Cycle No 230310 and Sample No 04

Report Date : 19/12/2023

Analyte	Instrument	Result Value	Standard Unit	Z-Score
No Outlier				
Legend @ : z score \leq 2.0 - Acceptable ! : 2.0 < z score < 3.0 - Warning X : z score \geq 3 - Unacceptable # : Not Evaluated ⌚ : Delayed Result Entry * : Not considered for evaluation.				

Problem Classification: _____

Corrective Action: _____

Reviewed by: _____

Dated: _____

Instrument : BC 700

Analyte	Standard Unit	Result Value	Mean	Z-Score	RMZ
@ Erythrocyte (RBC) Count	mill/cu.mm	1.45	--	-0.23	-0.34
@ Total Leucocytes (WBC) count	/c.mm	5000.00	--	0.69	-0.17
@ Haemoglobin	g/dL	4.50	--	1.99	1.68
@ MCV (Mean Corpuscular Volume)	fL	153.00	--	1.47	1.76
@ MCH (Mean Corpuscular Hb)	pg	30.80	--	0.77	0.66
@ MCHC (Mean Corpuscular Hb Concentration)	g/dL	20.10	--	-0.97	-0.93
@ Platelet Count	1000/ μ L	135.00	--	-0.56	-0.59
@ Hematocrit	%	22.20	--	0.90	0.87
@ RDW CV%	%	29.80	--	-0.86	-0.99

Legend @ : Acceptable

! : $2.0 < z \text{ score} < 3.0$ - Warning


X : $z \text{ score} \geq 3$ - Unacceptable

: Not Evaluated

⌚ : Delayed Result Entry

* : Not considered for evaluation.

Total Parameters	9
Not Evaluated Parameters	0
Evaluated Parameters	9
Outlier Parameters (X)	0
EQAS Score Haematology - CBC	100.00 %



Dr Puneet Kumar Nigam
PT coordinator & Technical Manager, MHL EQAS

Unit No. 409-416.

Commercial Building - 1A

Kohinoor Mall, Kirol Road, Kurla (W),

Mumbai - 400070



Metropolis Healthcare Ltd., EQAS

KRM HEALTHCARE (1019)

Outlier And Analyte Summary Report

Outlier Details For Cycle No 230412 and Sample No 03

Report Date : 13/02/2024

Analyte	Instrument	Result Value	Standard Unit
# Lymphocyte.	-	20-40%	%
Legend @ : Acceptable X : Unacceptable # : Not Evaluated ⌚ Delayed Result Entry			

Problem Classification: _____

Corrective Action: _____

Reviewed by: _____


Dated: _____

Instrument : -

Analyte	Standard Unit	Result Value	Accepted Value
# Lymphocyte.	-	20-40%	
@ Neutrophil.	-	40-70%	40-70%
@ Eosinophil.	-	<6%	<6%

Legend @ : Acceptable
 @ : Acceptable
 X : Unacceptable
 # : Not Evaluated
 ⌚ : Delayed Result Entry

Total Parameters	3
Not Evaluated Parameters	1
Evaluated Parameters	2
Outlier Parameters (X)	0
EQAS Score Haematology-PS Slide Count Qualitative	100.00 %



Dr Puneet Kumar Nigam
 PT coordinator & Technical Manager, MHL EQAS
 Unit No. 409-416.
 Commercial Building - 1A
 Kohinoor Mall, Kiroi Road, Kurla (W),
 Mumbai - 400070