



Metropolis Healthcare Ltd., EQAS

KRM HEALTHCARE (1019)

Outlier And Analyte Summary Report

Outlier Details For Cycle No 230310 and Sample No 01

Report Date : 19/12/2023

Analyte	Instrument	Result Value	Standard Unit
X Blood (urine)	Manual	Trace	-
Legend @ : Acceptable X : Unacceptable # : Not Evaluated ⌚ : Delayed Result Entry			

Problem Classification: _____

Corrective Action: _____

Reviewed by: _____

Dated: _____

Instrument : Manual

Analyte	Standard Unit	Result Value	Accepted Value
@ pH	-	6.5	6,6.5,7
@ Specific Gravity	-	1.020	1.015,1.020,1.025
@ Glucose (urine)	-	2+	+2,+3,+4
@ Protein (Urine)	-	2+	+1,+2,+3
@ Ketone (urine)	-	Negative	Neg
X Blood (urine)	-	Trace	+1,+2,+3
@ Bilirubin (urine)	-	Negative	Neg
@ Urobilinogen (urine)	-	Normal	Nor
@ Nitrite (urine)	-	Present	Pre

Legend @ : Acceptable	<table border="1"> <tbody> <tr> <td>Total Parameters</td> <td>9</td> </tr> <tr> <td>Not Evaluated Parameters</td> <td>0</td> </tr> <tr> <td>Evaluated Parameters</td> <td>9</td> </tr> <tr> <td>Outlier Parameters (X)</td> <td>1</td> </tr> <tr> <td>EQAS Score Clinical Path - Urine Routine Chemistry</td> <td>88.89 %</td> </tr> </tbody> </table>	Total Parameters	9	Not Evaluated Parameters	0	Evaluated Parameters	9	Outlier Parameters (X)	1	EQAS Score Clinical Path - Urine Routine Chemistry	88.89 %
Total Parameters		9									
Not Evaluated Parameters		0									
Evaluated Parameters		9									
Outlier Parameters (X)		1									
EQAS Score Clinical Path - Urine Routine Chemistry	88.89 %										
@ : Acceptable											
X : Unacceptable											
# : Not Evaluated											
⌚ : Delayed Result Entry											



Dr Puneet Kumar Nigam
PT coordinator & Technical Manager, MHL EQAS
Unit No. 409-416.
Commercial Building - 1A
Kohinoor Mall, Kirol Road, Kurla (W),
Mumbai - 400070