



Metropolis Healthcare Ltd., EQAS

KRM HEALTHCARE (1019)

Outlier And Analyte Summary Report

Outlier Details For Cycle No 230412 and Sample No 01

Report Date : 13/02/2024

| Analyte | Instrument | Result Value | Standard Unit |
|--|-------------------|--------------|---------------|
| X C-Reactive Protein Qualitative | VITROS 5600 (Wet) | Normal | mg/L |
| Legend @ : Acceptable X : Unacceptable # : Not Evaluated ⌚ : Delayed Result Entry | | | |

Problem Classification: _____

Corrective Action: _____

Reviewed by: _____

Dated: _____

Instrument : VITROS 5600 (Wet)


| Analyte | Standard Unit | Result Value | Accepted Value |
|----------------------------------|---------------|--------------|----------------|
| X C-Reactive Protein Qualitative | mg/L | Normal | Abn |

Instrument : Manual

| Analyte | Standard Unit | Result Value | Accepted Value |
|---|---------------|------------------|------------------|
| @ HIV Ag + Ab | S/CO | Non-reactive | Non-reactive |
| @ HBsAg | IU/mL | Non-reactive | Non-reactive |
| @ HCV | IU/mL | Non-reactive | Non-reactive |
| @ RPR (Syphilis Serology) | - | Non-reactive | Non-reactive |
| @ Dengue NS 1 | RU/ml | Negative | Negative |
| @ Typhidot IgM | - | Negative | Negative |
| @ Widal test - Typhi O antigen (TO) | - | No Agglutination | No Agglutination |
| @ Widal test - Typhi H antigen (TH) | - | No Agglutination | No Agglutination |
| @ Widal test - ParaTyphi A-H antigen (AH) | - | No Agglutination | No Agglutination |
| @ Widal test - ParaTyphi B-H antigen (BH) | - | No Agglutination | No Agglutination |
| @ Dengue IgM | RU/ml | Negative | Negative |
| @ Dengue IgG | RU/ml | Negative | Negative |

Legend @ : Acceptable
 @ : Acceptable
 X : Unacceptable
 # : Not Evaluated
 ⊙ : Delayed Result Entry

| | |
|---------------------------------|---------|
| Total Parameters | 13 |
| Not Evaluated Parameters | 0 |
| Evaluated Parameters | 13 |
| Outlier Parameters (X) | 1 |
| EQAS Score Immunoassay Extended | 92.31 % |



Dr Puneet Kumar Nigam
 PT coordinator & Technical Manager, MHL EQAS
 Unit No. 409-416.
 Commercial Building - 1A
 Kohinoor Mall, Kirol Road, Kurla (W),
 Mumbai - 400070



Metropolis Healthcare Ltd., EQAS

KRM HEALTHCARE (1019)

Outlier And Analyte Summary Report

Outlier Details For Cycle No 230412 and Sample No 02

Report Date : 13/02/2024

| Analyte | Instrument | Result Value | Standard Unit |
|--|-------------------|--------------|---------------|
| X C-Reactive Protein Qualitative | VITROS 5600 (Wet) | Normal | mg/L |
| # HCV | Manual | Non-reactive | - |
| Legend @ : Acceptable X : Unacceptable # : Not Evaluated ⌚ Delayed Result Entry | | | |

Problem Classification: _____

Corrective Action: _____

Reviewed by: _____

Dated: _____

Instrument : VITROS 5600 (Wet)


| Analyte | Standard Unit | Result Value | Accepted Value |
|----------------------------------|---------------|--------------|----------------|
| X C-Reactive Protein Qualitative | mg/L | Normal | Abn |

Instrument : Manual

| Analyte | Standard Unit | Result Value | Accepted Value |
|---|---------------|-------------------------|-------------------------|
| @ HIV Ag + Ab | S/CO | HIV-1/HIV-2 Ab Reactive | HIV-1/HIV-2 Ab Reactive |
| @ HBsAg | IU/mL | Reactive | Reactive |
| # HCV | IU/mL | Non-reactive | |
| @ RPR (Syphilis Serology) | - | Non-reactive | Non-reactive |
| @ Dengue NS 1 | RU/ml | Negative | Negative |
| @ Typhidot IgM | - | Negative | Negative |
| @ Widal test - Typhi O antigen (TO) | - | No Agglutination | No Agglutination |
| @ Widal test - Typhi H antigen (TH) | - | No Agglutination | No Agglutination |
| @ Widal test - ParaTyphi A-H antigen (AH) | - | No Agglutination | No Agglutination |
| @ Widal test - ParaTyphi B-H antigen (BH) | - | No Agglutination | No Agglutination |
| @ Dengue IgM | RU/ml | Negative | Negative |
| @ Dengue IgG | RU/ml | Negative | Negative |

Legend @ : Acceptable
 @ : Acceptable
 X : Unacceptable
 # : Not Evaluated
 ⊙ : Delayed Result Entry

| | |
|---------------------------------|---------|
| Total Parameters | 13 |
| Not Evaluated Parameters | 1 |
| Evaluated Parameters | 12 |
| Outlier Parameters (X) | 1 |
| EQAS Score Immunoassay Extended | 91.67 % |



Dr Puneet Kumar Nigam
 PT coordinator & Technical Manager, MHL EQAS
 Unit No. 409-416.
 Commercial Building - 1A
 Kohinoor Mall, Kirol Road, Kurla (W),
 Mumbai - 400070



Metropolis Healthcare Ltd., EQAS

KRM HEALTHCARE (1019)

Outlier And Analyte Summary Report

Outlier Details For Cycle No 230412 and Sample No 03

Report Date : 13/02/2024

| Analyte | Instrument | Result Value | Standard Unit |
|----------------------------------|-------------------|--------------|---------------|
| # HIV Ag + Ab | Manual | Non-reactive | - |
| # HBsAg | Manual | Non-reactive | - |
| # HCV | Manual | Reactive | - |
| X C-Reactive Protein Qualitative | VITROS 5600 (Wet) | Normal | mg/L |

Legend @ : Acceptable
X : Unacceptable
: Not Evaluated
⌚ Delayed Result Entry

Problem Classification: _____

Corrective Action: _____

Reviewed by: _____

Dated: _____

Instrument : Manual


| Analyte | Standard Unit | Result Value | Accepted Value |
|---------------|---------------|--------------|----------------|
| # HIV Ag + Ab | S/CO | Non-reactive | |
| # HBsAg | IU/mL | Non-reactive | |
| # HCV | IU/mL | Reactive | |

Instrument : VITROS 5600 (Wet)

| Analyte | Standard Unit | Result Value | Accepted Value |
|----------------------------------|---------------|--------------|----------------|
| X C-Reactive Protein Qualitative | mg/L | Normal | Abnormal |

Legend @ : Acceptable
 @ : Acceptable
 X : Unacceptable
 # : Not Evaluated
 ⌚ : Delayed Result Entry

| | |
|---------------------------------|--------|
| Total Parameters | 4 |
| Not Evaluated Parameters | 3 |
| Evaluated Parameters | 1 |
| Outlier Parameters (X) | 1 |
| EQAS Score Immunoassay Extended | 0.00 % |



Dr Puneet Kumar Nigam
 PT coordinator & Technical Manager, MHL EQAS
 Unit No. 409-416.
 Commercial Building - 1A
 Kohinoor Mall, Kirool Road, Kurla (W),
 Mumbai - 400070



Metropolis Healthcare Ltd., EQAS

KRM HEALTHCARE (1019)

Outlier And Analyte Summary Report

Outlier Details For Cycle No 230310 and Sample No 01

Report Date : 19/12/2023

| Analyte | Instrument | Result Value | Standard Unit |
|--|-------------------|--------------|---------------|
| # C-Reactive Protein Qualitative | VITROS 5600 (Wet) | Normal | mg/L |
| Legend @ : Acceptable X : Unacceptable # : Not Evaluated ⌚ : Delayed Result Entry | | | |

Problem Classification: _____

Corrective Action: _____

Reviewed by: _____

Dated: _____

Instrument : VITROS 5600 (Wet)


| Analyte | Standard Unit | Result Value | Accepted Value |
|----------------------------------|---------------|--------------|----------------|
| # C-Reactive Protein Qualitative | mg/L | Normal | |

Instrument : Manual

| Analyte | Standard Unit | Result Value | Accepted Value |
|---|---------------|------------------|----------------|
| @ Widal test - Typhi O antigen (TO) | - | No Agglutination | 1:80,NoA |
| @ Widal test - Typhi H antigen (TH) | - | No Agglutination | 1:80,NoA |
| @ Widal test - ParaTyphi A-H antigen (AH) | - | No Agglutination | 1:80,NoA |
| @ Widal test - ParaTyphi B-H antigen (BH) | - | No Agglutination | 1:80,NoA |
| @ HIV Ag + Ab | S/CO | Non-reactive | Non-reactive |
| @ HBsAg | IU/mL | Non-reactive | Non-reactive |
| @ HCV | IU/mL | Non-reactive | Non-reactive |
| @ RPR (Syphilis Serology) | - | Non-reactive | Non-reactive |
| @ Dengue NS 1 | RU/ml | Negative | Negative |
| @ Typhidot IgM | - | Negative | Negative |

Legend @ : Acceptable
 @ : Acceptable
 X : Unacceptable
 # : Not Evaluated
 ⌚ : Delayed Result Entry

| | |
|---------------------------------|----------|
| Total Parameters | 11 |
| Not Evaluated Parameters | 1 |
| Evaluated Parameters | 10 |
| Outlier Parameters (X) | 0 |
| EQAS Score Immunoassay Extended | 100.00 % |



Dr Puneet Kumar Nigam
 PT coordinator & Technical Manager, MHL EQAS
 Unit No. 409-416.
 Commercial Building - 1A
 Kohinoor Mall, Kirol Road, Kurla (W),
 Mumbai - 400070