



Metropolis Healthcare Ltd., EQAS

KRM HEALTHCARE (1019)

Outlier And Analyte Summary Report

Outlier Details For Cycle No 230412 and Sample No 01

Report Date : 13/02/2024

Analyte	Instrument	Result Value	Standard Unit
No Outlier			
Legend @ : Acceptable X : Unacceptable # : Not Evaluated ⌚ : Delayed Result Entry			

Problem Classification: _____

Corrective Action: _____

Reviewed by: _____


Dated: _____

Instrument : Manual

Analyte	Standard Unit	Result Value	Accepted Value
@ pH	-	6	5,5,5,6
@ Specific Gravity	-	1.025	1.020,1.025,1.030
@ Glucose (urine)	-	3+	+2,+3,+4
@ Protein (Urine)	-	2+	+1,+2,+3
@ Ketone (urine)	-	Negative	Neg
@ Blood (urine)	-	1+	+1,+2,+3
@ Bilirubin (urine)	-	Negative	Neg
@ Urobilinogen (urine)	-	Normal	Nor
@ Nitrite (urine)	-	Present	Pre

Instrument : -

Analyte	Standard Unit	Result Value	Accepted Value
@ Urine Pregnancy Test	-	Positive	Pos
Legend		Total Parameters	10
@ : Acceptable		Not Evaluated Parameters	0
@ : Acceptable		Evaluated Parameters	10
X : Unacceptable		Outlier Parameters (X)	0
# : Not Evaluated		EQAS Score Clinical Path - Urine Routine Chemistry	100.00 %
⊙ : Delayed Result Entry			



Dr Puneet Kumar Nigam
PT coordinator & Technical Manager, MHL EQAS
Unit No. 409-416.
Commercial Building - 1A
Kohinoor Mall, Kiroil Road, Kurla (W),
Mumbai - 400070