

- Test Parameter(F3)
- Profiles / Calc(F4)
- QC/Calibration(F5)
- Consumables(F6)
- Status Monitor(F7)
- Search(F8)
- Reports(F9)
- Master
- Utility(F11)
- Service Check
- Maintenance(F12)
- Settings
- Archive
- Wake Up
- Shut Down

No.	Test	Curve Type	Type	Consumable	Lot #	Conc.	Abs	Factor	Calib Exp. (Days)	Acceptable Limit
1	GLU	Linear	BLANK	NORMAL SALINE	123	0.000	0.0433	0	0	15
		Linear	STD1	MULTICAL	2206133	200.000	0.7274	292.3649	0	15
2	UREA	Linear	BLANK	NORMAL SALINE	123	0.000	-0.0074	0	0	15
		Linear	STD1	MULTICAL	2206133	100.000	-0.1384	-763.3588	0	15
3	CRENZ	Linear	BLANK	NORMAL SALINE	123	0.000	0.0195	0	0	15
		Linear	STD1	MULTICAL	2206133	3.840	0.0862	57.5712	0	15
4	UA	Linear	BLANK	NORMAL SALINE	123	0.000	0.0025	0	0	15
		Linear	STD1	MULTICAL	2206133	5.430	0.0765	73.3784	0	15
5	CHOL	Linear	BLANK	NORMAL SALINE	123	0.000	0.0993	0	0	15
		Linear	STD1	MULTICAL	2206133	155.000	0.3365	656.2235	0	15
6	TRIG	Linear	BLANK	NORMAL SALINE	123	0.000	0.0258	0	0	15
		Linear	STD1	MULTICAL	2206133	140.000	0.2365	664.4518	0	15
7	HDL	Linear	BLANK	NORMAL SALINE	123	0.000	0.0660	0	0	15
		Linear	STD1	HDL LDL CAL	S032302	62.000	0.1207	540.5405	0	15
8	LDL	Linear	BLANK	NORMAL SALINE	123	0.000	0.0009	0	0	15
		Linear	STD1	HDL LDL CAL	1807047	122.000	0.1624	672.1763	0	15
9	PRO	Linear	BLANK	NORMAL SALINE	123	0.000	0.0802	0	0	15
		Linear	STD1	MULTICAL	2206133	5.840	0.3479	18.827	0	15
10	ALB	Linear	BLANK	NORMAL SALINE	123	0.000	0.0890	0	0	15
		Linear	STD1	MULTICAL	2206133	3.520	0.5911	7.0245	0	15
11	ALPU	Linear	BLANK	NORMAL SALINE	123	0.000	0.0004	0	0	15
		Linear	STD1	MULTICAL	2206133	262.000	0.0550	4798.5347	0	15
12	BIT	Linear	BLANK	NORMAL SALINE	123	0.000	0.0003	0	0	15
		Linear	STD1	MULTICAL	2206133	4.300	0.1719	25.0583	0	15
13	BID	Linear	BLANK	NORMAL SALINE	123	0.000	0.0005	0	0	15
		Linear	STD1	MULTICAL	2206133	2.290	0.0996	23.1781	0	15
14	SGOTD	Linear	BLANK	NORMAL SALINE	123	0.000	-0.0010	0	0	15
		Linear	STD1	MULTICAL	2206133	100.000	-0.0368	-2645.5027	0	15

- Test Parameter(F3)
- Profiles / Calc(F4)
- QC/Calibration(F5)
- Consumables(F6)
- Status Monitor(F7)
- Search(F8)
- Reports(F9)
- Master
- Utility(F11)
- Service Check
- Maintenance(F12)
- Settings
- Archive
- Wake Up
- Shut Down

No.	Test	Curve Type	Type	Consumable	Lot #	Conc.	Abs	Factor	Calib Exp. (Days)	Acceptable Limit
13	BID	Linear	BLANK	NORMAL SALINE	123	0.000	0.0008	0	0	15
		Linear	STD1	MULTICAL	2206133	2.290	0.0996	23.1781	0	15
14	SGOTD	Linear	BLANK	NORMAL SALINE	123	0.000	-0.0010	0	0	15
		Linear	STD1	MULTICAL	2206133	100.000	-0.0388	-2645.5027	0	15
15	SGPTD	Linear	BLANK	NORMAL SALINE	123	0.000	-0.0010	0	0	15
		Linear	STD1	MULTICAL	2206133	88.800	-0.0339	-2699.0881	0	15
16	GGT	Linear	BLANK	NORMAL SALINE	123	0.000	0.0013	0	0	15
		Linear	STD1	MULTICAL	2206133	111.000	0.0610	1859.2965	0	15
17	CA	Linear	BLANK	NORMAL SALINE	123	0.000	0.2090	0	0	15
		Linear	STD1	MULTICAL	2206133	10.300	0.5772	27.0730	0	15
18	AMY	Linear	BLANK	NORMAL SALINE	123	0.000	-0.0003	0	0	15
		Linear	STD1	MULTICAL	2206133	184.000	0.0498	3672.6545	0	15
19	HBA1	Point-To-Point	BLANK	NORMAL SALINE	123	0.000	-0.0006	0	0	0
		Point-To-Point	STD1	HBA1C CAL 2	S052357	4.900	0.0959	50.7772	0	0
		Point-To-Point	STD2	HBA1C CAL 3	S052357	8.500	0.3730	22.7516	0	0
		Point-To-Point	STD3	HBA1C CAL 4	S052357	11.900	0.6477	18.3557	0	0
		Point-To-Point	STD4	HBA1C CAL 5	S052357	15.700	0.8325	16.0453	0	0
20	CRP	Cubic Spline	BLANK	NORMAL SALINE	123	0.000	0.0269	0	0	0
		Cubic Spline	STD1	CRP CAL	B091744B	60.000	0.1751	404.8583	0	0
		Cubic Spline	STD2	CRP CAL	B091744B	30.000	0.1061	378.7879	0	0
		Cubic Spline	STD3	CRP CAL	B091744B	15.000	0.0629	416.8667	0	0
		Cubic Spline	STD4	CRP CAL	B091744B	7.500	0.0421	493.4211	0	0
		Cubic Spline	STD5	CRP CAL	B091744B	3.750	0.0338	543.4785	0	0
		Cubic Spline	STD6	CRP CAL	B091744B	1.875	0.0289	637.6007	0	0
21	CRE	Linear	BLANK	NORMAL SALINE	123	0.000	0.0009	0	0	15
		Linear	STD1	MULTICAL	1707039A	3.940	0.0226	179.9007	0	15
22	LIP	Linear	BLANK	NORMAL SALINE	123	0.000	0.0541	0	0	15
		Linear	STD1	MULTICAL	2206133	99.400	0.0716	5580.0065	0	15