



## CALIBRATION CERTIFICATE

ADL/DO/SER/007/23-24

Customer Name & Address	K.B Diagnostics No:7, New No: 24, Othavadai Chetty Street Vandavasi, Tiruvannamalai -604408
-------------------------	--

INSTRUMENT	INSTALLED ON	CALIBRATED ON
Mispa I3	04-Feb-2019	27/April/2023

### EQUIPMENT DETAILS

Instrument name	Mispa I3
Serial No	21192210372
Environmental condition:	Temperature:
Equipment calibration	
Parameter calibration	Inhouse (Calibrated IC card provided along with each kit)

This is to certify that the above-mentioned product has been calibrated & tested for the performance of kits with Control.

## AGAPPE DIAGNOSTICS LTD. ISO 9001:2015 | EN ISO 13485:2016 CERTIFIED COMPANY | CIN : U24239MH1998PLC115413

**CORPORATE OFFICE / REAGENT PLANT**  
Agappe Hills, Pattimattom (PO), Dist. Ernakulam, Kerala - 683 562, India.  
Tel: +91 484 286 7000 | Email: agappe@agappe.in

**EQUIPMENT PLANT**  
X/588-CB, Block No. 32, KINFRA Small Industrial Park,  
Nellad, Cochin, Kerala, India - 686 721. Tel: +91 484 276 7477.

**KOLKATA OFFICE**  
406, Merlin Matrix, Plot No-10, Block-DN, Sector V,  
Salt Lake City, Kolkata - 700 091.  
Tel: +91 33 4003 0451 | Email: kolkattaoffice@agappe.in

**MUMBAI [REGISTERED OFFICE]**  
401 & 402, 4<sup>th</sup> Floor, Jaisingh Business Centre, 119,  
Sahar Road, Parsiwada, Andheri (East), Mumbai - 400 099, India.  
Tel: +91 22 4300 8000 | Email: mumbaioffice@agappe.in

**DELHI OFFICE**  
DSM 540, 5<sup>th</sup> Floor, DLF Tower, Shivaji Marg,  
New Delhi - 110 015, India.  
Tel: +91 11 4558 8416 | Email: delhioffice@agappe.in

**BANGALORE OFFICE**  
S-6, II<sup>nd</sup> Floor, Red Cross Bhavan, No:26 Race Course Road,  
Bangalore - 560 001. Tel: +91 80 2228 8288  
Email: bangaloreoffice@agappe.in



## Equipment Test report

Checked the IC Card Reader	: Found Ok
Checked the Probe	: Found Ok
Checked Mother Board	: Found Ok
Checked the Wash Assembly	: Found Ok
	:
Channel-1 (with cartridge)	:
Channel-1 (without cartridge)	:
Channel-2 (with cartridge)	:
Channel-2 (without cartridge)	:
	:

## Performance validation with Control

Protein Control

LOT NO:

EXP:

Parameter	Low	High	Result	Target
CRP				
RF				
ASO				

Next Calibration due on 26/Oct/2023

For Agappe Diagnostics Ltd,



Authorized Signatory

Name: R.Vignesh

Manager - Customer Service

**AGAPPE DIAGNOSTICS LTD.** ISO 9001:2015 | EN ISO 13485:2016 CERTIFIED COMPANY | CIN : U24239MH1998PLC115413

**CORPORATE OFFICE / REAGENT PLANT**  
 Agappe Hills, Pattimattom (PO), Dist. Ernakulam, Kerala - 683 562, India.  
 Tel: +91 484 286 7000 | Email: agappe@agappe.in

**EQUIPMENT PLANT**  
 X/588-CB, Block No. 32, KINFRA Small Industrial Park,  
 Nellad, Cochin, Kerala, India - 686 721. Tel: +91 484 276 7477.

**KOLKATA OFFICE**  
 406, Merlin Matrix, Plot No-10, Block-DN, Sector V,  
 Salt Lake City, Kolkata - 700 091.  
 Tel: +91 33 4003 0451 | Email: kolkattaoffice@agappe.in

**MUMBAI [REGISTERED OFFICE]**  
 401 & 402, 4<sup>th</sup> Floor, Jaisingh Business Centre, 119,  
 Sahar Road, Parsiwada, Andheri (East), Mumbai - 400 099, India.  
 Tel: +91 22 4300 8000 | Email: mumbaioffice@agappe.in

**DELHI OFFICE**  
 DSM 540, 5<sup>th</sup> Floor, DLF Tower, Shivaji Marg,  
 New Delhi - 110 015, India.  
 Tel: +91 11 4558 8416 | Email: delhioffice@agappe.in

**BANGALORE OFFICE**  
 S-6, II<sup>nd</sup> Floor, Red Cross Bhavan, No:26 Race Course Road,  
 Bangalore - 560 001. Tel: +91 80 2228 8288  
 Email: bangaloreoffice@agappe.in

 MAINT



Name	Lot No:	Expiry	Test	Name	Lot No:	Expiry	Test
ASO	0123092	31/08/24	01	IgE	0724011	31/12/24	06
CRP	0223123	30/11/24	25	CYSC	No Saved Calibrations		
HbA1C	0323123	30/11/24	25	D_DIM	No Saved Calibrations		
RF	0423111	31/10/24	11	Trop I	No Saved Calibrations		
MALB	0523121	30/11/24	25	PCT	No Saved Calibrations		
FERRIT	No Saved Calibrations			Vitamin D	1224011	31/12/24	24

Exit



MISPA-i3