



ASIAN TECHNOLOGY

(A HOUSE OF CALIBRATION)



CALIBRATION CERTIFICATE

| | | | |
|---------------------------------|--|----------------------------|--------------------------|
| Customer name And Address | M/S: Ever Green Health Centre, 241/43, G.T Road, Post –Liluah, District- Howrah, Pin-711204, State- West Bengal, | Service request No. & date | P/02/24/08/23&24/08/2023 |
| | | ULR no. | CC223923000004857 F |
| | | Cert. No. | AT23000004857 |
| | | Date of Receipt of DUC | 24/08/2023 |
| | | Date of calibration | 24/08/2023 |
| | | Date of issue | 25/08/2023 |
| | | Suggested due date | 23/08/2024 |

Instrument Details

| | | | |
|------------------|----------------|--------------------------|---------|
| Instrument name | Micro Pipette | Sr. No. | 2010335 |
| Make / Model No. | MCR | Location | Lab |
| Range / Size | 100 to 1000 µl | Calibration Performed at | Lab |
| Least Count | 5 µl | Visual Inspection | OK |

Detail of reference standards & Major equipments used

| | | | |
|----------------------|--------------------------|--------------------------|--|
| Equipment Name | A set of weight box | Digital Weighing Balance | |
| Make | WEIGHTRONICS | AND | |
| Model / SR No. | ----- | GH-252/ 78002 | |
| Certificate No. | TYCON/WB/12/274 | TYCON/WB/02/23/327 | |
| Calibration Validity | 25/12/2023 | 19/02/2024 | |
| Calibration by | Tycon Engineering | Tycon Engineering | |

| | | | | |
|----------------------------|-------------------|------------|-----------------------|----------------|
| Environmental Condition | Temperature | (23± 3) °C | Calibration Reference | IS:4162 (1985) |
| | Relative Humidity | (50±20) % | Work Instruction | AT/WI-20M |

Calibration Results

| Serial No. | Nominal Capacity in (µl) | Measured Capacity in (µl) | Uncertainty At 95% C.L. (coverage factor k=2) |
|------------|--------------------------|---------------------------|---|
| 01. | 200 | 199.88 | ±0.46 µl |
| 02. | 500 | 499.79 | |
| 03. | 1000 | 999.65 | |

Remarks:

| |
|--|
| ❖ (1) Standard equipment use for calibration are traceable to national/ international standards. |
| ❖ (2) The reported expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by coverage factor k = 2 such that the coverage probability corresponds to approximately 95%. (3) The above results are valid at the time of and under the stated conditions measurement. (4) This certificate is refers only to the particular item submitted for calibration. (5) Next calibration due date given as requested by the customer. |

Calibrated By
(Calibration Engg./TM)
(MAHIPAL)
Form No.- QF-47

Approved By
(QM/TM)
(NEERAJTYAGI)
Page No. 1 of 1



---End of Report---