



# ASIAN TECHNOLOGY

(A HOUSE OF CALIBRATION)



## CALIBRATION CERTIFICATE

Customer name And Address	M/S: <b>Ever Green Health Centre</b> , 241/43, G.T Road, Post – Liluah, District- Howrah, Pin- 711204, State- West Bengal,	Service request No. & date	P/02/24/08/23&24/08/2023
		ULR no.	CC223923000004862 F
		Cert. No.	AT23000004862
		Date of Receipt of DUC	24/08/2023
		Date of calibration	24/08/2023
		Date of issue	25/08/2023
		Suggested due date	23/08/2024

Instrument Details			
Instrument name	Centrifuge	I.D. / S. No.	-----
Make / Model No.	R-BC	Location	Lab
Range / Size	0 to 5250 rpm	Accuracy	-----
Least Count	10 rpm	Visual Inspection	OK

Detail of reference standards & Major equipments used			
Equipment Name	Digital Tachometer		
Make	LT Lutron		
Model / SR No.	-----/A.C92792		
Certificate No.	QCTS/000246/10		
Calibration Validity	20/10/2023		
Calibration by	<b>Quality Calibration Testing Sol.</b>		

Environmental Condition	Temperature	20±2° C	Calibration Reference	IS: 12508
	Relative Humidity	50±10 %	Calibration Procedure	CP-37

Calibration Results			
Serial No.	Standard value in (rpm)	Measured value in (rpm)	Uncertainty At 95% C.L. (coverage factor K=2)
01.	503.4	500	±0.2 %rdg.
02.	1002	1000	
03.	2006	2000	
04.	3007	3000	
05.	5008	5000	

Remarks:

❖	(1) Standard equipment use for calibration are traceable to national/ international standards.
❖	(2)The reported expended uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by coverage factor = 2 such that the coverage probability corresponds to approximately 95%. (3)The above results are valid at the time of and under the stated conditions measurement. (4)This certificate is refers only to the particular item submitted for calibration. (5) Next calibration due date given as requested by the customer.

Calibrated By  
(Calibration Engg./TM)  
(MAHPAL)  
Form No.-QF-47



Approved By  
(QM/TM)  
(NEERAJTYAGI)  
Page No. 1 of 1

---End of Report---