

HORIBA

Medical

HORIBA India Private Limited

246, Okhla Industrial Estate Phase-III,
New Delhi 110020, India
Tel : +91 (11) 4646 5000 / 4669 5001
Fax : +91 (11) 4646 5020
http : //www.horiba.com
CIN : U73100DL2006PTC153232

HIN/MED/2023-2024 /10556324
25TH APRIL 2023

CALIBRATION CERTIFICATE

This is to certify that the Hematology Analyzer Horiba YUMIZEN H550 bearing serial number 012YAXH03187 installed at KRSNAA DIAGNOSTIC LAB PATHANKOT as calibrated on 25TH APRIL 2023.

Calibrator : ABX MINOCAL
Lot No. : CX481
Expiry Date : 05th MAY 2023

The reports of Blank Cycle, Repeatability and Calibration Values were all found in acceptable range.

Next calibration cycle is due on 24TH APRIL 2024

A circular stamp with the text "HORIBA INDIA PVT. LTD." around the perimeter. Overlaid on the stamp is a handwritten signature in black ink.

Shrish Dixit

(Head- Products & Marketing)
For Horiba India Pvt. Ltd.

246, Okhla Industrial Estate Phase-
III, New Delhi 110020, India
Tel : +91 (11) 4546 5000 / 46695001
Fax : +91 (11) 4646 5020
http : //www.horiba.com
CIN : U73100DL2006PTC1M232

Calibration

Calibrator Information

Sample ID

Exp. date

Lot number

Name

Modified on

Coefficients		WBC	RBC	HGB	HCT	PLT	MPV
New		1.126	1.059	0.994	1.069	1.146	1.059
Current		1.126	1.025	0.994	1.030	1.047	1.059
Target		8.80	4.42	12.5	36.5	248	10.7
Mean		8.82	4.44	12.7	35.5	240	10.1
CV(%)		0.64	1.17	0.51	1.27	3.59	0.71

Selected runs
(5 minimum)

5/5

|||

	Run Date & Time	WBC (10 ³ /μL)	RBC (10 ⁶ /μL)	HGB (g/dL)	HCT (%)	PLT (10 ³ /μL)	MPV (μm ³)
<input checked="" type="checkbox"/>	25/04/2023 03:58:51 PM	8.89	4.38	12.7 h	34.9	238	10.0
<input checked="" type="checkbox"/>	25/04/2023 04:00:21 PM	8.75	4.45	12.8 h	35.6	254	10.1
<input checked="" type="checkbox"/>	25/04/2023 04:01:50 PM	8.83	4.50 h	12.7 h	36.0	241	10.1
<input checked="" type="checkbox"/>	25/04/2023 04:06:40 PM	8.77	4.39	12.6	35.0	230	10.1
<input checked="" type="checkbox"/>	25/04/2023 04:08:25 PM	8.85	4.46	12.7 h	35.7	237	10.0

Arjun
25/4/23.