



AACPL CALIBRATION PVT. LTD.

CALIBRATION CERTIFICATE

7.8F-01



CC - 2646

Certificate No	AACPL / 07106F	Field	Elec.Tech.	Group	Page No.	1 of 1
SRF No & Date	230929.7	DT - 29 / 09 / 2023	ULR No.	CC2646230000 07106F		
Calibration Method :	Direct Method			Accuracy		
Company Name	M/s Krsnaa Diagnostics Ltd.			Calibration Date	29.09.2023	
Address	Shri Prabodh Chander District Hospital Shahpur Road, Pathankot (PB)			Recommended Date for Next Calib.	28.09.2024	
				Certificate Issue Dt.	03.10.2023	
				UUC Condition	Good	

Calibration Instrument Detail

SI	Instrument Name	Range	Least Count	Make / SI. No	Party ID / Location
1	Dig. Timer (Centrifuge M/c.)	0 - 99 minute	1 minute	Remi ZIDN-16602	KDDHPBCCCF06 Clinical Chemistry

Standard Equipments Used (Traceable to National Standard)

SI	Instrument Name	Make/SI No	Calibrated By	Cal Certificate No	Cal. Valid Upto
1	Dig. Time Interval Calibrator	Voltagram / ----	VTC	CC293522000002078F	27/10/2023

Reference Standard	Calibration Purpose	Humidity	Temperature	Calibration Performed At
IS : 1248	Time Measurement	50 ± 20) % RH	(25 ± 10)°C	Site

Calibration Procedure : As per Calibration Procedure Manual AACPL / CP / 03, Section No : CPET - 02

Calibration Result

SI	Selected Range (In s)	Average Read. (Std.) (In s)	Average Read. (UUC) (In s)	Deviation (In s)	Uncertainty at approx 95% confidence Level and coverage factor $k = 2$ is ±
1	900	900.056	900	-0.056	34.64 s
2	1800	1800.096	1800	-0.096	34.66 s
3	3600	3600.157	3600	-0.157	34.85 s

Note

- 1 The calibration results reported in this calibration certificate are valid at the time of & under stated condition
- 2 This certificate cannot be reproduced except in full without our prior permission in writing
- 3 This certificate refers only to the particular item(s) submitted for calibration
- 4 UUC - Unit under calibration

----- End of Certificate -----

Award
Calibrated by



Approved by
[Signature]
Anand Chaurasia
(CEO)