



A A CALIBRATION PVT. LTD.

CALIBRATION CERTIFICATE

7.8F-01



CC - 2646

Certificate No	AACPL / 07111F	Field	Mech.	Group	Volume	Page No.	1 of 1
SRF No & Date	230929.8	DT - 29 / 09 / 2023		ULR No.	CC2646230000 07111F		
Calibration Method :	Comparison Method			Accuracy			
Company Name	M/s Krsnaa Diagnostics Ltd.			Calibration Date	29.09.2023		
Address	Shri Prabodh Chander District Hospital Shahpur Road, Pathankot (PB)			Recommended Date for Next Calib.	28.09.2024		
				Certificate Issue Dt.	03.10.2023		
				UUC Condition	Good		

Calibration Instrument Detail

SI	Instrument Name	Range	Least Count	Make / SI. No	Party ID / Location
1	Micro Pipette	20 to 200 µl	1 µl	Dragon Lab YE222AX0047996	KDDHPBCCMP09 Clinical Chemistry

Standard Equipments Used (Traceable to National Standard)

SI	Instrument Name	Make/SI No	Calibrated By	Cal Certificate No	Cal. Valid Upto
1	Weighing Balance	Radwag / 668683	Bluebox Techno	BBT/452/JAN/23	05/03/2024

Reference Standard	Calibration Purpose	Humidity	Temperature	Calibration Performed At
ISO - 8655 - 6	Volume Measurement	(50 ± 10)%RH	(23 ± 2)°C	Laboratory

Calibration Procedure : As per Calibration Procedure Manual AACPL / CP / 01C, Section No : CP - 03

Calibration Result

S. No	Nominal Value on UUC (in µl)	Average Volume at 27 °C (in µl)	Deviation (in µl)	Uncertainty at approx 95% confidence Level and coverage factor k = 2 is ± (in µl)
1	20	20.00143	-0.00143	0.25
2	100	100.00259	-0.00259	0.25
3	200	200.00317	-0.00317	0.25

Note

- 1 The calibration results reported in this calibration certificate are valid at the time of & under stated condition
- 2 This certificate cannot be reproduced except in full without our prior permission in writing
- 3 This certificate refers only to the particular item(s) submitted for calibration
- 4 UUC - Unit under calibration

----- End of Certificate -----

Archieh
Calibrated by



Approved by
Archieh
Anand Chaurasia
(CEO)