

A A CALIBRATION PVT. LTD. CALIBRATION CERTIFICATE



C - 2646

Certificate No	AACPL / 07112F Field	Mech.	Group	Volume	Page N		1 of 1
SRF No & Date	230929.8 DT - 29 / 09 / 2023		ULR No.	CC2646230000 07112F			
Calibration Method :	Comparison Method		Accuracy				
Company Name	M/s Krsnaa Diagnostics Ltd.		Calibratio	n Date		29.09.	
Address	Shri Prabodh Chander District Hospital		Recommended Date for Next Calib.		28.09.		
	Shahpur Road, Pathankot (PB)		Certificate Issue Dt.		03.10.	2023	
			UUC Cond	dition		Good	

Calibration Instrument Detail

SI	Instrument Name	Range	Least Count	Make / SI. No	Party ID / Location
1	Micro Pipette	100 to 1000 µl	5 µl	Dragon Lab	KDDHPBCCMP10
				YE221AW0154702	Clinical Chemistry

Standard Equipments Used (Traceable to National Standard)

ſ	SI	Instrument Name	Make/SI No	Calibrated By	Cal Certificate No	Cal. Valid Upto
	1	Weighing Balance	Radwag / 668683	Bluebox Techno	BBT/452/JAN/23	05/03/2024

Reference Standard	Calibration Purpose	Humidity	Temperature	Calibration Performed At
ISO - 8655 - 6	Volume Measurement	(50 ± 10)%RH	(23 ± 2)°C	Laboratory

Calibration Procedure: As per Calibration Procedure Manual AACPL / CP / 01C, Section No : CP - 03

	Calibration Result					
S. Nominal Value on UUC			Uncertainty at approx 95% confidence			
No	(in µl)	(in μl)	(in µl)	Level and coverage factor k = 2 is ± (in μl)		
1	100	99.99856	0.00144	0.25		
2	500	499.99563	0.00437	0.25		
3	1000	999.99243	0.00757	0.25		

Note

- 1 The calibration results reported in this calibration certificate are valid at the time of & under stated condition
- 2 This certificate cannot be reproduced except in full without our prior permission in writing
- 3 This certificate refers only to the particular item(s) submitted for calibration
- 4 UUC Unit under calibration

----- End of Certificate -----

CALIBRATED CALIBRATED

Approved by

Anadd Chaurasia (CEO)

Ashick.
Calibrated by