



The Accurate Calibration & Medical Solution

CALIBRATION CERTIFICATE

CERTIFICATE OF PERFORMANCE FOR MICROSCOPE

ISSUED TO M/S KRSNAA DIAGNOSTICS LTD.
ADDRESS SHRI PRABODH CHANDER DISTRICT HOSPITAL, SHAHPUR ROAD, PATHANKOT (PB) - 145001
CERTIFICATE NO. TAC/09-23/KDL/MS-1001
LOCATION/DEPT. CLINICAL PATHOLOGY DATE 29.09.2023
ID NO./BAR CODE KDDHPBCPMS05 MODEL NO. LX-300
MAKE GLOMAX SERIAL NO. NA

CALIBRATION STANDARD USED

Sl.No.	Master Equipment	Manufacturer	Model No.	Serial No.	Cal. Date	Cal Due Date
1	Electrical Safety Analyzer	DATREND Systems Inc.	vPed - Mini	VPM18100028	15.10.2022	15.10.2023

ENVIRONMENTAL CONDITIONS

Temperature : $22 \pm 2^{\circ}\text{C}$
Humidity : $50 \pm 10\%RH$

VISUAL TEST

Sr.No.	Tests	Remarks
1	Power Cords, Cable Checking	OK
2	Mains Socket Checking	OK
3	Equipment Type (B, BF, CF)	CF
4	Equipment Class (I, II, III)	I

FUNCTIONAL TEST

Sr.No.	Parameters	Measured Value	Remarks
1	Voltage Between Live and Neutral (VLN)	230.4 V (220-240 V)	OK
2	Voltage Between Live and Earth (VLE)	229.5 V (220-240V)	OK
3	Voltage Between Neutral and Earth (VNE)	0.3 V (0-5V)	OK
4	Load Current (Ia)	0.3 A	OK
5	Enclosure Current	0.6 μA (0-10 μA)	OK
6	Leaking Current (IL)	142 μA (0-300 μA)	OK
7	Time for Testing (t)	15 (min)	OK

Electrical Safety Test Passed as per IEC Standard 60601.1

Calibration Passed YES NO

Calibration Date: 29. 09. 2023

Calibration Due Date: 28. 09. 2024

Remarks:

1. Certificates shall not be reproduced except in full, without the written approval of The Accurate Calibration & Medical Solution.
2. The Calibration result reported in this certificate are valid at the time of Calibration and the stated condition of measurement.

Calibrated by



Approved by

Plot No. 72/G1, Ganesh Vihar Sirsi Road, Jaipur (Raj.) -302012

Mob.: 9672294758

E-mail : theaccuratecalibration@gmail.com