



# SURYAADARASH CALIBRATION AND CONSULTANT (OPC) PVT. LTD.

WE ALWAYS TRY TO GET BEST MEASUREMENT

Formerly UNIQUE CALIBRATION & CONSULTANT SERVICES  
AN ISO 9001:2015 CERTIFIED CALIBRATION LABORATORY  
NABL Accredited Calibration Laboratory As per ISO/IEC 17025:2017



## CALIBRATION CERTIFICATE

Format No.: SCCPL-QF-7.8.1		Certificate No.: SCCPLCC20240203-016		ULR No.: CC328724000002627F				
Group: Volume			Discipline: Mechanical					
Customer's Name & Address			Page No.: 1 of 1	Calibration Performed: At Lab				
M/s. MOLXPRT DIAGNOSTICS PVT LTD 106, FIRST FLOOR, MAHAGUN MART, PLOT GH-04, OPPOSITE GAUR CITY 2, SECTOR 16C, NOIDA PHASE-2, GREATER NOIDA, UTTAR PRADESH 201306			Date of Receipt: 03-02-2024		Service Request No.: SCCPLSRN20240203-006			
			Certificate Issue Date: 07-02-2024		Date of Calibration: 06-02-2024			
			Next Due Date (Suggested by Customer): 05-02-2025					
			Equipment Description (UUC) & Identification			Name of the Instrument: Micro pipette		Make: Dragon Lab
			Range: 10 µl to 100 µl			Least Count: 1 µl		Model: -----
Condition of Instrument: Ok			Sr. No.: YE223AX0088929		I.D. No.: -----			
Master Equipment Used			Location: Lab					
Sr. No.	Instrument's Name	Range & I.D. No.	Calibration Agency	Calibration Certificate No.	Valid Up to			
1	Digital Weighing Balance	0 to 220 g & SCCPL/CAL/005	SCCPL	SCCPLCC20240113-002	14-01-2025			
Environmental Conditions			Calibration Procedure		Calibration Reference Standard			
Temperature: 27 ± 3 °C, Humidity: 45 to 80 %RH Atmospheric Pressure : 14.2 psi			SCCPL 03 SOP 018		ISO 8655			

### CALIBRATION RESULTS

Sr. No.	Nominal Capacity (µl) at 27 °C	Actual Volume (µl) at 27 °C	Error (µl)	Expanded Uncertainty (±) µl
1	10	10.41	-0.41	0.13
2	25	25.94	-0.94	0.13
3	50	51.42	-1.42	0.13
4	75	76.89	-1.89	0.13
5	100	102.36	-2.36	0.92

\*\*\*End of Certificate\*\*\*

Calibrated By  
Rohit Prasad

Calibration Engineer

Reviewed/Approved By  
L.M. Mahato  
Chief Executive Officer (CEO)



- Note:**
- The calibration results reported in this certificate are valid at the time of and stated condition of measurement.
  - This report is not to be reproducing wholly or in part and cannot be used as evidence in the Court of Law and should not be used in any advertising media without our special permission in writing.
  - Total liability of our Organization is limited to the invoiced amount.
  - Calibration Certificate without signature are not valid.
  - All our certificates are metrological traceable.
  - The results relate only to the item calibrated.
  - UUC: Unit Under Calibration.
  - Expanded Uncertainty of measurement expressed at approximately the 95 % level of confidence, using a coverage factor of k = 2.
  - Scan this QR code to download our NABL Accreditation Scope & Certificate.



Working : W.Z.1390/Z-2 Nangal Raya Extn. Pankha Road New Delhi - 110046

Regd. Office : R.Z.56A, Indra Park Gali No.25, Palam Colony, New Delhi -110045

M.: 9958445875, 9313768426, 9999761885 E-mail: suryaadarshcalibration@gmail.com