



SURYAADARSH CALIBRATION AND CONSULTANT (OPC) PVT. LTD.

Formerly UNIQUE CALIBRATION & CONSULTANT SERVICES

AN ISO 9001:2015 CERTIFIED CALIBRATION LABORATORY

NABL Accrediated Calibration Laboratory As per ISO/IEC 17025:2017



CALIBRATION CERTIFICATE

Format No.: SCCPL-QF-7.8.1		Certificate No.: SCCPLCC20240203-018		ULR No.: CC328724000002629F	
Group: Volume			Discipline: Mechanical		
Customer's Name & Address M/s. MOLXPRT DIAGNOSTICS PVT LTD 106, FIRST FLOOR, MAHAGUN MART, PLOT GH-04, OPPOSITE GAUR CITY 2, SECTOR 16C, NOIDA PHASE-2, GREATER NOIDA, UTTAR PRADESH 201306			Page No.: 1 of 1	Calibration Performed: At Lab	
			Date of Receipt: 03-02-2024		
			Service Request No.: SCCPLSRN20240203-006		
			Certificate Issue Date: 07-02-2024		
			Date of Calibration: 06-02-2024		
Next Due Date (Suggested by Customer): 05-02-2025					
Equipment Description (UUC) & Identification					
Name of the Instrument: Micro pipette			Make: -----		
			Model: -----		
Range: 1000 µl			Sr. No.: NI473590		
Least Count: -----			I.D. No.: -----		
Condition of Instrument: Ok			Location: Lab		
Master Equipment Used					
Sr. No.	Instrument's Name	Range & I.D. No.	Calibration Agency	Calibration Certificate No.	Valid Up to
1	Digital Weighing Balance	0 to 220 g & SCCPL/CAL/005	SCCPL	SCCPLCC20240113-002	14-01-2025
Environmental Conditions			Calibration Procedure	Calibration Reference Standard	
Temperature: 27 ± 3 °C, Humidity: 45 to 80 %RH			SCCPL 03 SOP 018	ISO 8655	
Atmospheric Pressure : 14.2 psi					

CALIBRATION RESULTS

Sr. No.	Nominal Capacity (µl) at 27 °C	Actual Volume (µl) at 27 °C	Error (µl)	Expanded Uncertainty (±) µl
1	1000	1001.68	-1.68	0.92


End of Certificate

Calibrated By
Rohit Prasad
Calibration Engineer

Reviewed/Approved By
M. Mahato
Chief Executive Officer (CEO)



Note: - 1. The calibration results reported in this certificate are valid at the time of and stated condition of measurement.
 2. This report is not to be reproducing wholly or in part and cannot be used as evidence in the Court of Law and should not be used in any advertising media without our special permission in writing.
 3. Total liability of our Organization is limited to the invoiced amount.
 4. Calibration Certificate without signature are not valid.
 5. All our certificates are metrological traceable.
 6. The results relate only to the item calibrated.
 7. UUC: Unit Under Calibration.
 8. Expanded Uncertainty of measurement expressed at approximately the 95 % level of confidence, using a coverage factor of k = 2.
 9. Scan this QR code to download our NABL Accreditation Scope & Certificate.



Working : W.Z.1390/Z-2 Nangal Raya Extn. Pankha Road New Delhi - 110046

Regd. Office : R.Z.56A, Indra Park Gali No.25, Palam Colony, New Delhi -110045

M.: 9958445875, 9313768426, 9999761885 E-mail: suryaadarshcalibration@gmail.com