



SURYAADARASH CALIBRATION AND CONSULTANT (OPC) PVT. LTD.

WE ALWAYS TRY TO GET BEST MEASUREMENT

Formerly UNIQUE CALIBRATION & CONSULTANT SERVICES
AN ISO 9001:2015 CERTIFIED CALIBRATION LABORATORY
NABL Accredited Calibration Laboratory As per ISO/IEC 17025:2017



CALIBRATION CERTIFICATE

Format No.: SCCPL-QF-7.8.1	Certificate No.: SCCPLCC20240203-013	ULR No.: CC328724000002624F
Group: Temperature	Discipline: Thermal	

Customer's Name & Address M/s. MOLXPRT DIAGNOSTICS PVT LTD 106, FIRST FLOOR, MAHAGUN MART, PLOT GH-04, OPPOSITE GAUR CITY 2, SECTOR 16C, NOIDA PHASE-2, GREATER NOIDA, UTTAR PRADESH 201306	Page No.: 1 of 1	Calibration Performed: At Site
	Date of Receipt: 03-02-2024	
	Service Request No.: SCCPLSRN20240203-006	
	Certificate Issue Date: 07-02-2024	
	Date of Calibration: 06-02-2024	
Next Due Date (Suggested by Customer): 05-02-2025		

Equipment Description (UUC) & Identification	
Name of the Instrument: Digital Temperature Indicator with Sensor (Refrigerator)	Make: -----
Range: -50 °C to 70 °C	Model: -----
Least Count: 0.1 °C	Sr. No.: -----
Condition of Instrument: Ok	I.D. No.: MDPL/RF-01
Location: Lab	

Master Equipment Used					
Sr. No.	Instrument's Name	Range & I.D. No.	Calibration Agency	Calibration Certificate No.	Valid Up to
1	High Precision Thermometer with 4 wire RTD Sensor	-100 °C to 600 °C & SCCPL/CAL/022A & SCCPL/CAL/022B	Bansal Calibration Services	BCS/N/23120002-002	06-12-2024

Environmental Conditions Temperature: 20 ± 15 °C, Humidity: 30 to 75 %RH	Calibration Procedure SCCPL 03 SOP 026	Calibration Reference Standard DKD-R 5-7
--	--	--

CALIBRATION RESULTS

Sr. No.	UUC Value (°C)	Ref. Standard Value (°C)	Error (°C)	Expanded Uncertainty (± °C)
1	-30.0	-30.085	0.085	0.44
2	-10.1	-10.157	0.057	0.44
3	0.0	-0.214	0.214	0.44
4	2.0	1.719	0.281	0.44
5	6.1	5.747	0.353	0.44
6	10.1	9.631	0.469	0.44

End of Certificate

Calibrated By
Vikash Kumar

Calibration Engineer

Reviewed/Approved By
L.M. Mahate

Chief Executive Officer (CEO)

<p>Note: - 1. The calibration results reported in this certificate are valid at the time of and stated condition of measurement.</p> <p>2. This report is not to be reproducing wholly or in part and cannot be used as evidence in the Court of Law and should not be used in any advertising media without our special permission in writing.</p> <p>3. Total liability of our Organization is limited to the invoiced amount.</p> <p>4. Calibration Certificate without signature are not valid.</p> <p>5. All our certificates are metrological traceable.</p> <p>6. The results relate only to the item calibrated.</p> <p>7. UUC: Unit Under Calibration.</p> <p>8. Expanded Uncertainty of measurement expressed at approximately the 95 % level of confidence, using a coverage factor of k = 2.</p> <p>9. Scan this QR code to download our NABL Accreditation Scope & Certificate.</p>	
---	--

Working : W.Z.1390/Z-2 Nangal Raya Extn. Pankha Road New Delhi - 110046

Regd. Office : R.Z.56A, Indra Park Gali No.25, Palam Colony, New Delhi -110045
M.: 9958445875, 9313768426, 9999761885 E-mail: suryaadarshcalibration@gmail.com