



ASIAN TECHNOLOGY

(A HOUSE OF CALIBRATION)



CALIBRATION CERTIFICATE

Customer name And Address	M/S: Dr. Garg's Diagnostics 11702/3, Main G.T. Road, Shakti Nagar Delhi-110007	Service request No. & date	A/01/13/05/23 & 13/05/2023
		ULR no.	CC223923000002145F
		Cert. No.	AT23000002145
		Date of Receipt of DUC	13/05/2023
		Date of calibration	13/05/2023
		Date of issue	14/05/2023
		Suggested due date	12/05/2024

Instrument Details

Instrument name	Refrigerator	Location	-----
Make	Croma	Accuracy	-----
Range / Size	2 to 8 °C	Visual Inspection	OK
Least Count	1 °C	Calibration Performed at	Site
I.D. No. / Sr. no.	DGD/LAB/EQP/01 / -----		

Detail of reference standards & Major equipments used

Equipment Name	SPRT with Indicator		
Make	Fluke		
Model / SR No.	05640		
Certificate No.	CC288822000001839F		
Calibration Validity	27/10/2023		
Calibration by	Sigma		

Detail of reference standards & Major equipments used

Environmental Condition	Temperature	(25±4) °C	Calibration Reference	ITS:90 (1990)
	Relative Humidity	(50±20) %	Work Instruction	AT/WI-08TH

Calibration Results

Serial No.	Std. Reading in (°C)	Measured Reading in (°C)	Uncertainty At 95% C.L. (coverage factor K=2)
01.	1.988	2	±0.3 °C
02.	3.997	4	
03.	5.977	6	
04.	7.964	8	
05.	9.953	10	

Remarks:

- ❖ (1) Standard equipment use for calibration are traceable to national/ international standards
- ❖ (2) The reported expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by coverage factor = 2 such that the coverage probability corresponds to approximately 95% (3) The above results are valid at the time of and under the stated conditions measurement (4) This certificate is refers only to the particular item submitted for calibration. (5) Next calibration due date given as requested by the customer.

Calibrated By
(Calibration Engg./TM)
Q.B SAXENA)
Form No.- QF-47



Approved By
(QM/TM)
(NEERAJ TYAGI)
Page No. 1 of 1

---End of Report---