

**Beckman Coulter India Pvt. Ltd.**

Regional Office:

The DLF Galleria Building, Unit - 1201

Plot No: BG-08, Action Area IB, New Town

Kolkata – 700 156, West Bengal

Tel: +91 033 4005 0630

Fax: +91 033 4005 0631

05th May ,2023.**Calibration Certificate**

Name of the customer	KINS CARE & RESEARCH FOUNDATION PVT.LTD. Jhankar More, Siliguri, West Bengal-734001
Make	Beckman Coulter
Product Name	Dx560 – Hematology Analyzer
Serial No	BD090062

This is to certify that the cell counter installed at the above mentioned customer site has been calibrated with Calibrator (Lot no : 492316200, Expiry – 05th ,May 2023)

SL No	Procedures	Parameter	Remarks
1	Startup	WBC/RBC/HGB/PLT	PASS
2	Reproducibility	WBC/RBC/HGB/MCV/RDW/PLT/MPV	PASS
4	Carry Over	WBC/RBC/HGB/PLT/DIFF/NRBC	PASS
5	Calibration	WBC/RBC/HGB/MCV/PLT	PASS
6.	QC Tri level	WBC/RBC/HGB/MCV/PLT	PASS

Date of Calibration: 05/05/2023	Due date for Calibration : 04/05/2024
--	--

For Beckman Coulter India Pvt. Ltd.**Subhrabaran Chakrabarti****Area Applications Manager****Beckman Coulter India Pvt. Ltd.**Registered Office: 3rd Floor, B Wing, Art Guild House Phoenix Marketcity, LBS Marg, Kurla (W)

Mumbai - 400093 Tel:+91 22 6838 5000 Fax:+91 22 6838 5060

CIN: U33119MH2005PTC157177 www.beckmancoulter.com