



Aspen Diagnostics Pvt. Ltd.
(Rapid Diagnostic Group Of Companies)



Calibration Certificate


Customer Name	ESI Model Hospital.
Address	jaipur
Contact Person	Dr. Ratan Chaudhary
Contact No.	+91-9413561416
Instrument	Hematology Analyzer
Model	BC-6200
Serial No.	TW-93000389
Instrument Status	Under Warranty
Calibration Done On	12/03/2024
Next Calibration Due On	11/03/2025
Calibration Done By	Satish Sharma

For Aspen Diagnostics Pvt. Ltd.

Service Engineer

Satish

(Satish Sharma)





Aspen Diagnostics Pvt. Ltd.
(Rapid Diagnostic Group Of Companies)



CALIBRATION DETAIL ESI HOSPITAL, JAIPUR

Hematology Analyzer	Model: BC -6200	Serial No.: TW-93000389
---------------------	-----------------	-------------------------

Mode: OV-WB	Calibrator Lot No : PLUS 0324	EXP.DATE:05-04-2024
-------------	-------------------------------	---------------------

Parameter	WBC	RBC	HGB	MCV	PLT	RBC -O	PLT-O	WBC-D
Target	9.19	4.64	13.5	93.3	248			9.19
1.	8.67	5.09	14.1	92.6	264			8.40
2.	8.76	5.13	14.2	92.4	266			8.40
3.	8.41	5.10	14.2	92.6	274			8.24
4.	8.66	5.07	14.1	92.9	265			8.30
5.	8.55	5.12	14.2	92.6	270			8.50
6.	8.85	5.09	14.2	92.4	274			8.35
Mean	8.65	5.10	14.2	92.6	269			8.36
CV	1.792	0.441	0.35	0.17	1.7			1.062
Old Cal.Factors (%)	103.0	90.06	91.00	91.00	91.92	110.10	109.11	-
New Cal.Factors (%)	106.24	90.99	95.27	100.76	92.19			-
Old Tc %								101.10
New Tc %								103.44

For Aspen Diagnostics Pvt. Ltd.

Service Engineer

(Satish Sharma)

Date : 12/03/2024



Aspen Diagnostics Pvt. Ltd.
(Rapid Diagnostic Group Of Companies)




ISO 9001:2015
CERTIFIED COMPANY

Calibration Certificate

Customer Name	ESI Model Hospital.
Address	jaipur
Contact Person	Dr. Ratan Chaudhary
Contact No.	+91-9413561416
Instrument	Hematology Analyzer
Model	BC-6200
Serial No.	TW-93000440
Instrument Status	Under Warranty
Calibration Done On	12/03/2024
Next Calibration Due On	11/03/2025
Calibration Done By	Satish Sharma

For Aspen Diagnostics Pvt. Ltd.

Service Engineer


(Satish Sharma)



Aspen Diagnostics Pvt. Ltd.
(Rapid Diagnostic Group Of Companies)



CALIBRATION DETAIL E.S.I . HOSPITAL, JAIPUR

Hematology Analyzer	Model: BC -6200	Serial No.: TW-93000440
---------------------	-----------------	-------------------------

Mode: OV-WB	Calibrator Lot No : PLUS 0324	EXP.DATE:05-04-2024
-------------	-------------------------------	---------------------

Parameter	WBC	RBC	HGB	MCV	PLT	RBC -O	PLT-O	WBC-D
Target	9.19	4.64	13.5	93.3	248			9.19
1.	8.86	5.14	14.3	93.6	267			8.50
2.	8.62	5.08	14.3	93.5	261			8.81
3.	8.59	5.06	14.3	93.5	281			8.56
4.	8.89	5.05	14.3	93.8	268			8.67
5.	8.72	5.11	14.4	93.2	256			8.41
6.	8.70	5.12	14.4	92.9	271			8.69
Mean	8.73	5.09	14.3	93.4	267			8.60
CV	1.405	0.701	0.41	0.36	3.1			1.701
Old Cal.Factors (%)	100.0	93.00	95.00	98.90	95.00			
New Cal.Factors (%)	105.27	91.11	94.32	99.87	92.71			
Old Tc %								103.04
New Tc %								101.45

For Aspen Diagnostics Pvt. Ltd.

Service Engineer

(Satish Sharma)



Date : 12/03/2024