



CALIBRATION CERTIFICATE

Customer Name And Address	MS/ SUN SHINE PATHCARE PVT LTD D1/32MINAL JK ROAD BHOPAL	Certificate No.	MT/10/200224/AH10/002630
		ULR No.	CC333323000002630F
		SRF No. & Date	MT/10/20.02.2024
		Receipt Date	20/02/2024
		Date of Calibration	21/02/2024
		Suggested Due Date	21/02/2025
		Date of Issue	22/02/2024

Instrument Details			
Instrument Name	Digital Centrifuge (RPM Meter)	Sr. No.	-----
Make / Model No.	Remi	Location	Lab
Range / Size	Upto 4000rpm	Accuracy	-----
Least Count	10rpm	Visual Inspection	OK
I.D. No.	CFG-01	Calibration Performed At	Site

Detail of Reference Standard & Major Equipment Used			
Equipment Name	Techometer		
Make	Line-Seiki		
Model / SR No.	164-0142		
Certificate No.	IALME/02881/23		
Calibration Validity	05/12/2024		
Calibration by	Belz Calibration		

Environmental Condition	Temperature	25±4° C	Calibration Reference	NABL 129,130
	Relative Humidity	50±20 %	Calibration Procedure	CP//M/&S/01

Calibration Results

Serial No.	Std. (Master) Value in (RPM)	UUC Value in (RPM)	Uncertainty At 95% C.L. (coverage factor K=2)
01.	200	199.8	±1.2%rdg
02.	500	499.5	
02.	1000	998.5	
03.	2000	1998	
04.	4000	3996	

Remarks:

- ❖ (1) Standard equipment use for calibration are traceable to national/ international standards.
- ❖ (2) The reported expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by coverage factor $k = 2$ such that the coverage probability corresponds to approximately 95%.
- ❖ (3) The above results are valid at the time of and under the stated conditions measurement.
- ❖ (4) This certificate is refers only to the particular item submitted for calibration.
- ❖ (5) Next calibration due date given as requested by the customer.

Calibrated By
(Calibration Engineer)
(Sunil Yadav)
Format No. F01(7.8)

.....End of the Certificate.....

Approved By
(Quality Manager)
(Anil Saini)
Page No. 1 of 1

