

Certificate of Calibration

Date : 4 March 2023

Name of Instrument : BTS 350 (S.No. 8017512371)

Customer Name: MOLXPERT Diagnostics Pvt Ltd, 106, First Floor, Mahagun Mart, GH-04, Sector-16, Greater Noida West

Following parameters have been checked and adjusted as per specification of Manufacturer.

Details of Calibration:

| ADJUSTMENT PUMP | | | LIGHT INTENSITY | | | |
|-------------------|---------|-------|-----------------|------|--------|-----------|
| 04/03/23 13:52:26 | | | 04/03/23 13:46 | | | |
| TEST VOLUME (uL): | | | LEDS | TINT | MAIN | REFERENCE |
| VALUE | CURRENT | NEW | 340 | 14 | 976971 | 227284 |
| SAMPLE VOLUME | 40325 | 44384 | 405 | 13 | 942715 | 669064 |
| POSITIONING | 2032 | 2920 | 505 | 27 | 982007 | 698847 |
| PUMP DELAY | 2 | | 535 | 14 | 956834 | 482568 |
| 836935 | | | 560 | 11 | 913274 | 520068 |
| 5700 | | | 600 | 23 | 972614 | 595480 |
| | | | 635 | 12 | 969920 | 559192 |
| | | | 570 | 11 | 939431 | 590634 |

Date of Calibration: 4 March 2023

Next Date of Calibration: 5 March 2024

For Innogen Healthcare

Authorized Signatory



INNOGEN HEALTH CARE

Off : No. 4, 1st Floor, G-83, Main Bazar Laxmi Nagar, Delhi-110092
+91-9312793140, +91-9891762590 innogenhealthcare@gmail.com