



Premium Calibration Services



PLOT No - 73, ADARSH NAGAR COLONY, MODINAGR ROAD, HAPUR-245101

Ph:- 0122-7965803, 08279757019

E-mail: Pcslab141@gmail.com

CALIBRATION CERTIFICATE

ULR No. : CC374023000000025F	Calibration Done On : 20.10.2023	S.R.F.No : 0001/20
Issue Date : 21.10.2023	Calibration Valid Upto : 19.10.2024	S.R.F Date : 20.10.2023
UUC Receipt Date :	certificate No. PCS/00025F	Pages : 1 of 1

Calibrated for M/S.

Molxpert Diagnostic Pvt. Ltd.
106, 1st Floor Mahagun Mywoods Mart Opp Gaur City-2
Plot No-GH04, Sector-16C Greater Noida

Identification on UUC (Unit Under Calibration)

Instrument Name	Centrifuge	Size/Range	0 - 3500 RPM
Least Count	10 RPM	Conition Of Item	Good
Manufactured By	Bio Plus	Calibration Performed at	Site
Manufacturer M/Sr No.	----	Location	-----
Party ID Mark	CF-01	Ref. Standard/Procedure	IS: 12508

Calibration Procedure / Method : As per Calibration Procedure Document No. PCS/CP/01, Section No.M/CP-21

Environmental conditions (s) : Temperature : 25 ± 5 °C. Relative Humidity : 30 to 80 % RH

Standard & Major Equipment(s)

Standard(s)& Equipments	Make	ID No. / Serial No.	Calibrating Agency	Traceability ULR No. / Cert. No.	Valid Up to
Dig. Tachometer	Line Seiki	PCS-TM-01 /	KL Technology Solutions	CC325723000000071F	21.02.2024

Uncertainty of measurement (at Approx 95% Confidence Level With Coverage Factor k=2) = ± 0.04 %

Calibration Results:

U.U.C Value (RPM)	Standard Value (RPM)	Error (RPM)
500	500.0	0.0
1000	1000.0	0.0
1500	1500.1	-0.1
2000	2001.3	-1.3
2500	2501.8	-1.8
3000	3002.4	-2.4

***** End of Certificate *****

NOTE:

- The result are valid only for this Calibrated Item.
- The calibration report shall not be reproduced without the written approval of calibration laboratory.
- Result reported are valid at the time of & under the stated conditions of measurement
- Laboratory standards are traceable to National/International Standards.
- Calibration Certificates issued for Weight & Measure parameters i.e. Mass, Balance, Volumetric equipment Measuring Seales / Tapes etc. are for scientific purpose only and should not be used for Trade / Commercial use.

CALIBRATED / CHECKED BY

Ra

(CALIBRATION ENGINEER)



AUTHORISED SIGNATORY

Tejb
TEJ BAHADUR