

## Certificate of Calibration

Date : 29 March 2024

**Name of Instrument :** BTS 350 (S.No. 8017512371)

**Customer Name:** MOLXPRT Diagnostics Pvt Ltd, 106, First Floor, Mahagun Mart,  
GH-04, Sector-16, Greater Noida West

Following parameters have been checked and adjusted as per specification of Manufacturer.

Details of Calibration:

LIGHT INTENSITY				ADJUSTMENT PUMP				
29/03/24 19:47:50				29/03/24 19:45:52				
LEDS	TINT	MAIN	REFERENCE	TEST VOLUME (μL):		VALUE	CURRENT	NEW
340	13	924405	215948	SAMPLE VOLUME	44430	44833		
405	13	924254	651589	POSITIONING	2324	2304		
505	28	983416	711453	PUMP DELAY	2			
535	14	925115	472867					
560	12	975555	559572					
600	23	967244	599978					
635	12	939655	548004					
670	11	921752	583541					

Date of Calibration: 29 March 2024

Next Date of Calibration: 30 March 2025

For Innogen Healthcare

Authorized Signatory



## INNOGEN HEALTH CARE

© Off : No. 4, 1st Floor, G-83, Main Bazar Laxmi Nagar, Delhi-110092  
© +91-9312793140, +91-9891762590 © innogenhealthcare@gmail.com