

VIEW LAB MONTHLY SUMMARY

Lab Name **SADAR HOSPITAL**

[Details About Robust Analysis](#)

Lab No **18465**

[Detail About Monthly Summary](#)

Month **February**

[Detail about Z-Score](#)

Year **2024**

Constituent Group **Chemistry II**

[Click on the analyte to view Graphical Data](#)

All Analyser Result

Print

Date of Result Entered : **19/02/2024**

Date of Report Published : **05/03/2024**

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	AV	Participants CV	SDPA	Your Value	Z Score	u*
1	GLUCOSE II	HEXOKINASE II	Any Analyser (Automation / Semi Automation)	84	97.18	6.95	6.76	93.6 mg/dL	-0.53	1.47
2	UREA II	Urease UV / GLDH II	Any Analyser (Automation / Semi Automation)	1168	26.64	15.95	4.25	31.2 mg/dL	1.07	0.25
3	CREATININE II	Jaffes Kinetic- Alkaline Picrate II	Any Analyser (Automation / Semi Automation)	1273	2.26	12.99	0.29	2.5 mg/dL	0.82	0.02
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	Any Analyser (Automation / Semi Automation)	1130	1.76	16.01	0.28	1.3 mg/dL	-1.63	0.02
5	T-PROTEIN II	Biuret - Colorimetric II	Any Analyser (Automation / Semi Automation)	1322	5.99	9.60	0.57	5.6 g/dL	-0.68	0.03
6	ALBUMIN II	BCG - Colorimetric II	Any Analyser (Automation / Semi Automation)	1294	3.61	9.21	0.33	3.4 g/dL	-0.63	0.02
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Any Analyser (Automation / Semi Automation)	1499	5.37	17.70	0.95	6.3 mg/dL	0.98	0.05
8	CHOLESTROL II	CHOD-PAP II	Any Analyser (Automation / Semi Automation)	1685	125.93	11.88	14.96	115.6 mg/dL	-0.69	0.73
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Any Analyser (Automation / Semi Automation)	1488	159.25	13.19	21.01	158.5 mg/dL	-0.04	1.09

u* - Method of Uncertainty

Z-Score	Interpretation
$ z \leq 2.0$	Acceptable
$2.0 < z < 3.0$	Warning Signal
$ z \geq 3.0$	Unacceptable (action Signal)

Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out