



**SVR BIOTECH**

H.No : 9-107/3/2, GVR Heights, Nagaram Main Road,  
Near Vijaya Hospital, Secunderabad - 500 083.  
GSTIN/UIN: 36ADQFS4180E1Z5  
Mail : [info@svrbiotech.com](mailto:info@svrbiotech.com)

## CALIBRATION CERTIFICATE

INSTRUMENT NAME : URINE ANALYSER - YD 200  
SERIAL NUMBER : 200190808005  
INSTRUMENT IDENTIFICATION : EQP - 261  
MANUFACTURER NAME : MR INTERNATIONAL HEALTHCARE  
TECHNOLOGY CO LTD  
IMPORTED AND MARKETED IN INDIA BY: SVR BIOTECH  
STANDARD USED : SELF CALIBRATION MODE  
STANDARD TRACEABILITY : IN HOUSE CERTIFICATE

### CALIBRATION:

The instrument function well in all the systems check, self check and maintenance procedures. The equipment was calibration and the Red, Blue & Green (R,B & G) filters of all the parameters and top & bottom were found to be within the acceptable range of  $333 \pm 10$ .

QC PERFORMANCE : ACCEPTABLE  
SELF CALIBRATION : ACCEPTABLE  
PERFORMANCE QUALIFICATION : PASSED  
CALIBRATION DATE : 01/01/2024  
CALIBRATION : PASSED  
CALIBRATION VALID UP TO : 30/06/2024

**K. RAJASHEKAR RAJU**

*K. Rajashekar Raju*



**TECHNICAL SERVICE ENGINEER**



# calibration raw Data

Date:01-01-2024 16:18:58  
Calibration OK

Date:01-01-2024 16:20:55  
No. 0016  
URO 17 umol/L  
BIL Neg  
KET Neg  
BLD Neg  
PRO Neg  
\*NIT Pos  
\*LEU 125 cells/uL  
GLU 5.5 mmol/L  
SG >=1.030  
pH 5.0

*Raw*  
1/1/24

