



Metropolis Healthcare Ltd., EQAS

SHRI MAHAVIR MEDICAL CENTRE (1077)

Outlier And Analyte Summary Report

Outlier Details For Cycle No 230412 and Sample No 01

Report Date : 13/02/2024

Analyte	Instrument	Result Value	Standard Unit	Z-Score
No Outlier				
Legend @ : z score \leq 2.0 - Acceptable ! : 2.0 < z score < 3.0 - Warning X : z score \geq 3 - Unacceptable # : Not Evaluated ⌚ : Delayed Result Entry * : Not considered for evaluation.				

Problem Classification: _____

Corrective Action: _____

Reviewed by: _____

Dated: _____

Instrument : Mindray BS-240

Analyte	Standard Unit	Result Value	Mean	Z-Score	RMZ
@ Urea	mg/dl	23.06	21.31	0.43	0.19
@ Creatinine	mg/dl	0.81	0.74	0.56	0.22
@ Bilirubin Total	mg/dl	0.11	0.13	-0.55	-0.65
@ Total Protein	g/dL	6.20	5.76	1.44	0.83
@ Albumin	g/dL	3.86	3.46	1.80	-3.48
@ Cholesterol	mg/dl	145.60	131.61	1.81	0.95
@ Triglycerides	mg/dl	138.20	123.58	0.99	0.64
@ HDL	mg/dl	26.60	25.94	-0.11	0.21
@ ALT	U/L	13.10	12.13	1.25	3.42
@ AST	U/L	17.70	17.54	0.75	2.31

Instrument : BeneSphera Electrolyte Analyzer

Analyte	Standard Unit	Result Value	Mean	Z-Score	RMZ
@ Chloride	mmol/L	83.80	80.45	0.30	0.43
@ Sodium	mmol/L	116.90	111.65	0.80	1.13
@ Potassium	mmol/L	3.85	4.10	0.12	0.28

Legend @ : Acceptable
! : 2.0 < z score < 3.0 - Warning
X : z score ≥3 - Unacceptable
: Not Evaluated
⊙ : Delayed Result Entry
* : Not considered for evaluation.

Total Parameters	13
Not Evaluated Parameters	0
Evaluated Parameters	13
Outlier Parameters (X)	0
EQAS Score Biochemistry	100.00 %

Dr Puneet Kumar Nigam

Dr Puneet Kumar Nigam
PT coordinator & Technical Manager, MHL EQAS
Unit No. 409-416.
Commercial Building - 1A
Kohinoor Mall, Kirod Road, Kurla (W),
Mumbai - 400070



Metropolis Healthcare Ltd., EQAS

SHRI MAHAVIR MEDICAL CENTRE (1077)

Outlier And Analyte Summary Report

Outlier Details For Cycle No 230412 and Sample No 02

Report Date : 13/02/2024

Analyte	Instrument	Result Value	Standard Unit	Z-Score
No Outlier				
Legend @ : z score \leq 2.0 - Acceptable ! : 2.0 < z score < 3.0 - Warning X : z score \geq 3 - Unacceptable # : Not Evaluated ⌚ : Delayed Result Entry * : Not considered for evaluation.				

Problem Classification: _____

Corrective Action: _____

Reviewed by: _____

Dated: _____

Instrument : Mindray BS-240

Analyte	Standard Unit	Result Value	Mean	Z-Score	RMZ
@ Glucose	mg/dl	80.30	70.82	1.39	0.28
@ Urea	mg/dl	23.16	21.51	0.48	0.19
@ Creatinine	mg/dl	0.87	0.73	0.51	0.22
@ Bilirubin Total	mg/dl	0.37	0.32	0.25	-0.65
@ Bilirubin Direct	mg/dl	0.13	0.13	-0.05	-0.29
@ Total Protein	g/dL	6.66	5.86	1.78	0.83
@ Albumin	g/dL	3.97	3.45	1.78	-3.48
@ Cholesterol	mg/dl	148.30	129.57	1.74	0.95
@ Triglycerides	mg/dl	170.50	147.56	1.53	0.64

Instrument : BeneSphera Electrolyte Analyzer

Analyte	Standard Unit	Result Value	Mean	Z-Score	RMZ
@ Chloride	mmol/L	86.20	84.55	0.14	0.43
@ Sodium	mmol/L	120.30	116.60	1.13	1.13
@ Potassium	mmol/L	4.07	4.13	0.27	0.28

Legend @ : Acceptable

! : 2.0 < z score < 3.0 - Warning


X : z score ≥ 3 - Unacceptable

: Not Evaluated

⊙ : Delayed Result Entry

* : Not considered for evaluation.

Total Parameters	12
Not Evaluated Parameters	0
Evaluated Parameters	12
Outlier Parameters (X)	0
EQAS Score Biochemistry	100.00 %



Dr Puneet Kumar Nigam
PT coordinator & Technical Manager, MHL EQAS

Unit No. 409-416.

Commercial Building - 1A

Kohinoor Mall, Kirod Road, Kurla (W),

Mumbai - 400070