



## Calibration Certificate

Date 23 Jan 2024

To,  
Shree Vighnharta Superspeciality Hospital ,Dhule.

Dear Sir,  
As desired by you please find enclosed herewith Technical calibration report  
for your Urine Analyzer Model: Laura Smart Sr. No. 6527-1-202009.

1. Date of Calibration - : 23 Jan 2024
2. Next Calibration date- : 22 Jan 2025

### Calibration status

Ref	R=330	G=334	B=334
LEU	R=330	G=334	B=335
NIT	R=330	G=334	B=335
URO	R=330	G=334	B=335
PRO	R=330	G=334	B=335
pH	R=330	G=334	B=335
BLD	R=330	G=334	B=334
S.G.	R=330	G=334	B=334
KET	R=331	G=334	B=334
BIL	R=330	G=334	B=335
GLU	R=330	G=334	B=335
VTC	R=330	G=334	B=335

Calibration value Range 333+/- 10 ( 323 -343 )

Calibration of instrument is ok.  
Thanking you and assuring of our prompt attention at all times.

Yours faithfully,

For TRANSASIA BIO-MEDICALS LTD.

Rajesh Wankhede  
Regional Service Manager

