

Parashar Micro Measurement Pvt. Ltd.

B-59, SECTOR-64, NOIDA (U.P.) 201 307
 PH.: 0120-4252360-65, Email : pmmpl.india@gmail.com
 Website : www.pmmpl.in, www.parashar.info



Calibration Certificate

ULR - CC279524000022279F		Page 1 of 1
Certificate Number : PMM/100424/4-1		Date of Request 10.04.2024
Calibrated For M/s. Sadar Hospital Central Pathology , Aurangabad.	Calibrated on 10.04.2024	Suggested Due Date 09.04.2025
	Cert. Issue Date 11.04.2024	

Description of Equipments

Nomenclature	Centrifuge	Accuracy/Class	Not Mentioned
Range	0 - 3500 rpm	Condition of UUC	Physically Ok
Resolution / Least Count	10 rpm	Serial No.	ZDAN-00042
Make & Model No.	Remi / R-8C	Party ID Mark No.	-----
Material / Type	-----	Location	-----

Master Equipment / Standard Used

Sr.No.	Nomenclature	Make & Model	Serial No. / I/D Mark	Calibrated From	Certificate No.	Due Date of Calibration
1	Dig. Tachometer	Metravi	210813911 PMM/TM/09	PMMPL	PMM/1529	01.12.2024

Standard used for calibration are traceable to Accredited lab, for Standard ISO/IEC : 17025 - 2017 or National Standards Through unbroken chain of calibration

Calibration Performed at	Site	Environmental	Temperature = 23 ± 5° C
Calibration Procedure No.	CP/MMV/S/03	Condition(s)	Humidity = 30 to 75 % RH
Reference Standard / Guideline	-----		

Mechanical Calibration

CALIBRATION RESULTS

Sr.No.	Nominal Value in UUC (in rpm)	Actual Observed Value in Std. (in rpm)	Error (in rpm)	± Uncertainty in Measurement (in rpm)
1	1000	1001	-1	13
2	1500	1501	-1	13
3	2000	2002	-2	13
4	3000	3003	-3	13
5	3500	3504	-4	13

UUC - Unit Under Calibration

Std. - Standard Instrument

Uncertainty of Measurement (at approx 95% Confidence Level with Coverage factor $k = 2$)

Calibrated By : *Sudhir*
 Designation : Calibration Engineer



Approved By : *Jai Prakash*
 Designation : Tech. Manager / Auth. Sign.

- Conditions : 1. This certificate refers only to the particular item submitted for calibration.
 2. The calibration result reported in this certificate are valid at the time of and under the stated conditions of measurement.
 3. This particular certificate can not be reproduced except in full, without prior permission of chief executive officer of the lab.