

VIEW LAB MONTHLY SUMMARY

Lab Name	BISWESWARI ROY DIAGNOSTIC & POLYCLINIC CENTRE	Details About Robust Analysis
Lab No	19513	Detail About Monthly Summary
Month	March	Detail about Z-Score
Year	2024	
Constituent Group	Chemistry II	

Click on the analyte to view Graphical Data

[All Analyser Result](#) [Print](#)

Date of Result Entered : 18/03/2024

Date of Report Published : 06/04/2024

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	AV	Participants		Your Value	Z Score	u*
						CV	SDPA			
1	GLUCOSE II	GOD-POD II	Any Analyser (Automation / Semi Automation)	1640	385.53	10.46	40.34	432 mg/dL	1.15	1.99
2	UREA II	Urease UV / GLDH II	Any Analyser (Automation / Semi Automation)	1166	64.17	12.02	7.71	59.6 mg/dL	-0.59	0.45
3	CREATININE II	Jaffes Kinetic-Alkaline Picrate II	Any Analyser (Automation / Semi Automation)	1233	6.66	11.71	0.78	6.49 mg/dL	-0.22	0.04
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	Any Analyser (Automation / Semi Automation)	1435	2.60	16.30	0.42	2.61 mg/dL	0.02	0.02
5	T-PROTEIN II	Biuret - Colorimetric II	Any Analyser (Automation / Semi Automation)	1251	4.33	12.37	0.54	3.98 g/dL	-0.65	0.03
6	ALBUMIN II	BCG - Colorimetric II	Any Analyser (Automation / Semi Automation)	1285	2.65	11.67	0.31	2.38 g/dL	-0.87	0.02
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Any Analyser (Automation / Semi Automation)	1238	4.30	14.99	0.64	3.26 mg/dL	-1.62	0.04
8	CHOLESTROL II	CHOD-PAP II	Any Analyser (Automation / Semi Automation)	1453	90.30	11.84	10.69	86.1 mg/dL	-0.39	0.56
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Any Analyser (Automation / Semi Automation)	1296	81.75	17.62	14.41	65.2 mg/dL	-1.15	0.80

u* - Method of Uncertainty

Z-Score	Interpretation
$ z \leq 2.0$	Acceptable
$2.0 < z < 3.0$	Warning Signal
$ z \geq 3.0$	Unacceptable (action Signal)

Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
--------------------------------	--------------------------------	------------------------------------	--

VIEW LAB MONTHLY SUMMARY

Lab Name	BISWESWARI ROY DIAGNOSTIC & POLYCLINIC CENTRE	Details About Robust Analysis
Lab No	19513	Detail About Monthly Summary
Month	February	Detail about Z-Score
Year	2024	
Constituent Group	Chemistry II	

Click on the analyte to view Graphical Data

[All Analyser Result](#) [Print](#)

Date of Result Entered : 09/02/2024

Date of Report Published : 05/03/2024

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	AV	Participants		Your Value	Z Score	u*
						CV	SDPA			
1	GLUCOSE II	GOD-POD II	Any Analyser (Automation / Semi Automation)	1627	96.57	9.21	8.90	99.8 mg/dL	0.36	0.44
2	UREA II	Urease UV / GLDH II	Any Analyser (Automation / Semi Automation)	1168	26.64	15.95	4.25	34.5 mg/dL	1.85	0.25
3	CREATININE II	Jaffes Kinetic-Alkaline Picrate II	Any Analyser (Automation / Semi Automation)	1273	2.26	12.99	0.29	2.82 mg/dL	1.90	0.02
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	Any Analyser (Automation / Semi Automation)	1130	1.76	16.01	0.28	1.88 mg/dL	0.43	0.02
5	T-PROTEIN II	Biuret - Colorimetric II	Any Analyser (Automation / Semi Automation)	1322	5.99	9.60	0.57	6.18 g/dL	0.33	0.03
6	ALBUMIN II	BCG - Colorimetric II	Any Analyser (Automation / Semi Automation)	1294	3.61	9.21	0.33	3.43 g/dL	-0.54	0.02
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Any Analyser (Automation / Semi Automation)	1499	5.37	17.70	0.95	4.62 mg/dL	-0.79	0.05
8	CHOLESTROL II	CHOD-PAP II	Any Analyser (Automation / Semi Automation)	1685	125.93	11.88	14.96	128 mg/dL	0.14	0.73
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Any Analyser (Automation / Semi Automation)	1488	159.25	13.19	21.01	211 mg/dL	2.46	1.09

u* - Method of Uncertainty

Z-Score	Interpretation
$ z \leq 2.0$	Acceptable
$2.0 < z < 3.0$	Warning Signal
$ z \geq 3.0$	Unacceptable (action Signal)

Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
--------------------------------	--------------------------------	------------------------------------	--

VIEW LAB MONTHLY SUMMARY

Lab Name	BISWESWARI ROY DIAGNOSTIC & POLYCLINIC CENTRE	Details About Robust Analysis
Lab No	19513	Detail About Monthly Summary
Month	January	Detail about Z-Score
Year	2024	
Constituent Group	Chemistry II	

Click on the analyte to view Graphical Data

[All Analyser Result](#)

[Print](#)

Date of Result Entered : 18/01/2024

Date of Report Published : 06/02/2024

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	AV	Participants		Your Value	Z Score	u*
						CV	SDPA			
1	GLUCOSE II	GOD-POD II	Any Analyser (Automation / Semi Automation)	1480	151.16	9.82	14.84	156 mg/dL	0.33	0.77
2	UREA II	Urease UV / GLDH II	Any Analyser (Automation / Semi Automation)	992	46.22	11.22	5.18	49.1 mg/dL	0.56	0.33
3	CREATININE II	Jaffes Kinetic- Alkaline Picrate II	Any Analyser (Automation / Semi Automation)	1130	1.67	12.25	0.20	1.78 mg/dL	0.54	0.01
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	Any Analyser (Automation / Semi Automation)	1042	3.41	12.84	0.44	3.4 mg/dL	-0.02	0.03
5	T-PROTEIN II	Biuret - Colorimetric II	Any Analyser (Automation / Semi Automation)	1105	5.12	9.00	0.46	4.79 g/dL	-0.72	0.03
6	ALBUMIN II	BCG - Colorimetric II	Any Analyser (Automation / Semi Automation)	1176	3.10	9.86	0.31	2.8 g/dL	-0.98	0.02
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Any Analyser (Automation / Semi Automation)	1217	7.71	11.58	0.89	7.3 mg/dL	-0.46	0.05
8	CHOLESTROL II	CHOD-PAP II	Any Analyser (Automation / Semi Automation)	1312	104.57	9.94	10.40	103 mg/dL	-0.15	0.57
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Any Analyser (Automation / Semi Automation)	1178	145.99	11.26	16.44	173 mg/dL	1.64	0.96

u* - Method of Uncertainty

Z-Score	Interpretation
$ z \leq 2.0$	Acceptable
$2.0 < z < 3.0$	Warning Signal
$ z \geq 3.0$	Unacceptable (action Signal)

Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
--------------------------------	--------------------------------	------------------------------------	--