




CALIBRATION CERTIFICATE												
ULR NO :	CC310224000001042F					CERTIFICATE No. :	UCSL/04-24/088_04					
Customer Name & Address:						Instrument receipt Date		15.04.2024				
M/s. MICRO HEALTH LABORATORIES -ENDO BIOLAB THANNEER PANDAL BUS STOP,PATHAIKKARA PERINTHALMANNA,679322						SRF No.		088				
						Calibration Date		17.04.2024				
						Next calibration date (As per customer request)		17.04.2025				
						Certificate Issue date		19.04.2024				
Identification on DUC (Device Under Calibration)												
Instrument Name		MICROPIPETTE				Make		DRAGON LAB				
Range		5-50 µl				Model		DRAGON LAB				
Least count		0.5 µl				Instrument Condition		OK				
Serial No.		YE19AAL0575568				Calibration Performed at		Mass & Volume Lab				
ID No.		MHL/E/C/2				Location		---				
Environmental Condition			Avg. Temperature (°C)			Avg. RH (%)			Avg. Atmospheric Pressure (hpa)			
			23.4			56			1010			
Equipment & Master Used For calibration												
Instrument Used		Calibrated		ID No:		Certificate No		Valid Upto		Sr No.		Accreditation No
Digital Weighing Balance		EIE		-		EIE/M/231046		30-07-2024		T1004465		CC-2222
Method Used :						Gravimetric method			Discipline & Category :			Mechanical - Volume
Calibration Reference Standard						Calibration Procedure						
ISO-8655-6 & ISO /TR 20461						UCSL/SOP/01-MPT						
CALIBRATION RESULTS @ 27 °C :-												
1. Lower Volume 5 µl												
1	2	3	4	5	6	7	8	9	10	Mean Volume V		
5.06	5.04	5.04	5.05	5.03	5.04	5.06	5.04	5.04	5.03	5.04		
Systematic Error es µl :			-0.04			Error Limits (± 0.50 µl)			Random Error in sr µl :			0.01
Systematic Error es in % :			-0.86			(±10.0%)			Random Error in Cv in % :			0.20
Error Limits (± 0.2 µl)						Random Error in sr µl :			0.15			
Error Limits (± 0.8%)						Random Error in Cv in % :			0.59			
2. Middle Volume 25 µl												
1	2	3	4	5	6	7	8	9	10	Mean Volume V		
25.24	25.24	25.32	25.26	25.15	25.20	25.58	25.09	25.48	25.30	25.29		
Systematic Error es µl :			-0.29			Error Limits (± 0.50 µl)			Random Error in sr µl :			0.15
Systematic Error es in % :			-1.15			(± 2.0%)			Random Error in Cv in % :			0.59
Error Limits (± 0.2 µl)						Random Error in sr µl :			0.10			
Error Limits (± 0.8%)						Random Error in Cv in % :			0.20			
3. Nominal Volume 50 µl												
1	2	3	4	5	6	7	8	9	10	Mean Volume V		
50.46	50.53	50.52	50.41	50.29	50.50	50.52	50.34	50.59	50.60	50.47		
Systematic Error es µl :			-0.47			Error Limits (± 0.50 µl)			Random Error in sr µl :			0.10
Systematic Error es in % :			-0.94			(± 1.0%)			Random Error in Cv in % :			0.20
Error Limits (± 0.2 µl)						Random Error in sr µl :			0.10			
Error Limits (± 0.4%)						Random Error in Cv in % :			0.20			
Decision Rule is Applied or Not :								Yes		No		
										✓		
<b>Conclusion /Remarks:</b> The Reported Uncertainty <u>from 0.5 to 10 µl is 0.06 µl and at 10 to 50 µl is 0.12 µl</u> at ,Coverage Factor K=2,which corresponds to a confidence level at approximately 95 % 1. The calibration results reported in this certificate are valid at the time of and the stated condition of measurement. 2. <b>This report is valid for Scientific &amp; Industrial Purpose Only</b> 3. This report should not be reproduced except in full without our prior permission in writing. 4. Calibration certificate without signature are not valid. 5. This Calibration Certificate relates only to the above DUC 6. All the measurements are traceable to SI units through unbroken chain of calibration from the competent laboratories as per ISO/IEC/17025 standard requirements. 7. DUC : Device under calibration												
Calibrated by				Checked By				Issued / Approved By				
												
								(APPU K MANI)				
								(Technical Director)				

