Ace Medical Solutions (India) Pvt. Ltd.

An ISO 9001-2015 Certified Company

No: 3/751, 1st Cross Street, Surveyor Colony, K.Pudur, Madurai - 625 007.

Phone: 0452 - 2561691 Mobile: +91 94434 74610



GSTIN. 33AAHCA8053H1ZR

DL No. MDU/4205/20B,MDU/3971/21B

CIN: U33110TN2009PTC072140

CALIBRATION CERTIFICATE

Customer	M/S. Vishnu Clinical Lab,	Certificate No:	AMS/LAB/289/2009
Name &	#Rasipuram Main Road,	Issuing Date:	03/09/2022
Address:	Vaiyappamalai.	Page No:	1 of 1

Details of Device under Calibration (DUC):

DUC:	Micropipette	Id. No:	VL/LAB/FMP/001
Make:	Dragon Lab	Department:	Laboratory
Model:		Calibrated On:	02/09/2022
Sl. No:	YE185AK0027089	Calibration Due	
Type:	Fixed	Suggested:	01/09/2023
Range &	1000μl	DUC Condition	. = T
Least Count:	3 3	on Receipt:	OK

Traceability Details:

Sl.No.	Test Equipment	Make	Certificate Number	Valid Up to
1	Electronic Weighing Scale	EWC/SF400C	PC/01/22/ATL/0432-02	30/01/2023
2	Thermo Hygrometer	Beetech/088	PC/01/22/ATL/0265-01	30/01/2023

Room Temperature: 25°C±2

Room Humidity: 30-60%

(At the time of Calibration)

Test Results Volume µl:

SL.	DUC Reading	Standard Reading (µl)		Correction
No.	(µl)	Trial 1	Trial 2	(µl)
1	1000.0	1010.0	1010.0	+10.0
2	1000.0	1010.0	1010.0	+10.0

Allowable Tolerance: User Defined

Note:

- 1. Calibration Points were selected as per customer request.
- 2. Each reading was taken by Repeatability Test (i.e.) each reading in a row is a mean of 10 readings.
- 3. Results Reported are valid at the time of and under the stated conditions of measurements.

Calibrated By:

(Calibration Engineer)



Approved by:

(Technical Manager)

93455 56140, 93455 56148

aceagm@gmail.com