



# Metropolis Healthcare Ltd., EQAS

AAKASH HOSPITAL-LAB (1090)

## Outlier And Analyte Summary Report

Outlier Details For Cycle No 230412 and Sample No 02

Report Date : 13/02/2024

Analyte	Instrument	Result Value	Standard Unit	Z-Score
! Glucose	Erba EM 200	85.40	mg/dl	2.48
! Total Protein	Erba EM 200	7.09	g/dL	2.09
Legend @ : z score $\leq$ 2.0 - Acceptable ! : 2.0 < z score < 3.0 - Warning X : z score $\geq$ 3 - Unacceptable # : Not Evaluated ⌚ : Delayed Result Entry * : Not considered for evaluation.				

**Problem Classification:** \_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_

**Corrective Action:** \_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_

**Reviewed by:** \_\_\_\_\_

**Dated:** \_\_\_\_\_

Instrument : Erba EM 200

Analyte	Standard Unit	Result Value	Mean	Z-Score	RMZ
! Glucose	mg/dl	85.40	74.09	2.48	0.63
@ Urea	mg/dl	23.80	21.32	1.25	1.56
@ Creatinine	mg/dl	0.61	0.77	-1.05	-0.42
@ Bilirubin Total	mg/dl	0.31	0.33	-0.03	0.24
@ Bilirubin Direct	mg/dl	0.14	0.13	0.36	0.29
! Total Protein	g/dL	7.09	6.08	2.09	0.90
@ Albumin	g/dL	3.60	3.53	0.17	-0.09
@ Uric Acid	mg/dl	4.50	4.19	0.68	-0.04
@ Calcium	mg/dl	7.20	6.57	0.92	0.09
@ Cholesterol	mg/dl	128.00	124.17	0.37	0.41
@ Triglycerides	mg/dl	155.60	146.07	0.72	0.87
@ ALP	U/L	45.00	43.50	0.11	0.38
@ GGT	U/L	24.00	22.57	0.70	0.47
@ Amylase	U/L	41.00	38.00	0.60	0.88

Legend @ : Acceptable

! : 2.0 < z score < 3.0 - Warning


X : z score ≥ 3 - Unacceptable

# : Not Evaluated

⊙ : Delayed Result Entry

\* : Not considered for evaluation.

Total Parameters	14
Not Evaluated Parameters	0
Evaluated Parameters	14
Outlier Parameters (X)	0
EQAS Score Biochemistry	100.00 %



**Dr Puneet Kumar Nigam**  
PT coordinator & Technical Manager, MHL EQAS

Unit No. 409-416.

Commercial Building - 1A

Kohinoor Mall, Kirool Road, Kurla (W),

Mumbai - 400070