



# Metropolis Healthcare Ltd., EQAS

AAKASH HOSPITAL-LAB (1090)

## Outlier And Analyte Summary Report

Outlier Details For Cycle No 230310 and Sample No 01

Report Date : 19/12/2023

Analyte	Instrument	Result Value	Standard Unit
# C-Reactive Protein Qualitative	Erba EM 200	Normal	mg/L
Legend @ : Acceptable X : Unacceptable # : Not Evaluated ⌚ : Delayed Result Entry			

Problem Classification: \_\_\_\_\_

\_\_\_\_\_

\_\_\_\_\_

\_\_\_\_\_

Corrective Action: \_\_\_\_\_

\_\_\_\_\_

\_\_\_\_\_

\_\_\_\_\_

Reviewed by: \_\_\_\_\_

Dated: \_\_\_\_\_

Instrument : Erba EM 200

Analyte	Standard Unit	Result Value	Accepted Value
# C-Reactive Protein Qualitative	mg/L	Normal	

Instrument : Manual

Analyte	Standard Unit	Result Value	Accepted Value
@ Widal test - Typhi O antigen (TO)	-	No Agglutination	1:80,NoA
@ Widal test - Typhi H antigen (TH)	-	No Agglutination	1:80,NoA
@ Widal test - ParaTyphi A-H antigen (AH)	-	No Agglutination	1:80,NoA
@ Widal test - ParaTyphi B-H antigen (BH)	-	No Agglutination	1:80,NoA
@ HBsAg	IU/mL	Non-reactive	Non-reactive
@ HCV	IU/mL	Non-reactive	Non-reactive
@ RPR (Syphilis Serology)	-	Non-reactive	Non-reactive
@ HIV Ab	S/CO	Non-reactive	Non-reactive
@ Typhidot IgM	-	Negative	Negative

Legend @ : Acceptable


@ : Acceptable

X : Unacceptable

# : Not Evaluated

⊙ : Delayed Result Entry

Total Parameters	10
Not Evaluated Parameters	1
Evaluated Parameters	9
Outlier Parameters (X)	0
EQAS Score Immunoassay Extended	100.00 %



**Dr Puneet Kumar Nigam**  
PT coordinator & Technical Manager, MHL EQAS

Unit No. 409-416.

Commercial Building - 1A

Kohinoor Mall, Kiroil Road, Kurla (W),

Mumbai - 400070



# Metropolis Healthcare Ltd., EQAS

AAKASH HOSPITAL-LAB (1090)

## Outlier And Analyte Summary Report

Outlier Details For Cycle No 230310 and Sample No 03

Report Date : 19/12/2023

Analyte	Instrument	Result Value	Standard Unit
# C-Reactive Protein Qualitative	Erba EM 200	Normal	mg/L
Legend @ : Acceptable X : Unacceptable # : Not Evaluated ⌚ : Delayed Result Entry			

**Problem Classification:** \_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_

**Corrective Action:** \_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_

**Reviewed by:** \_\_\_\_\_

**Dated:** \_\_\_\_\_

Instrument : Erba EM 200


Analyte	Standard Unit	Result Value	Accepted Value
# C-Reactive Protein Qualitative	mg/L	Normal	

Instrument : Manual

Analyte	Standard Unit	Result Value	Accepted Value
@ HBsAg	IU/mL	Reactive	Reactive
@ HCV	IU/mL	Reactive	Reactive
@ HIV Ab	S/CO	HIV-1 Ab Reactive	HIV-1 Ab Reactive

Legend @ : Acceptable  
 @ : Acceptable  
 X : Unacceptable  
 # : Not Evaluated  
 ⌚ : Delayed Result Entry

Total Parameters	4
Not Evaluated Parameters	1
Evaluated Parameters	3
Outlier Parameters (X)	0
EQAS Score Immunoassay Extended	100.00 %



**Dr Puneet Kumar Nigam**  
 PT coordinator & Technical Manager, MHL EQAS  
 Unit No. 409-416.  
 Commercial Building - 1A  
 Kohinoor Mall, Kiroi Road, Kurla (W),  
 Mumbai - 400070