



Date:
Effective Date:

Certificate of Calibration

Customer Name: Aakash Hospital

Model : Automated Hematology Analyzer H560

Serial No. : K11041910022

Calibration Done Date: 4.8.23

Next Calibration Due Date On or Before: 04/08/2024

Lab In-charge: . Dr Meena Metre

This is to certify that the above-mentioned product has been verified of calibration for CBC 5 parameters (WBC, RBC, HGB, MCV and PLT) according to the standard procedures provided by Erba Lachema s.r.o, Karasek.

Calibration at site performed by
Vineet Kumar Rai
Sr. Service Engineer
Transasia Bio-Medicals Ltd
Delhi

Encl:

1. Certificate of Inspection
2. Assay Sheet of Hematology Calibrator (H Cal)
3. Printouts
4. Traceability Document



Date: 04/08/2023
Effective Date: 04/08/2023

Certificate of Inspection

1. Model: Automated Hematology Analyzer H560
2. Serial No.: K11041910022
3. Calibration Date: 04/08/2023
4. Material used: H Cal (Lot No. PLUS0623, Expiry date: 10-July-2023)

By comparing your data to the results of the standard counters in Erba Lachema, the calibration for CBC 5 parameters using the measurement standard material (H Cal) was completed. The calibration result of 5 runs is summarized in the following table. Please refer to the attached sheets for the details.

Technical Service Department
Transasia Bio-Medicals Ltd