



Avi Scientific (India)

NABL ACCREDITED CALIBRATION LABORATORY
(Thermal, Pressure, Mass, Volume, RPM)



CC - 2645

CALIBRATION CERTIFICATE

Format No. CC-12G				Page No: 1 of 1			
Date of Calibration:		Next Cal. due on:		Date of Issue:		ULR No.:	
15.01.2024		14.01.2025		16.01.2024		CC26452400000016F	
SRF No. 2400014		SRF Date		15.01.2024		Certificate No: AVI 2401016	
2400014		15.01.2024		Job No. 2401016		Job Date 15.01.2024	
Discipline : Mechanical				Group : Volume			
Customer's Name & Address:				Shri Mahavir Medical Centre Shop No.4/5, Bhanukant Complex, Aarey Checknaka, Goregaon (E) Mumbai-400063			
Customer Reference/Challan No				-			
Receipt Date				15.01.2024			
Condition on receipt				Satisfactory			
Calibration Carried out:				INHOUSE			
Description of item Calibrated:							
Instrument				Micropipette			
Make				Labserv			
Model				-			
Serial No.				HW02502			
Identification No.				SMMC/INST/09			
Range				100 to 1000 µl			
Least Count				5 µl			
Accuracy				-			
Location				-			
Work Instruction No.				AVI/WI-12			
Environmental Condition		Water Temp		23.5		Temp. (°C): (25 ±2)	
		Z correction factor		1.0036		Relative Humidity (% RH) (40 - 60)	
						Atmospheric Pressure (hpa) 970-1030	
Equipment & Master Used for Calibration				Used Standard equipments are traceable to the National Standard.			
Equipment		Sr.No./ ID No.		Certificate No		Validity up to	
Digital Weighing Balance		AVI/DWB/04		AVI 2309010		10.09.2024	
						Traceability with AVI Scientific (India),CC2645	

OBSERVATION TABLE

Uni of Measurement µl

Sr No.	UUC Reading	Standard Reading	Systematic error	Accuracy (A) as per ISO 8655-2	Random error	Random error as per ISO 8655-2	Expanded Uncertainty (±)	Status
1	100	101.19	-1.19	0.8	0.000332	0.3	0.28	Within Accuracy
2	500	500.45	-0.45	4.0	0.000426	1.5	0.28	Within Accuracy
3	1000	1002.02	-2.02	8.0	0.000675	3	0.28	Within Accuracy

(The reported expanded uncertainty of measurement is multiplied by coverage factor k=2, which corresponds to a coverage probability of approximately 95% for a normal distribution)

Note: Above readings are the average of 10 readings.

Calibrated By	Authorized Signatory
(Niranjan Rajguru) Calibration Engineer	(Janardan Chavan) Technical Manager

—End of the certificate—

Please Refer Note Backside of the Page

ARCHERCAL Private Limited

(Formerly known as Archerchem Calibrations Pvt Ltd)

Laboratory : Unit No. B/103, Tanvi's Tanishka Industrial Estate, Tanvi Complex,

Near HP Petrol Pump, Dahisar (East), Mumbai - 400 068. India. • Tel.: +91 22 20892984

E-mail : sales1@archercal.com / admin@archercal.com • Website : www.archercal.com

CIN : U74999MH2018PTC304510



CALIBRATION CERTIFICATE

7.8/R/M&V-01, Rev 00

Certificate No.: ACPL/MV/L/0206/01/23-24

ULR No.: CC377224000001222F

Date of Calibration:	01-02-2024	Issue Date:	02-02-2024
Next Calibration Due Date:	31-07-2024	Page:	01 of 02

CUSTOMER NAME & ADDRESS : SHREE MAHAVIR MEDICAL CENTRE.
Shop 4/5 , Bhanukant Complex, Checknaka, Aarey Colony,
Goregoan (East), Mumbai -400 063.

DATE OF RECEIPT : 30-01-2024

CALIBRATED AT : Lab

INSTRUMENT : Micropipette

ID. NO. : SMMC/INST/07

CONDITION OF ITEM : Good

CALIBRATION PROCEDURE : WI/APL/CAL/MV/04
ISO 8655-6 (Latest Edition)

ENVIRONMENTAL CONDITION : Temperature : 23 ± 1 °C
Air Pressure : 1010 ± 30 hPa
Relative Humidity : 50 ± 10 % RH
Water Temperature: 21.5 °C
Z Correction Factor (µl/mg) 1.00336
Y Correction Factor ((1/°C)x°C) 1.00207

TRACEABILITY : This Certificate is issued in the field of calibration and provides traceability of measurement results to International systems of units (SI)

CALIBRATIONS RESULTS : The results have been presented on pages(s) 2 of this certificate including uncertainty of measurements.

UNCERTAINTY OF MEASUREMENT : The uncertainty stated is the expanded uncertainty of measurement obtained by multiplying the standard uncertainty by the coverage factor k=2 correspondes to confidence level of 95%.

CONFORMITY STATEMENT : On the Basis of Calibration results, it has been found that instruments submitted for calibration meets the requirements specified in manufacturers specification & ISO8655-2:2002 (E).

REMARKS : The measured values mentioned are the average of 10 readings
: The Reported Volume at 27°C is by the formula,
 $V_{27^{\circ}\text{C}}(\mu\text{l}) = m(\text{mg}) \times Z(\mu\text{l}/\text{mg}) \times Y((1/^{\circ}\text{C}) \times ^{\circ}\text{C})$.
: The test liquid used for calibration is distilled water as per ISO3696 requirement.

Calibrated By:

Prathamesh Mestri

Prathamesh Mestri
Calibration Engineer



Approved By :

Kailas Chilap

Kailas Chilap
Technical Director

ARCHERCAL Private Limited

(Formerly known as Archerchem Calibrations Pvt Ltd)

Laboratory : Unit No. B/103, Tanvi's Tanishka Industrial Estate, Tanvi Complex,
Near HP Petrol Pump, Dahisar (East), Mumbai - 400 068. India. • Tel.: +91 22 20892984

E-mail : sales1@archercal.com / admin@archercal.com • Website : www@archercal.com

CIN : U74999MH2018PTC304510



CALIBRATION CERTIFICATE

7.8/R/M&V-01,Rev.00

Certificate No.: ACPL/MV/L/0206/01/23-24

ULR No.: CC377224000001222F

Date of Calibration:	01-02-2024	Issue Date:	02-02-2024
Next Calibration Due Date:	31-07-2024	Page:	02 of 02

Micropipette

Make :	Labserv	Range :	5 to 50 μ l
Model :	---	Least Count :	0.5 μ l
Sr.No. :	HW01104	Accuracy :	As per ISO 8655-2
Tag/Id.No. :	SMMC/INST/07	Location :	---

Discipline : Mechanical Calibration	Product Group : Mass & Volume - Volume
-------------------------------------	--

Measurement Result :

Sr. No.	Nominal Volume in μ l	1	2	3	4	5	6	7	8	9	10	Volume at 27°C (Average) in μ l
		Mass of Water in mg										
1	5	5.007	5.006	5.083	5.096	5.112	5.002	5.043	5.021	5.011	5.003	5.066
2	25	24.880	24.856	24.862	24.888	24.856	24.862	24.890	24.896	24.873	24.869	25.009
3	50	49.933	49.911	49.984	49.967	49.903	49.943	49.922	49.987	49.906	49.961	50.213

Error (A) in μ l	Accuracy ($\pm A$) as per ISO 8655 in μ l	Precision (σ) in μ l	Precision ($\pm CV$) as per ISO 8655 in μ l	Expanded Uncertainty (\pm) in μ l	Status
0.066	0.50	0.043	0.20	0.015	Within Accuracy
0.009	0.50	0.015	0.20	0.015	Within Accuracy
0.213	0.50	0.032	0.20	0.015	Within Accuracy

REFERENCE STANDARDS USED

DESCRIPTION	RANGE	Sr.No./ ID.NO.	TRACEABILITY	VALID UPTO
Digital Micro Balance	Max Capacity : 0 - 5.1 g	513955/ ACPL/MUM/M&V/01	ACPL,MUMBAI (ACPL/MV/L/0919/01/23-24)	31-05-2024

- NOTE**
- : The report refer only the particular item calibrated at site / laboratory.
 - : The calibration result reported are valid at the time of and under the condition of measurement
 - : Certificate should not be reproduced, except in full without the prior permission of Laboratory
 - : Any correction in this certificate invalidates the certificate
 - : The calibration of under test is meant for scientific and industrial purpose only.

Calibrated By:  Prathamesh Mestri Calibration Engineer		Approved By:  Kailas Chilap Technical Director
---	---	--

***** End of Certificate *****