

INSTALLATION REPORT

AGAPPE  
YOUR BEST PARTNER IN DIAGNOSTICS

CUSTOMER DETAILS

Name of the customer: ESIC MC & Model Hospital

Installation Address: Rajaji Nagar, Bangalore

Tel: ..... PIN: 560010

Contact Person: Shakthi Email: esicmicroxy@gmail.com

Mobile: 9916486095

EQUIPMENT DETAILS

Model: Mispa 13

Serial No.: 21212221690

Invoice No.: ..... Date: ...../...../.....

Delivery Date: .....

Installation Date: 15/02/2022

Warranty expiry date: 14/02/2025

Invoiced from AGAPPE / Distributor (If through distributor, please mention the name)

Distributor Name: ..... Place: .....

SITE DETAILS

Vibration free platform  Yes  No

AC line input voltage: 229 ✓

Dust free  Yes  No

Earth Neutral Voltage: 1 ✓

Away from Sunlight  Yes  No

Air Conditioned  Yes  No

Away from TFR/AC/UPS  Yes  No

UPS available  Yes  No

TO BE FILLED BY THE CUSTOMER

Received the machine in good condition  Yes  No

Packing of the machine  Good  Not Satisfactory

Received all accessories  Yes  No  Partially

Comments if any: .....

Declaration: The items as per the order have been delivered, installed, programmed and handed over in good condition. Training has been imparted to our satisfaction.

For AGAPPE DIAGNOSTICS LTD.

For CUSTOMER

Staff Name: Naveen JOY

Name: Dr. Raksha Yoganand

Regions: SBZ6 HO: Blr-3

Sign: Raksha Yoganand

Sign: [Signature]

Date: 15/2/2022

Date: 15/02/2022

Seal: [Seal]

Dr. Raksha Yoganand / Professor and Head, Department of Microbiology