

Parashar Micro Measurement Pvt. Ltd.



The way to Technology
(21 Years of Excellence)

B-59, SECTOR-64, NOIDA (U.P.) 201 307
PH.: 0120-4252360-65, Email : pmmpl.india@gmail.com
Website : www.pmmpl.in, www.parashar.info



CC - 2795

Calibration Certificate

ULR - CC279524000022352F		Page 1 of 1				
Certificate Number : PMM/100424/12-1		Date of Request	10.04.2024			
Calibrated For	M/s. Sadar Hospital Central Pathology Lab., Sitamarhi.	Calibrated on	10.04.2024			
		Suggested Due Date	09.04.2025			
		Cert. Issue Date	12.04.2024			
Description of Equipments						
Nomenclature	Centrifuge	Accuracy/Class	Not Mentioned			
Range	0 - 3500 rpm	Condition of UUC	Physically Ok			
Resolution / Least Count	10 rpm	Serial No.	11178840			
Make & Model No.	Remi / R-8C	Party ID Mark No.	-----			
Material / Type	-----	Location	-----			
Master Equipment / Standard Used						
Sr.No.	Nomenclature	Make & Model	Serial No. / I/D Mark	Calibrated From	Certificate No.	Due Date of Calibration
1	Dig. Tachometer	Metravi	210813911 PMM/TM/09	PMMPL	PMM/1529	01.12.2024
Standard used for calibration are traceable to Accredited lab, for Standard ISO/IEC : 17025 - 2017 or National Standards Through unbroken chain of calibration						
Calibration Performed at		Site	Environmental Condition(s)	Temperature = 23 ± 5° C		
Calibration Procedure No.		CP/M/MV/S/03		Humidity = 30 to 75 % RH		
Reference Standard / Guideline		-----				
Mechanical Calibration						
CALIBRATION RESULTS						
Sr.No.	Nominal Value in UUC (in rpm)	Actual Observed Value in Std. (in rpm)	Error (in rpm)	± Uncertainty in Measurement (in rpm)		
1	1000	1001	-1	13		
2	1500	1502	-2	13		
3	2000	2003	-3	13		
4	3000	3004	-4	13		
5	3500	3505	-5	13		

UUC - Unit Under Calibration

Std. - Standard Instrument

Uncertainty of Measurement (at approx 95% Confidence Level with Coverage factor $k = 2$)

Calibrated By : *Sudhir*
Designation : Calibration Engineer



Approved By : *Jai Prakash*
Designation : Tech. Manager / Auth. Sign.

- Conditions : 1. This certificate refers only to the particular item submitted for calibration.
2. The calibration result reported in this certificate are valid at the time of and under the stated conditions of measurement.
3. This particular certificate can not be reproduced except in full, without prior permission of chief executive officer of the lab.