

Date: 17/04/2024

Dear Sir,

Lab ID – NABL BCL007495

Lab Name – KRM Healthcare Private Limited

For clarification of HIV test with three methodologies according to NACO Guidelines and some points in our Scope:

1. Kindly remove Urine WBC, Monocyte and basophil from scope as it is not covered by Metropolis PT provider.
2. Kindly remove CRP, ASO, RA, VDRL by TPHA method & LDL from Scope as it is not done in laboratory due to insufficient request from our patients. We outsource the test if received.
3. The actual methodology for hemoglobin is Non Cyan meth.
4. Three Different Methodologies for Rapid HIV Testing are as below:

Sr. No	Company Name of Kit Used	Principle
1	Vitros anti HIV 1 & 2	Immunometric Technique is used for HIV 1 & 2
2	Meril Quadro HIV 1-2 antibody (2 nd generation assay)	Immunochromatography Method
3	Determine HIV 1 & 2 antigen & antibody combo	Antigen & Antibody 4 th generation kit

5. The agreement with Biomedical waste disposal vendor has been renewed and attached for your reference

Regards,



Dr. Neeraj Gujar

MD. (Path)

Mobile: 9820833488

KRM Healthcare Private Limited

CIN No. U85300MH2022PTC377481

+91 9820833488



info@krmhealthcare.com
www.krmhealthcare.com



KRM Healthcare Pvt Ltd

436/437, Laxmi Plaza, Laxmi Industrial Estate, Andheri West, Mumbai, 400053





Metropolis Healthcare Ltd., EQAS

KRM HEALTHCARE (1019)

Outlier And Analyte Summary Report

Outlier Details For Cycle No 230412 and Sample No 01

Report Date : 13/02/2024

Analyte	Instrument	Result Value	Standard Unit
No Outlier			
Legend @ : Acceptable X : Unacceptable # : Not Evaluated ⌚ : Delayed Result Entry			

Problem Classification: _____

Corrective Action: _____

Reviewed by: _____


Dated: _____

Instrument : Manual

Analyte	Standard Unit	Result Value	Accepted Value
@ pH	-	6	5,5,5,6
@ Specific Gravity	-	1.025	1.020,1.025,1.030
@ Glucose (urine)	-	3+	+2,+3,+4
@ Protein (Urine)	-	2+	+1,+2,+3
@ Ketone (urine)	-	Negative	Neg
@ Blood (urine)	-	1+	+1,+2,+3
@ Bilirubin (urine)	-	Negative	Neg
@ Urobilinogen (urine)	-	Normal	Nor
@ Nitrite (urine)	-	Present	Pre

Instrument : -

Analyte	Standard Unit	Result Value	Accepted Value
@ Urine Pregnancy Test	-	Positive	Pos
Legend		Total Parameters	10
@ : Acceptable		Not Evaluated Parameters	0
@ : Acceptable		Evaluated Parameters	10
X : Unacceptable		Outlier Parameters (X)	0
# : Not Evaluated		EQAS Score Clinical Path - Urine Routine Chemistry	100.00 %
⊙ : Delayed Result Entry			



Dr Puneet Kumar Nigam
PT coordinator & Technical Manager, MHL EQAS
Unit No. 409-416.
Commercial Building - 1A
Kohinoor Mall, Kiroil Road, Kurla (W),
Mumbai - 400070