

Date: 17/04/2024

Dear Sir,

Lab ID – NABL BCL007495

Lab Name – KRM Healthcare Private Limited

For clarification of HIV test with three methodologies according to NACO Guidelines and some points in our Scope:

1. Kindly remove Urine WBC, Monocyte and basophil from scope as it is not covered by Metropolis PT provider.
2. Kindly remove CRP, ASO, RA, VDRL by TPHA method & LDL from Scope as it is not done in laboratory due to insufficient request from our patients. We outsource the test if received.
3. The actual methodology for hemoglobin is Non Cyan meth.
4. Three Different Methodologies for Rapid HIV Testing are as below:

Sr. No	Company Name of Kit Used	Principle
1	Vitros anti HIV 1 & 2	Immunometric Technique is used for HIV 1 & 2
2	Meril Quadro HIV 1-2 antibody (2 nd generation assay)	Immunochromatography Method
3	Determine HIV 1 & 2 antigen & antibody combo	Antigen & Antibody 4 th generation kit

5. The agreement with Biomedical waste disposal vendor has been renewed and attached for your reference

Regards,



Dr. Neeraj Gujar

MD. (Path)

Mobile: 9820833488

KRM Healthcare Private Limited

CIN No. U85300MH2022PTC377481

+91 9820833488



info@krmhealthcare.com
www.krmhealthcare.com



KRM Healthcare Pvt Ltd
436/437, Laxmi Plaza, Laxmi Industrial
Estate, Andheri West, Mumbai, 400053





Metropolis Healthcare Ltd., EQAS

KRM HEALTHCARE (1019)

Outlier And Analyte Summary Report

Outlier Details For Cycle No 230412 and Sample No 01

Report Date : 13/02/2024

Analyte	Instrument	Result Value	Standard Unit
X C-Reactive Protein Qualitative	VITROS 5600 (Wet)	Normal	mg/L
Legend @ : Acceptable X : Unacceptable # : Not Evaluated ⌚ : Delayed Result Entry			

Problem Classification: _____

Corrective Action: _____

Reviewed by: _____

Dated: _____

Instrument : VITROS 5600 (Wet)


Analyte	Standard Unit	Result Value	Accepted Value
X C-Reactive Protein Qualitative	mg/L	Normal	Abn

Instrument : Manual

Analyte	Standard Unit	Result Value	Accepted Value
@ HIV Ag + Ab	S/CO	Non-reactive	Non-reactive
@ HBsAg	IU/mL	Non-reactive	Non-reactive
@ HCV	IU/mL	Non-reactive	Non-reactive
@ RPR (Syphilis Serology)	-	Non-reactive	Non-reactive
@ Dengue NS 1	RU/ml	Negative	Negative
@ Typhidot IgM	-	Negative	Negative
@ Widal test - Typhi O antigen (TO)	-	No Agglutination	No Agglutination
@ Widal test - Typhi H antigen (TH)	-	No Agglutination	No Agglutination
@ Widal test - ParaTyphi A-H antigen (AH)	-	No Agglutination	No Agglutination
@ Widal test - ParaTyphi B-H antigen (BH)	-	No Agglutination	No Agglutination
@ Dengue IgM	RU/ml	Negative	Negative
@ Dengue IgG	RU/ml	Negative	Negative

Legend @ : Acceptable
 @ : Acceptable
 X : Unacceptable
 # : Not Evaluated
 ⊙ : Delayed Result Entry

Total Parameters	13
Not Evaluated Parameters	0
Evaluated Parameters	13
Outlier Parameters (X)	1
EQAS Score Immunoassay Extended	92.31 %



Dr Puneet Kumar Nigam
 PT coordinator & Technical Manager, MHL EQAS
 Unit No. 409-416.
 Commercial Building - 1A
 Kohinoor Mall, Kirol Road, Kurla (W),
 Mumbai - 400070



Metropolis Healthcare Ltd., EQAS

KRM HEALTHCARE (1019)

Outlier And Analyte Summary Report

Outlier Details For Cycle No 230412 and Sample No 02

Report Date : 13/02/2024

Analyte	Instrument	Result Value	Standard Unit
X C-Reactive Protein Qualitative	VITROS 5600 (Wet)	Normal	mg/L
# HCV	Manual	Non-reactive	-

Legend @ : Acceptable
X : Unacceptable
: Not Evaluated
⌚ Delayed Result Entry

Problem Classification: _____

Corrective Action: _____

Reviewed by: _____

Dated: _____

Instrument : VITROS 5600 (Wet)


Analyte	Standard Unit	Result Value	Accepted Value
X C-Reactive Protein Qualitative	mg/L	Normal	Abn

Instrument : Manual

Analyte	Standard Unit	Result Value	Accepted Value
@ HIV Ag + Ab	S/CO	HIV-1/HIV-2 Ab Reactive	HIV-1/HIV-2 Ab Reactive
@ HBsAg	IU/mL	Reactive	Reactive
# HCV	IU/mL	Non-reactive	
@ RPR (Syphilis Serology)	-	Non-reactive	Non-reactive
@ Dengue NS 1	RU/ml	Negative	Negative
@ Typhidot IgM	-	Negative	Negative
@ Widal test - Typhi O antigen (TO)	-	No Agglutination	No Agglutination
@ Widal test - Typhi H antigen (TH)	-	No Agglutination	No Agglutination
@ Widal test - ParaTyphi A-H antigen (AH)	-	No Agglutination	No Agglutination
@ Widal test - ParaTyphi B-H antigen (BH)	-	No Agglutination	No Agglutination
@ Dengue IgM	RU/ml	Negative	Negative
@ Dengue IgG	RU/ml	Negative	Negative

Legend @ : Acceptable
 @ : Acceptable
 X : Unacceptable
 # : Not Evaluated
 ⊙ : Delayed Result Entry

Total Parameters	13
Not Evaluated Parameters	1
Evaluated Parameters	12
Outlier Parameters (X)	1
EQAS Score Immunoassay Extended	91.67 %



Dr Puneet Kumar Nigam
 PT coordinator & Technical Manager, MHL EQAS
 Unit No. 409-416.
 Commercial Building - 1A
 Kohinoor Mall, Kirol Road, Kurla (W),
 Mumbai - 400070



Metropolis Healthcare Ltd., EQAS

KRM HEALTHCARE (1019)

Outlier And Analyte Summary Report

Outlier Details For Cycle No 230310 and Sample No 01

Report Date : 19/12/2023

Analyte	Instrument	Result Value	Standard Unit
# C-Reactive Protein Qualitative	VITROS 5600 (Wet)	Normal	mg/L
Legend @ : Acceptable X : Unacceptable # : Not Evaluated ⌚ : Delayed Result Entry			

Problem Classification: _____

Corrective Action: _____

Reviewed by: _____

Dated: _____

Instrument : VITROS 5600 (Wet)


Analyte	Standard Unit	Result Value	Accepted Value
# C-Reactive Protein Qualitative	mg/L	Normal	

Instrument : Manual

Analyte	Standard Unit	Result Value	Accepted Value
@ Widal test - Typhi O antigen (TO)	-	No Agglutination	1:80,NoA
@ Widal test - Typhi H antigen (TH)	-	No Agglutination	1:80,NoA
@ Widal test - ParaTyphi A-H antigen (AH)	-	No Agglutination	1:80,NoA
@ Widal test - ParaTyphi B-H antigen (BH)	-	No Agglutination	1:80,NoA
@ HIV Ag + Ab	S/CO	Non-reactive	Non-reactive
@ HBsAg	IU/mL	Non-reactive	Non-reactive
@ HCV	IU/mL	Non-reactive	Non-reactive
@ RPR (Syphilis Serology)	-	Non-reactive	Non-reactive
@ Dengue NS 1	RU/ml	Negative	Negative
@ Typhidot IgM	-	Negative	Negative

Legend @ : Acceptable
 @ : Acceptable
 X : Unacceptable
 # : Not Evaluated
 ⌚ : Delayed Result Entry

Total Parameters	11
Not Evaluated Parameters	1
Evaluated Parameters	10
Outlier Parameters (X)	0
EQAS Score Immunoassay Extended	100.00 %



Dr Puneet Kumar Nigam
 PT coordinator & Technical Manager, MHL EQAS
 Unit No. 409-416.
 Commercial Building - 1A
 Kohinoor Mall, Kirol Road, Kurla (W),
 Mumbai - 400070