



**PROFICIENCY TESTING REPORT**  
**ISHTM-AIIMS EXTERNAL QUALITY ASSURANCE PROGRAMME**  
 NABL accredited program as per ISO/IEC 17043:2010 standard  
 Organized By Department of Hematology, AIIMS, New Delhi-110029



*Duration of stability testing - minimum upto 8 days at ambient temp. after dispatch of specimens*

EQAP CODE No. : 3430

Distribution No.: 162-I

Month/Year: January/2024

Instrument ID: BENESPHERA AVANTOR

Model Name.: H33s

Serial No.: TH-87005308

Name & Contact No. of PT Co-ordinator: Dr. Manoranjan Mahapatra ( Prof. & Head), Hematology, AIIMS, Delhi,  
 Tel: 9013085730 , E-Mail : accuracy2000@gmail.com

Date of issue &amp; status of the report: 20-03-2024[Final].

### CBC and Retic Assessment

Test Parameters	S.No.	Among Lab (Accuracy Testing)						Within Lab (Precision Testing)			
		Your Result 1	Your Result 2	Your Results Sum of 2 Value	Consensus result sum of 2 values (Assigned Value)	Uncertainty of Assigned Values	Z Score	Yours Results Diff. of 2 Values	Consensus Result Diff. of 2 values (Assigned Value)	Uncertainty of Assigned Values	Z Score
WBC x10 <sup>3</sup> /µl	1	5.3	5	10.3	9.5	0.047	1.08	0.3	0.1	0.009	2.08
RBC x10 <sup>6</sup> /µl	1	3.78	3.76	7.54	7.66	0.014	-0.54	0.02	0.03	0.004	-0.27
Hb g/dl	1	12.8	12.8	25.6	25.5	0.039	0.17	0	0.1	0.012	-0.67
HCT%	1	38.1	37.7	75.8	80.7	0.289	-0.82	0.4	0.4	0.038	0.00
MCV-fl	1	100.9	100.4	201.3	209.6	0.687	-0.64	0.5	0.3	0.034	0.54
MCH-Pg	1	34	33.8	67.8	66.2	0.110	0.83	0.2	0.3	0.024	-0.45
MCHC-g/dl	1	33.8	33.5	67.3	63.2	0.233	0.89	0.3	0.3	0.028	0.00
Plt. x10 <sup>3</sup> /µl	1	143	132	275	381.5	2.344	-2.71	11	5	0.500	1.01
Retic %	2	11	10.8	21.8	18.2	0.462	0.46	0.2	0.5	0.054	-0.51

### P.S . Assesment

YOUR REPORT			CONSENSUS REPORT
DLC%	3	Nrbcs=0 , Poly=10 L=89, E=0, Mono/Promono=1 , B1=0 P.M.=, Mye=, Meta=, Other=occasional smudge cells	Lymp: 84-92, Poly: 6.5-11, nRBC/Blast/Myelo/Meta/Mono/Eosino: 0-5
RBC Morphology	3	normocytic normochromic	Predominantly: Normocytic/Normochromic with Mild Anisocytosis, Smudge Cells.
Diagnosis	3	CHRONIC LYMPHOCYTIC LEUKEMIA	Chronic Lymphoproliferative Disorder/CLL

**COMBINED DATA VALUES OF TOTAL PARTICIPANTS**

Test parameters	S.No.	Total participants covered in the current dist. 162--I	Total No. responded	% of Labs with Z Score 0-2		% of Labs with Z Score 2-3		% of Labs with Z Score >3	
				Among labs	Within lab	Among labs	Within lab	Among labs	Within lab
<b>WBC x10<sup>3</sup>/µl</b>	1	138	138	82.61	87.68	3.62	8.7	13.77	3.62
<b>RBC x10<sup>6</sup>/µl</b>	1	138	138	87.68	89.13	7.25	4.35	5.07	6.52
<b>Hb g/dl</b>	1	138	138	85.51	85.51	9.42	7.25	5.07	7.24
<b>HCT%</b>	1	138	137	93.43	87.59	4.38	7.3	2.19	5.11
<b>MCV-fl</b>	1	138	137	95.62	94.16	3.65	2.19	0.73	3.65
<b>MCH-Pg</b>	1	138	137	88.32	93.43	4.38	2.92	7.3	3.65
<b>MCHC-g/dl</b>	1	138	137	92.7	93.43	6.57	4.38	0.73	2.19
<b>Plt. x10<sup>3</sup>/µl</b>	1	138	138	89.86	91.3	7.25	5.07	2.89	3.63
<b>ReticCount%</b>	2	138	125	92.8	88.8	4	8.8	3.2	2.40
<b>PS Assessment</b>	3	138	126	Satisfactory :91.99%, Borderline Sat. :2.91%, Unsatisfactory :5.10%					

**\*Comments:**

**1). Among Lab (EQA) : Results acceptable.**

**2). Within Lab (IQA) : Precision acceptable.**

**Note-1: EQA** (External Quality Assurance) : Your Performance among various of participating labs in PT, to determine the accuracy of your results.

**IQA** ( Internal Quality Assurance) : Your Performance of comparison of two consecutive measurement values within your lab to test the precision of your autoanalyzer.

**Note-2:** Z score among & within lab were calculated, as per to ISO/IEC 13528:2015 standard. Z score among lab (EQA)= (Your Result Sum of two values - Consensus Result sum of two values)/(Normalised IQR)

Z score within lab (IQA)= (Your Result Difference of two values - Consensus Result difference of two values)/(Normalised IQR)

IQR = Quartile 3 - Quartile 1 of participant data, Normalised IQR = 0.7413 x IQR

**Note-3:** Z score 0 to  $\pm 2$ : Acceptable, Z score  $\pm 2$  to  $\pm 3$  :Warning Signal, Z score  $> \pm 3$  : Unacceptable [As per ISO/IEC 13528:2015 standard]

**Note-4:** Z score value between " $0$  to  $\pm 2$ " are texted in green colour. Z score value between " $\pm 2$  to  $\pm 3$ " are texted in orange colour. Z score value  $> \pm 3$  are texted in red colour.

**Note-5:** Homogeneity and stability testing of PT sample were done as per ISO 13528:2015 standard. To pass homogeneity test, between sample SD (Ss) should be smaller than the check value ( $0.3*SDPA$ ). To pass the stability test, average difference in measurement values of first and last day sample ( $\bar{x}-\bar{y}$ ) should be smaller than the check value ( $0.3*SDPA$ ).

**Note-6:** ISHTM-AIIMS-EQAP does not subcontract any task of its scheme

**Note-7:** Participants are free to use methods/analyzer of their own choice.

**Note-8:** Proficiency testing (PT ) samples are sent quarterly to each participant.

**Note-9:** All the necessary details regarding design and implementation of PT, are provided in the instruction sheet as well as on programme's website [www.ishtmaimseqap.com](http://www.ishtmaimseqap.com).

**Note 10:** Reports are kept confidential.

Report authorized by,



Dr. Manoranjan Mahapatra ( Prof. & Head)

PT Co-ordinator: ISHTM-AIIMS-EQAP

Department of Hematology, AIIMS, New Delhi

-----End Of Report-----



**PROFICIENCY TESTING REPORT**  
**ISHTM-AIIMS EXTERNAL QUALITY ASSURANCE PROGRAMME**  
 NABL accredited program as per ISO/IEC 17043:2010 standard  
 Organized By Department of Hematology, AIIMS, New Delhi-110029



*Duration of stability testing - minimum upto 8 days at ambient temp. after dispatch of specimens*

EQAP CODE No. : 3430

Distribution No.: 161-I

Month/Year: August/2023

Instrument ID: MINDRAY

Model Name.:

Serial No.:

**Name & Contact No. of PT Co-ordinator:** Dr. Manoranjan Mahapatra ( Prof. & Head), Hematology, AIIMS, Delhi,  
 Tel: 9013085730 , E-Mail : accuracy2000@gmail.com

**Date of issue & status of the report:** 10-01-2024[Final].

### CBC and Retic Assessment

Test Parameters	S.No.	Among Lab (Accuracy Testing)						Within Lab (Precision Testing)			
		Your Result 1	Your Result 2	Your Results Sum of 2 Value	Consensus result sum of 2 values (Assigned Value)	Uncertainty of Assigned Values	Z Score	Yours Results Diff. of 2 Values	Consensus Result Diff. of 2 values (Assigned Value)	Uncertainty of Assigned Values	Z Score
WBC x10 <sup>3</sup> /µl	1	8.3	8.1	16.4	10.2	0.047	8.12	0.2	0.1	0.010	0.90
RBC x10 <sup>6</sup> /µl	1	3.99	3.88	7.87	8.02	0.012	-0.70	0.11	0.04	0.003	1.89
Hb g/dl	1	12.8	12.5	25.3	25	0.032	0.51	0.3	0.1	0.011	2.70
HCT%	1	38.7	37.6	76.3	78	0.272	-0.33	1.1	0.4	0.036	1.89
MCV-fl	1	97	96.9	193.9	195.6	0.595	-0.14	0.1	0.2	0.027	-0.34
MCH-Pg	1	32.3	32.2	64.5	62.5	0.100	1.12	0.1	0.2	0.023	-0.45
MCHC-g/dl	1	33.3	33.2	66.5	63.7	0.223	0.65	0.1	0.2	0.026	-0.39
Plt. x10 <sup>3</sup> /µl	1	127	125	252	407	2.242	-3.87	2	4	0.411	-0.39
Retic %	2	10	8	18	16	0.345	0.30	2	0.5	0.048	2.45

### P.S . Assesment

YOUR REPORT			CONSENSUS REPORT
<b>DLC%</b>	3	Nrbcs=0 , Poly=5 L=90, E=2, Mono/Promono=3 , B1=0 P.M.=0, Mye=0, Meta=0, Other=SMUDGE CELLS	Lymp: 83-94, Poly: 3-8, nRBC/blast/mono/Myelo/Meta/Eosino: 0-5
<b>RBC Morphology</b>	3	ANISOPOIKILOCYTOSIS WITH NARMOCYTIC HYPOCHROMIC CELLS AND MACROVALOCYTES	Predominantly: Normocytic/Normochromic; Moderate: Anisocytosis, Microcytic, Hypochromia; Mild: Poikilocytosis, Tear Drop Cells.
<b>Diagnosis</b>	3	CHRONIC LYMPHOCYTIC LEUKEMIA (CLL)	Chronic Lymphoproliferative Disorder

**COMBINED DATA VALUES OF TOTAL PARTICIPANTS**

Test parameters	S.No.	Total participants covered in the current dist. 161--I	Total No. responded	% of Labs with Z Score 0-2		% of Labs with Z Score 2-3		% of Labs with Z Score >3	
				Among labs	Within lab	Among labs	Within lab	Among labs	Within lab
<b>WBC x10<sup>3</sup>/µl</b>	1	144	143	85.31	90.21	2.1	3.5	12.59	6.29
<b>RBC x10<sup>6</sup>/µl</b>	1	144	144	90.97	90.28	5.56	5.56	3.47	4.16
<b>Hb g/dl</b>	1	144	144	86.81	89.58	5.56	4.86	7.63	5.56
<b>HCT%</b>	1	144	143	95.8	90.91	2.8	3.5	1.4	5.59
<b>MCV-fl</b>	1	144	143	97.2	89.51	2.1	4.2	0.7	6.29
<b>MCH-Pg</b>	1	144	143	88.11	72.73	8.39	19.58	3.5	7.69
<b>MCHC-g/dl</b>	1	144	143	94.41	93.01	4.9	1.4	0.69	5.59
<b>Plt. x10<sup>3</sup>/µl</b>	1	144	142	89.44	90.85	6.34	4.93	4.22	4.22
<b>ReticCount%</b>	2	144	132	93.18	93.94	3.79	4.55	3.03	1.51
<b>PS Assessment</b>	3	144	130	Satisfactory :89.59%, Borderline Sat. :2.08%, Unsatisfactory :8.33%					

**\*Comments:**

1). **Among Lab (EQA) : CBC result for WBC & PLT unacceptable, please check calibration/human error. Remaining results acceptable.**

2). **Within Lab (IQA) : Precision acceptable.**

**Note-1: EQA** (External Quality Assurance) : Your Performance among various of participating labs in PT, to determine the accuracy of your results.

**IQA** ( Internal Quality Assurance) : Your Performance of comparison of two consecutive measurement values within your lab to test the precision of your autoanalyzer.

**Note-2:** Z score among & within lab were calculated, as per to ISO/IEC 13528:2015 standard. Z score among lab (EQA)= (Your Result Sum of two values - Consensus Result sum of two values)/(Normalised IQR)

Z score within lab (IQA)= (Your Result Difference of two values - Consensus Result difference of two values)/(Normalised IQR)

IQR = Quartile 3 - Quartile 1 of participant data, Normalised IQR = 0.7413 x IQR

**Note-3:** Z score 0 to ±2: Acceptable, Z score ±2 to ±3 :Warning Signal, Z score > ±3 : Unacceptable [As per ISO/IEC 13528:2015 standard]

**Note-4:** Z score value between "0 to ±2" are texted in green colour. Z score value between "±2 to ±3" are texted in orange colour. Z score value > ±3 are texted in red colour.

**Note-5:** Homogeneity and stability testing of PT sample were done as per ISO 13528:2015 standard. To pass homogeneity test, between sample SD (Ss) should be smaller than the check value (0.3\*SDPA). To pass the stability test, average difference in measurement values of first and last day sample ( $\bar{x}-\bar{y}$ ) should be smaller than the check value (0.3\*SDPA).

**Note-6:** ISHTM-AIIMS-EQAP does not subcontract any task of its scheme

**Note-7:** Participants are free to use methods/analyzer of their own choice.

**Note-8:** Proficiency testing (PT ) samples are sent quarterly to each participant.

**Note-9:** All the necessary details regarding design and implementation of PT, are provided in the instruction sheet as well as on programme's website [www.ishtmaiimseqap.com](http://www.ishtmaiimseqap.com).

**Note 10:** Reports are kept confidential.

Report authorized by,



Dr. Manoranjan Mahapatra ( Prof. & Head)

PT Co-ordinator: ISHTM-AIIMS-EQAP

Department of Hematology, AIIMS, New Delhi

-----End Of Report-----



**PROFICIENCY TESTING REPORT**  
**ISHTM-AIIMS EXTERNAL QUALITY ASSURANCE PROGRAMME**  
 NABL accredited program as per ISO/IEC 17043:2010 standard  
 Organized By Department of Hematology, AIIMS, New Delhi-110029



*Duration of stability testing - minimum upto 8 days at ambient temp. after dispatch of specimens*

EQAP CODE No. : 3430

Distribution No.: 160-I

Month/Year: June/2023

Instrument ID: TH87005308

Model Name.:

Serial No.:

**Name & Contact No. of PT Co-ordinator:** Dr. Manoranjan Mahapatra ( Prof. & Head), Hematology, AIIMS, Delhi,  
 Tel: 9013085730 , E-Mail : accuracy2000@gmail.com

**Date of issue & status of the report:** 08-08-2023[Final].

### CBC and Retic Assessment

Test Parameters	S.No.	Among Lab (Accuracy Testing)						Within Lab (Precision Testing)			
		Your Result 1	Your Result 2	Your Results Sum of 2 Value	Consensus result sum of 2 values (Assigned Value)	Uncertainty of Assigned Values	Z Score	Yours Results Diff. of 2 Values	Consensus Result Diff. of 2 values (Assigned Value)	Uncertainty of Assigned Values	Z Score
WBC x10 <sup>3</sup> /μl	1	4.7	4.7	9.4	12	0.089	-1.38	0	0.1	0.013	-0.90
RBC x10 <sup>6</sup> /μl	1	4.87	4.7	9.57	9.61	0.014	-0.17	0.17	0.04	0.004	2.92
Hb g/dl	1	15.6	15.3	30.9	30.3	0.054	0.65	0.3	0.1	0.011	2.70
HCT%	1	52.4	50.9	103.3	92.3	0.320	1.76	1.5	0.4	0.041	2.47
MCV-fl	1	108	107.7	215.7	193	0.544	2.13	0.3	0.3	0.031	0.00
MCH-Pg	1	32.4	32	64.4	63	0.106	0.71	0.4	0.3	0.023	0.45
MCHC-g/dl	1	30	29.7	59.7	64.6	0.229	-1.09	0.3	0.3	0.028	0.00
Plt. x10 <sup>3</sup> /μl	1	167	161	328	305	2.716	0.51	6	5	0.459	0.18
Retic %	2	4	2	6	5.4	0.169	0.19	2	0.3	0.032	4.59

### P.S . Assesment

YOUR REPORT		CONSENSUS REPORT
<b>DLC%</b>	3	Nrbcs=3 , Poly=7 L=4, E=0, Mono/Promono=0 , B1=48 P.M.=28, Mye=4, Meta=7, Other=Thrombocytopenia
<b>RBC Morphology</b>	3	Predominantly: Normocytic/Normochromic; Moderate: Microcytosis, Hypochromia; Mild: Anisocytosis
<b>Diagnosis</b>	3	Acute Myeloid Leukemia



**COMBINED DATA VALUES OF TOTAL PARTICIPANTS**

Test parameters	S.No.	Total participants covered in the current dist. 160--I	Total No. responded	% of Labs with Z Score 0-2		% of Labs with Z Score 2-3		% of Labs with Z Score >3	
				Among labs	Within lab	Among labs	Within lab	Among labs	Within lab
<b>WBC x10<sup>3</sup>/µl</b>	1	145	143	86.71	86.71	4.2	4.9	9.09	8.39
<b>RBC x10<sup>6</sup>/µl</b>	1	145	145	84.14	88.97	7.59	2.07	8.27	8.96
<b>Hb g/dl</b>	1	145	145	85.52	86.9	6.21	6.9	8.27	6.2
<b>HCT%</b>	1	145	143	93.01	90.21	4.9	5.59	2.09	4.2
<b>MCV-fl</b>	1	145	143	93.71	88.11	4.2	7.69	2.09	4.2
<b>MCH-Pg</b>	1	145	143	87.41	93.01	5.59	2.8	7	4.19
<b>MCHC-g/dl</b>	1	145	143	93.71	90.21	4.2	2.8	2.09	6.99
<b>Plt. x10<sup>3</sup>/µl</b>	1	145	143	90.21	93.01	8.39	1.4	1.4	5.59
<b>ReticCount%</b>	2	145	134	92.54	95.52	5.97	0.00	1.49	4.48
<b>PS Assessment</b>	3	145	129	Satisfactory :91.05%, Borderline Sat. :2.06%, Unsatisfactory :6.89%					

**\*Comments:**

1). **Among Lab (EQA) : Results acceptable.**

2). **Within Lab (IQA) : RETIC result is unacceptable, may be due to random/human error.**

**Note-1: EQA** (External Quality Assurance) : Your Performance among various of participating labs in PT, to determine the accuracy of your results.

**IQA** ( Internal Quality Assurance) : Your Performance of comparison of two consecutive measurement values within your lab to test the precision of your autoanalyzer.

**Note-2:** Z score among & within lab were calculated, as per to ISO/IEC 13528:2015 standard. Z score among lab (EQA)= (Your Result Sum of two values - Consensus Result sum of two values)/(Normalised IQR)

Z score within lab (IQA)= (Your Result Difference of two values - Consensus Result difference of two values)/(Normalised IQR)

IQR = Quartile 3 - Quartile 1 of participant data, Normalised IQR = 0.7413 x IQR

**Note-3:** Z score 0 to ±2: Acceptable, Z score ±2 to ±3 :Warning Signal, Z score > ±3 : Unacceptable [As per ISO/IEC 13528:2015 standard]

**Note-4:** Z score value between“0 to ±2” are texted in green colour. Z score value between“±2 to ±3” are texted in orange colour. Z score value > ±3 are texted in red colour.

**Note-5:** Homogeneity and stability testing of PT sample were done as per ISO 13528:2015 standard. To pass homogeneity test, between sample SD (Ss) should be smaller than the check value (0.3\*SDPA). To pass the stability test, average difference in measurement values of first and last day sample ( $\bar{x}-\bar{y}$ ) should be smaller than the check value (0.3\*SDPA).

**Note-6:** ISHTM-AIIMS-EQAP does not subcontract any task of its scheme

**Note-7:** Participants are free to use methods/analyzer of their own choice.

**Note-8:** Proficiency testing (PT ) samples are sent quarterly to each participant.

**Note-9:** All the necessary details regarding design and implementation of PT, are provided in the instruction sheet as well as on programme’s website www.ishtmaiimseqap.com.

**Note 10:** Reports are kept confidential.

Report authorized by,



Dr. Manoranjan Mahapatra ( Prof. & Head)

PT Co-ordinator: ISHTM-AIIMS-EQAP

Department of Hematology, AIIMS, New Delhi

-----End Of Report-----



**PROFICIENCY TESTING REPORT**  
**ISHTM-AIIMS EXTERNAL QUALITY ASSURANCE PROGRAMME**  
 NABL accredited program as per ISO/IEC 17043:2010 standard  
 Organized By Department of Hematology, AIIMS, New Delhi-110029



*Duration of stability testing - minimum upto 8 days at ambient temp. after dispatch of specimens*

EQAP CODE No. : 3430

Distribution No.: 159-I

Month/Year: March/2023

Instrument ID: TH87005308

Model Name.:

Serial No.:

Name & Contact No. of PT Co-ordinator: Dr. Seema Tyagi (Prof.), Hematology, AIIMS, Delhi,  
 Tel: 9013085730 , E-Mail : accuracy2000@gmail.com

Date of issue &amp; status of the report: 12-05-2023[Final].

### CBC and Retic Assessment

Test Parameters	S.No.	Among Lab (Accuracy Testing)						Within Lab (Precision Testing)			
		Your Result 1	Your Result 2	Your Results Sum of 2 Value	Consensus result sum of 2 values (Assigned Value)	Uncertainty of Assigned Values	Z Score	Yours Results Diff. of 2 Values	Consensus Result Diff. of 2 values (Assigned Value)	Uncertainty of Assigned Values	Z Score
WBC x10 <sup>3</sup> /µl	1	2.8	2.7	5.5	6.4	0.076	-0.69	0.1	0.1	0.009	0.00
RBC x10 <sup>6</sup> /µl	1	3.73	3.7	7.43	7.6	0.012	-0.85	0.03	0.03	0.003	0.00
Hb g/dl	1	11.9	11.6	23.5	23.3	0.030	0.39	0.3	0.1	0.011	2.70
HCT%	1	34.9	34.6	69.5	72	0.202	-0.57	0.3	0.3	0.033	0.00
MCV-fl	1	93.6	93.4	187	189	0.481	-0.21	0.2	0.3	0.031	-0.27
MCH-Pg	1	32	31.2	63.2	61.6	0.097	1.03	0.8	0.3	0.026	1.69
MCHC-g/dl	1	34.3	33.3	67.6	65.5	0.184	0.56	1	0.3	0.028	1.89
Plt. x10 <sup>3</sup> /µl	1	163	157	320	371	1.993	-1.40	6	6	0.505	0.00
Retic %	2	10	8	18	18	0.553	0.00	2	0.6	0.063	2.36

### P.S . Assesment

YOUR REPORT		CONSENSUS REPORT
DLC%	3 Nrbcs=3 , Poly=46 L=5, E=1, Mono/Promono=0 , B1=2 P.M.=18, Mye=16, Meta=9, Other=Platelets: mild thrombocytopenia	Poly: 45 - 60, Myelo: 13 - 25, Meta: 7- 15, Lympho: 2- 6, Eosino: 1-2, Promyelo: 1-6, Blast: 1-4, Mono: 1 - 2, nRBC/Baso: 0-5
RBC Morphology	3 Normocytic Normochromic, occasional stomatocytes and nRBCs seen.	Predominantly: Normocytic/Normochromic; Moderate: Anisocytosis, hypochromia, Microcytosis; Mild: Macrocytosis, Poikilocytosis
Diagnosis	3 Atypical CML in chronic phase with absence of basophilia	Chronic Myeloid Leukemia (Chronic Phase)

**COMBINED DATA VALUES OF TOTAL PARTICIPANTS**

Test parameters	S.No.	Total participants covered in the current dist. 159--I	Total No. responded	% of Labs with Z Score 0-2		% of Labs with Z Score 2-3		% of Labs with Z Score >3	
				Among labs	Within lab	Among labs	Within lab	Among labs	Within lab
<b>WBC x10<sup>3</sup>/µl</b>	1	146	145	84.14	88.28	5.52	5.52	10.34	6.2
<b>RBC x10<sup>6</sup>/µl</b>	1	146	146	84.93	91.1	9.59	4.79	5.48	4.11
<b>Hb g/dl</b>	1	146	146	88.36	86.3	5.48	6.16	6.16	7.54
<b>HCT%</b>	1	146	145	97.24	91.72	2.07	4.83	0.69	3.45
<b>MCV-fl</b>	1	146	145	95.17	85.52	4.14	7.59	0.69	6.89
<b>MCH-Pg</b>	1	146	145	82.07	90.34	9.66	5.52	8.27	4.14
<b>MCHC-g/dl</b>	1	146	145	95.86	95.17	2.07	2.76	2.07	2.07
<b>Plt. x10<sup>3</sup>/µl</b>	1	146	145	93.1	89.66	2.76	3.45	4.14	6.89
<b>ReticCount%</b>	2	146	128	96.09	84.38	2.34	11.72	1.57	3.90
<b>PS Assessment</b>	3	146	129	Satisfactory :95.28%, Borderline Sat. :2.02%, Unsatisfactory :2.70%					

**\*Comments:**

**1). Among Lab (EQA) : Results acceptable.**

**2). Within Lab (IQA) : Precision acceptable.**

**Note-1: EQA** (External Quality Assurance) : Your Performance among various of participating labs in PT, to determine the accuracy of your results.

**IQA** ( Internal Quality Assurance) : Your Performance of comparison of two consecutive measurement values within your lab to test the precision of your autoanalyzer.

**Note-2:** Z score among & within lab were calculated, as per to ISO/IEC 13528:2015 standard. Z score among lab (EQA)= (Your Result Sum of two values - Consensus Result sum of two values)/(Normalised IQR)

Z score within lab (IQA)= (Your Result Difference of two values - Consensus Result difference of two values)/(Normalised IQR)

IQR = Quartile 3 - Quartile 1 of participant data, Normalised IQR = 0.7413 x IQR

**Note-3:** Z score 0 to  $\pm 2$ : Acceptable, Z score  $\pm 2$  to  $\pm 3$  :Warning Signal, Z score  $> \pm 3$  : Unacceptable [As per ISO/IEC 13528:2015 standard]

**Note-4:** Z score value between "0 to  $\pm 2$ " are texted in green colour. Z score value between " $\pm 2$  to  $\pm 3$ " are texted in orange colour. Z score value  $> \pm 3$  are texted in red colour.

**Note-5:** Homogeneity and stability testing of PT sample were done as per ISO 13528:2015 standard. To pass homogeneity test, between sample SD (Ss) should be smaller than the check value (0.3\*SDPA). To pass the stability test, average difference in measurement values of first and last day sample ( $\bar{x}-\bar{y}$ ) should be smaller than the check value (0.3\*SDPA).

**Note-6:** ISHTM-AIIMS-EQAP does not subcontract any task of its scheme

**Note-7:** Participants are free to use methods/analyzer of their own choice.

**Note-8:** Proficiency testing (PT ) samples are sent quarterly to each participant.

**Note-9:** All the necessary details regarding design and implementation of PT, are provided in the instruction sheet as well as on programme's website [www.ishtmaiimseqap.com](http://www.ishtmaiimseqap.com).

**Note 10:** Reports are kept confidential.

Report authorized by,



Dr. Seema Tyagi (Prof.)

PT Co-ordinator: ISHTM-AIIMS-EQAP

Department of Hematology, AIIMS, New Delhi

-----End Of Report-----